

MEASURE TECHNO LAB

**2, B.T. ROAD (JAYANTI CINEMA COMPLEX)
BARRACKPORE, KOLKATA - 700120, W.B.**

Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519,
Mobile: 9831190974 , LAB:- 8100143376,
E-mail: measuretechno@yahoo.co.in

Form No. - MTL/22/2006

**CALIBRATION CERTIFICATE OF
DEEP FREEZER**



CALIBRATION CERTIFICATE NO.: MTL / MMC&H(SL) / R03 / 02 - 22

ULR - CC254522000006615F

Page: 1 of 1

1.0 Service Request No.: MTL / 23C / 02 / 21 - 22

1.1 Issued to: Prof. & HOD,
Department of Microbiology (Serology Lab)
Midnapore Medical College & Hospital,
Paschim Medinipur, West Bengal - 721101.

1.2 Description & Identification of item to be Calibrated:

| | | | |
|-------------------------|----------------------|----------------|------------------------------|
| a) Name: | Deep Freezer | b) Code No.: | MMC / MICRO / SERO / DF - 01 |
| c) SI No.: | 201904145 | d) Make: | Terumo Penpol |
| e) Model / Type: | DF 80U | f) Range: | (-) 80 °C |
| g) Sensor: | N.A. | h) Resolution: | 0.1 °C |
| i) End User: | Serology Lab | j) Accuracy: | N.S. |
| k) Calibration done at: | √ On Site / In House | | |

1.3 Date of receipt of item : 23-02-22

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 23-02-22

1.6 Recommended date of next calibration : 23-02-23

1.7 Date of Issue : 26-02-22

1.8 Environmental Conditions During Calibration: Temperature: 25.3 °C
Humidity: 64.7 % RH

1.9 Method of Calibration: SOP / TH / 05

2.0 Traceability :

a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.

b) The following standards / Equipment have been used.

i) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2021 - 1227 (JRPM, Chennai) (Cal. Date: 26/08/21, Due Date: 25/08/22)

ii) RTD (PT - 100) Cal. Certificate No. TL / 021 / 1091.2.1 (TEMPESENS, Udaipur) (Cal. Date: 25/10/21, Due Date: 24/10/22)

2.1 Result :

Thermal Calibration

| Sl. No. | Parameter/ Range | Temperature Set at °C | Measured Value on DUC °C | | | Corrected Std. Value °C | | | Error °C | Rack | Measurement Expanded Uncertainty ± °C |
|---------|-----------------------|-----------------------|--------------------------|-------|---------|-------------------------|--------|---------|----------|------|---------------------------------------|
| | | | Min. | Max. | Average | Min. | Max. | Average | | | |
| 1. | Temperature (-) 80 °C | -80.0 | -80.0 | -80.0 | -80.00 | -80.28 | -80.24 | -80.249 | 0.249 | 1 | 0.83 |
| | | | | | | -80.31 | -80.29 | -80.288 | 0.288 | 2 | 0.83 |
| | | | | | | -80.40 | -80.37 | -80.371 | 0.371 | 3 | 0.83 |
| | | | | | | -80.56 | -80.45 | -80.492 | 0.492 | 4 | 0.83 |

Remarks: i) This result has an expanded uncertainty with a coverage factor k=2 at approximately 95% confidence level.

ii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

iii) Average Reading obtained by 10 Readings.

iv) Error = Average DUC Reading - Average Standard Reading.

DUC - Device Under Calibration N.S. - Not Specified N.A. - Not Applicable Std. - Standard Min. - Minimum
Max. - Maximum

Opinions and Interpretations

| | | |
|-------------|---|--------------------------|
| Calibrated | √ | Accepted / Valid for use |
| Limited Use | | Rejected / Out of use |

Measure Techno Lab
J. Dey
Testing Engineer

J. Dey
26/02/22

Checked / Approved by:

S. Pandey
Quality & Technical Manager

S. Pandey