

Certificate of calibration

Date of calibration: 16/10/2020
Next Calibration due on: 15/10/2021

Name of the Equipment: Thermometer
Serial No. / Model No.: Mextech
Identification No.: MDACS/SPMR/ICFC/TM/01
Acceptable range: 2 to 8 degree Celsius ($^{\circ}\text{C}$)

Point of Calibration: 2°C , 4°C , 8°C

Tolerance (If Applicable): $\pm 2^{\circ}\text{C}$

Location of Calibration: At Site

Condition of Receipt Item: Good

Method of calibration: In- house calibration done using Calibrated Thermometer 2020-21

Traceability Certificate no: MA 309/ 2020-21

Calibration Result:

| Calibration Points ($^{\circ}\text{C}$) | UUC Reading ($^{\circ}\text{C}$) | | | Standard Reading ($^{\circ}\text{C}$) | Mean |
|---|------------------------------------|-----------|-----------|---|------|
| | Reading 1 | Reading 2 | Reading 3 | | |
| 2 | 2.1 | 1.9 | 2 | 2 | 2 |
| 4 | 3.9 | 4.1 | 4 | 4 | 4 |
| 8 | 8.1 | 7.9 | 8 | 8 | 8 |

Acceptable Variation $\pm 2^{\circ}\text{C}$ of the Mean Value

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

Rebut
15/10/2020
Ms. Anita Bhargava
SRL LTMMC & GH, Lion
Technical Officer, MDACS
Mumbai 400022

**State Reference Laboratory
I. C. T. C.
Department of Microbiology
L. T. M. Medical College and General Hospital Mumbai-22**

Certificate of calibration

Name of the Equipment: Thermometer

Date of calibration: 18/10/2019

Serial No. / Model No.: Mextech

Next Calibration due on: 18/10/2020

Identification No.: MDAB/SPMH/ICTC/TM/01

Acceptable range: 2 to 8 degree Celsius ($^{\circ}\text{C}$)

Point of Calibration: 2°C , 4°C , 8°C

Tolerance (If Applicable): $\pm 2^{\circ}\text{C}$

Location of Calibration: At Site

Condition of Receipt Item: Good

Method of calibration: In- house calibration done using Calibrated Thermometer 2020-21

Traceability Certificate no: MA 309/ 2020-21

Calibration Result:

| Calibration Points ($^{\circ}\text{C}$) | UUC Reading ($^{\circ}\text{C}$) | | | Standard Reading ($^{\circ}\text{C}$) | Mean |
|---|------------------------------------|-----------|-----------|---|------|
| | Reading 1 | Reading 2 | Reading 3 | | |
| 2 | 2 | 2.1 | 1.9 | 2 | 2 |
| 4 | 4 | 3.9 | 4.1 | 4 | 4 |
| 8 | 8 | 7.8 | 8.2 | 8 | 8 |

Acceptable Variation $\pm 2^{\circ}\text{C}$ of the Mean Value

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

Anita

Ms Anita Bharadwaj
Technical Officer, MDACS

SRL TMMC & GH,
SION, Mumbai 400022.