



STAR HUMAN Sciences Pvt. Ltd.

25-B/7, New Rohtrak Road, Karol Bagh, New Delhi - 110005, India

Tel. : +91-11-28721843, Mob. : 9873686675, 9810028448

E-mail : info@starhuman.in, www.starhuman.in

CERTIFICATE OF CALIBRATION

Narayan Pathology & Biopsy Centre
117/22, Sarvodya Nagar
(Opp J.L Rohtagi Eye Hospital Near Baba Bhothnath Ashram)
Kanpur-208005

This is to certify that the Instrument **Humastar 200 Biochemistry Analyzer** bearing serial no. 21191146004 manufactured by **HUMAN GERMANY** is in full working and has been calibrated to the standard specifications for the period of 09th December 2021 to 8th December 2022.

Date of calibration: 09-12-2021

This calibration certificate is valid up to 08-12-2022

For Star Human Sciences Pvt. Ltd.



(Authorized Signatory)

Technical Calibration Detail


HS200

S. no. 21191146004

ITEMS	value	Target range
<u>Lamp</u>		
mV @ 340 nm	1230	900-1500
Current mV	6.5	OK
Efficiency@340	69	54-130
<u>Pump</u>		
Pump 1-5 volume uL	In range	220-520
Pump 1-5 flow rate uL/Sec	OK	550-850

Other mechanical parameters in OS are in range as per standard criteria.

Dated-09-Dec-2021


09/12/2021
Service Engineer

For Star Human Sciences Pvt. Ltd.

Performance Qualification

HumaStar 100, HumaStar 200, HumaStar 300SR and HumaStar 600

Revision list

Revision	Date	Description	Editor
1	2018/08/14	First revision	Silvia Fischer

Lab Name

Narayan Pathology And Biopsy Centre
Add- 117/22 Sarvodaya Nagar Kanpur

Contents

1. Special materials.....	2
1.1. Customer using Human reagents	2
1.2. Customer using non Human reagents	2
2. Accompanying documents	2
3. Performance quality checks	2
4. Documentation.....	3
5. Calculation of the CV.....	3
6. Example for calculations in EXCEL	3
7. Closure	5
8. Additional notes	5

Introduction

With the Performance Qualification we assure that HumaStar 100/200/300SR/600 will work under real life conditions (in the lab of the customer) according Human specifications.

All assays the customer is supposed to use should be checked during this test. The PQ can be done during the end user training or ahead in the laboratory of the customer.

For each test calibration, QC (two levels) and precision should be measured and analysed to check if the results are in the specification of Human.

If non Human reagents are used, the Performance Qualification should be done with Human reagents according the IFU (using Human calibrator and QC material)

1. Special materials

- Wash additive (18971) – All HumaStar Systems
- Special wash solution (18974) – HS 100/200
- Special wash solution-2 (18974/2) – HS 300SR
- Cuvette Clean (16663/20) – HS 600
- 0.9% NaCl solution – HS 100/200
- Diluent (16663/10) – HS300SR/600

1.1. Customer using Human reagents

All reagents, calibrator and QC material which the customer has chosen for the routine.

1.2. Customer using non Human reagents

- AutoCal (Ref 13160)
- Serodos (Ref 13951)
- HumaTrol P (Ref 13512)
- GLUCOSE liquicolor reagent (Ref 10121 – HS 100/200; Ref 10260300 – HS 300SR/600)
- CALCIUM liquicolor reagent (Ref 10011 – HS 100/200; Ref 10011300 – HS 300SR/600)
- GOT (ASAT) IFCC mod. liquidUV (Ref 12211 – HS 100/200; Ref 12021300 – HS 300SR/600)

2. Accompanying documents

- A) User Manuals (last revision, please check Human web page)
- B) Application sheets (on the PC)
- C) Package inserts (in the kits)

3. Performance quality checks

Step	Check	Expected result	Result	Final remark
1	Maintenance (perform all user maintenance)	All maintenance tasks finish without an error	DONE	OKAY
1	Calibration	Calibration factors should be in the specified range	IN RANGE	DONE
2	Quality Control	Both QC samples should be in range given by Human	IN RANGE	DONE

Step	Check	Expected result	Result	Final remark
3	Precision (n= 20), a Qc sample can be used	The CV in % for each test should be in the specified range	In all text < cv=3.1%	OKAY

4. Documentation

Please document all results in an additional document (excel file).

5. Calculation of the CV

<p>Evaluate the results as following:</p> $\text{mean} = \sum_{i=1}^{10} \text{sample } i$ $\text{SD} = \sqrt{\frac{\sum_{i=1}^{10} (\text{sample } i - \text{mean})^2}{9}}$ $\text{CV}\% = \frac{\text{SD} * 100}{\text{mean}}$ <p>Easy to calculate in an EXCEL sheet using the appropriate formulas.</p> <p>See also example below</p>			
---	--	--	--

6. Example for calculations in EXCEL

	A	B	C	D
1		target	min.	max.
2	Serodos	8,33	7,41	9,24
3				
4	sample 1	8,4		
5	sample 2	8,3		
6	sample 3	8,5		
7	sample 4	8,0		
8	sample 5	8,2		
9	sample 6	8,2		
10	sample 7	8,4		
11	sample 8	8,2		
12	sample 9	9,3		
13	sample 10	8,4		
14	mean	8,4		
15	SD	0,35		
16	CV%	4,18		
17	single value <> 20% mean?	0		
18	mean within reference range?	yes		
19	CV% < 5%?	yes		



Cell	Function
A14	=AVERAGE(B4:B13)
A15	=STDEV(B4:B13)
A16	=B15*100/B14
A17	=COUNTIFS(B4:B13,"<"&(B14-(B14*0.2)))+COUNTIFS(B4:B13,">"&(B14+(B14*0.2)))
A18	=IF(ISNUMBER(B4)=FALSE,"",IF(B14>\$D\$2,"no",IF(B14<\$C\$2,"no","yes")))
A19	=IF(ISNUMBER(B4)=FALSE,"",IF(B16<=5,"yes","no"))

7. Closure

Study data has determined that the system described in this document meets all criteria outlined in this Performance Qualification protocol. All exceptional conditions if any have been addressed. The system is ready for specific usage.

The analyzer passed the performance quality check.

The analyzer didn't pass the performance quality check (see additional notes).

System	<input type="checkbox"/> HumaStar 100 <input checked="" type="checkbox"/> HumaStar 200 <input type="checkbox"/> HumaStar 300 SR <input type="checkbox"/> HumaStar 600	Serial Number: 21191146004 SW Version: 0.44.2.15 Setting-Database Version: 1.36
Date 30-NOV-2019	Service engineer / Application Specialist (printed name) MUBASHIR WALI	Service engineer / Application Specialist (signature) 
Date 30-NOV-2019	Lab Manager (printed name) NEELIMA SACHAN	Lab Manager (signature) 

8. Additional notes

HumaStar 100

HumaStar 200 ✓

HumaStar 300SR

Installation quality check

Revision list

Revision	Date	Description	Editor
1	2016/02/26	First revision	Mathias Kamprath
2	2018/06/20	Review and extension to HumaStar 300SR	Mathias Kamprath
3	2019/01/28	Repetition of all calibration steps added	Mathias Kamprath
4	2019/02/22	Serial number field added	Mathias Kamprath

Lab Name Narayan Pathology And Biopsy Centre
Add- 117/22 Sarvodaya Nagar Kanpur

Special tools

Volt meter or multi meter (V and mV). E. g. Human catalog number 60200224

Accompanying documents

- A) HSB_Power_Surge_Protector_&_online_UPS_incl. ground check.pdf
- B) HumaStar 100/200 Service Manual, revision 04/2015/09
- C) HumaStar 300SR Service Manual, revision 01/2018-FEB-09
- D) Touch screen monitor documentation (optional)
- E) Printer documentation (optional)

Installation quality check

Step	Check	Expected result	Result	Final remark
Main power supply				
1	Measure the grounding voltage between neutral and ground of the main power line. See document A).	Less than 2V.	1.3v	OKAY

Step	Check	Expected result	Result	Final remark
2	Connect the power surge protector socket to the main power. If the EU plug doesn't fit use either the UK or the US adapter. See document A).	Green and red LEDs are on.	SUPPLY OKAY	DONE
3	Connect the on-line UPS to one of the power protection sockets and switch on the on-line UPS. See document A).	The on-line UPS starts charging the battery. The wrong wiring alarm is off. Wait for 100% charge of the battery before switching on the analyzer.	CONNECTED	OKAY
Installation of the analyzer, PC (incl. mouse and keyboard), external barcode scanner (HumaStar 300SR) monitor (touch screen optional), printer (optional) and bottles for system solution (blue), special wash solution (green), normal waste (red) and special waste (yellow, HumaStar 300SR)				
4	Place the analyzer on a work bench. Remove all transportation protection for the: <ul style="list-style-type: none"> • top cover, • sampling arm(s), • internal barcode reader cover (HumaStar 100/200), • wash station top cover (HumaStar 200). 	Enough space (150 mm recommended) on the left, rear and right side and approx. 55-60 mm underneath. The analyzer is horizontally aligned.	ALL DONE	OKAY
5	Place the other electronic components on the right side of the analyzer. (recommended)		PLACED	DONE
6	Place the system solution, special wash solution, normal waste and special waste (HumaStar 300SR) bottles on the left side of the analyzer. (recommended)		PLACED AT BOTTOM	DONE

Step	Check	Expected result	Result	Final remark
7	Establish all connections.	As described in document B) "2.3.5 Installation" (HumaStar 100/200) or document C) "3.3.5 Installation" (HumaStar 300SR).	ESTABLISHED	DONE
8	Switch on the analyzer.	The inner plate shakes three times, the peristaltic pump 7 (and 9, HumaStar 300SR) turn for a second and the pinch valve(s) switch on and off.	SWITCHED ON	DONE
9	Connect the external barcode scanner (HumaStar 300SR) to the PC.		N/A	N/A
10	Switch on the monitor and the PC.	Log on as "Support" user. If not pre-installed, install the HI software as described in document B) "18 HI Software installation/update" (HumaStar 100/200) or document C) "8 HI Software installation" (HumaStar 300SR).	SWITCHED ON INSTALLED HI	DONE
11	Touch screen monitor and printer only.	If not pre-installed, install the driver(s) as described in the accompanying documentation D) and E).	N/A	N/A
12	Start the HI program as "Installer".	Analyzer connects to the HI software.	STARTED	DONE
13	Go to the Terminal program and repeat all calibration steps. HS100/200: 21 steps HS300SR: 39 steps	Every calibration step has to be finished successfully.	SUCCESSFULLY DONE	DONE

Closure

Serial number of the analyzer: 2 1 1 9 1 1 4 6 0 0 4

The analyzer passed the installation quality check.

The analyzer didn't pass the installation quality check.

Note the next steps to get the analyzer in condition to pass the installation quality check.

Date	Service engineer (printed name)	Service engineer (signature)
30-NOV-2019	MUBASHIR WALI	

1. Additional notes

HumaStar 100 HumaStar 200 ✓ HumaStar 300SR Operation quality check

Revision list

Revision	Date	Description	Editor
1	2016/02/26	First revision	Mathias Kamprath
2	2016/03/02	SD formula and EXCEL example changed	Mathias Kamprath
3	2018/06/20	Review and extension to HumaStar 300SR	Mathias Kamprath
4	2019/02/22	Serial number field added	Mathias Kamprath

Lab Name Narayan Pathology And Biopsy Centre
Add- 117/22 Sarvodaya Nagar Kanpur

Special materials

- Wash additive (18971)
- Special wash solution (18974)
- 0.9% NaCl solution
- Serodos control (13951)
- HumaTrol P control (13512)
- AutoCal multi-calibrator (13160)
- Glucose liquicolor reagent
- Calcium liquicolor reagent

Accompanying documents

- A) HumaStar 100/200 User Manual, revision 02/2013-03
- B) HumaStar 100/200 Service Manual, revision 04/2015/09
- C) HumaStar 300SR User Manual, revision 02/2017-11
- D) HumaStar 300SR Service Manual, revision 01/2018-FEB-09

Operation quality check

Step	Check	Expected result	Result	Final remark
1	Prepare the system solution, the special wash solution and the dilution bottles.	PREPARED	OKAY	DONE
2	Switch on the analyzer and run the start-up procedure.	No error messages. All cuvettes shown in green in HI > Maintenance > reaction cuvettes (HumaStar 100/200) or HI > Maintenance > Special > Reaction cuvettes (HumaStar 300SR).	NO MESSAGE ALL GREEN	DONE
3	Prepare the AutoCal multi-calibrator and the two controls Serodos and HumaTrol.	Place a cup of AutoCal, Serodos and HumaTrol on the sample tray.	PREPARED & PLACE	DONE
4	Prepare the Glucose and/or GOT and/or Calcium reagent(s). GLU & GOT	Place the reagent bottle(s) on the reagent tray and run the level check. All volumes have to be recognized.	PLACED & CHECKED THE LEVEL	DONE
5	Use Serodos as sample material.	Place a cup of "sample" material on the sample tray. When all tests are to be performed, better place two cups on the sample tray.	PLACED 2 CUPS	DONE
6	Generate a work list.	Per reagent the "sample" material has to be tested 10 times.	PLACED THE SAMPLE	DONE
7	Run the work list.		RUN	DONE

Step	Check	Expected result	Result	Final remark
8	<p>Evaluate the results as following:</p> $\text{mean} = \sum_{i=1}^{10} \text{sample } i$ $\text{SD} = \sqrt{\frac{\sum_{i=1}^{10} (\text{sample } i - \text{mean})^2}{9}}$ $\text{CV\%} = \frac{\text{SD} * 100}{\text{mean}}$ <p>Easy to calculate in an EXCEL sheet using the appropriate formulas.</p> <p>See also 6. Example for calculations in EXCEL.</p>	<p>The mean of all ten "sample" results has to be in the reference range of the Serodos control.</p> <p>The CV% has to be less than 5%.</p> <p>Only one "sample" result may deviate more than +/- 20% from the mean.</p> <p>The mean and the SD of the remaining nine results have to be recalculated.</p>	<p>GLU</p> <p>CV= 2.5%</p> <p>SGOT</p> <p>CV= 1.8%</p>	<p>OKAY</p>

Closure

Serial number of the analyzer: 2 1 1 9 1 1 4 6 0 0 4



The analyzer passed the operation quality check.



The analyzer didn't pass the operation quality check.

Note the next steps to get the analyzer in condition to pass the operation quality check.

Date	Service engineer (printed name)	Service engineer (signature)
30-NOV-2019	MUBASHIR WALI	

Additional notes

Example for calculations in EXCEL

	A	B	C	D
1		target	min.	max.
2	Serodos	8,33	7,41	9,24
3				
4	sample 1	8,4		
5	sample 2	8,3		
6	sample 3	8,5		
7	sample 4	8,0		
8	sample 5	8,2		
9	sample 6	8,2		
10	sample 7	8,4		
11	sample 8	8,2		
12	sample 9	9,3		
13	sample 10	8,4		
14	mean	8,4		
15	SD	0,35		
16	CV%	4,18		
17	single value <> 20% mean?	0		
18	mean within reference range?	yes		
19	CV% < 5%?	yes		

Cell	Function
A14	=AVERAGE(B4:B13)
A15	=STDEV(B4:B13)
A16	=B15*100/B14
A17	=COUNTIFS(B4:B13,"<"&(B14-(B14*0.2)))+COUNTIFS(B4:B13,">"&(B14+(B14*0.2)))
A18	=IF(ISNUMBER(B4)=FALSE,"",IF(B14>\$D\$2,"no",IF(B14<\$C\$2,"no","yes")))
A19	=IF(ISNUMBER(B4)=FALSE,"",IF(B16<=5,"yes","no"))

Training Certificate

This is to certify that

*Ms Smriti Gupta
Has been trained on*

Humastar 200

*Model of Humastar 200 Biochemistry Analyzer
The training was conducted by Product Experts*

Narayan Pathology & Biopsy Centre

117/22, Sarvodya Nagar
Opp J.L RohtagiEye Hospital Near Baba Bhoothnath Ashram
Kanpur-208005

Dated: 16-12-2019

Star Human Sciences Pvt. Ltd.
New Delhi, INDIA -110 005

Training Certificate

This is to certify that

*Mr Ranjeet Kumar
Has been trained on*

Humastar 200

*Model of Humastar 200 Biochemistry Analyzer
The training was conducted by Product Experts*

Narayan Pathology & Biopsy Centre

117/22, Sarvodya Nagar
Opp J.L RohtagiEye Hospital Near Baba Bhoothnath Ashram
Kanpur-208005

Dated: 16-12-2019

Star Human Sciences Pvt. Ltd.
New Delhi, INDIA -110 005

SAMPLE REPORT

Default laboratory
(to be replaced)

NARAYAN PATHOLOGY & BIOPSY
CENTRE

ID: serodos plus-00000007

Name:
Family Name:
Date of birth:
Department:

Reference:
Date: 09-Feb-22
Type: Control
Nature: Serum

Analysis:

Method Name	Result	Unit	Evaluation	Min	Max
Bilirubin total	3.78	mg/dl		3.59	4.13
Bilirubin direct	2.62	mg/dl		2.2	3.08
ASAT/GOT	133	U/l		124	144
ALAT/GPT	127	U/l		113	131
Alk Phos AMP IFCC	342	U/l		301	351
Total Protein	9.74	g/dl		8.88	11
Albumin IFCC	5.17	g/dl		4.63	5.97
Urea	148.96	mg/dl		153	153
Creatinine	4.40	mg/dl		3.69	6.47
Uric acid	10.35	mg/dl		8.07	11.79
Phosphorus	8.05	mg/dl		7.98	9.22
Glucose	190.0	mg/dl		180	200

Comments

Human

LG

LABORATORY CENTER
Phone: 916311022
650.00
650.00
650.00
and Fifty Units

8/0 ZWALI
EP - 1-2
PC - 2-4
RBC - Nil
WBC - Nil

Deep-
APTT - NCD
PT - NCD
Brid-
APTT - NCD
PT

16
17
18
19
20
21
22
23

my

SAMPLE REPORT

HS200 Sample report

Default laboratory
(to be replaced)

NARAYAN PATHOLOGY & BIOPSY
CENTRE

ID: serodos plus-0000007


Name:
Family Name:
Date of birth:
Department:

Reference:
Date: 23-Feb-22
Type: Control
Nature: Serum

Analysis:

Method Name	Result	Unit	Evaluation	Min	Max
Bilirubin total	3.73	mg/dl		3.59	4.13
Bilirubin direct	2.79	mg/dl		2.2	3.08
ASAT/GOT	144	U/l		124	144
ALAT/GPT	118	U/l		113	131
Alk.Phos.AMP IFCC	309	U/l		301	351
Total Protein	10.08	g/dl		8.88	11
Albumin IFCC	5.11	g/dl		4.63	5.97
Urea	148.83	mg/dl		133	153
Creatinine	4.44	mg/dl		3.65	6.47
Uric acid	9.88	mg/dl		8.07	11.79
Phosphorus	8.92	mg/dl		7.98	9.22

Comment



344
1-2
3-4
5-6
7-8
9-10
11-12
13-14
15-16
17-18
19-20
21-22
23-24
25-26
27-28
29-30
31-32
33-34
35-36
37-38
39-40
41-42
43-44
45-46
47-48
49-50
51-52
53-54
55-56
57-58
59-60
61-62
63-64
65-66
67-68
69-70
71-72
73-74
75-76
77-78
79-80
81-82
83-84
85-86
87-88
89-90
91-92
93-94
95-96
97-98
99-100

Deep-
APTT - NCD
PT - NCD
Brid -
APTT - NCD

18-Ab
19-Ab
20-Ab
21-Ab
22-Ab
23-Ab
24-Ab

REAGENTS

MONITOR

SAMPLES

WORKLIST

METHODS

SAMPLE REPORT

HS200

Find | Next

Sample report

Default laboratory
(to be replaced)NARAYAN PATHOLOGY & BIOPSY
CENTRE

ID: serodos plus-00000007

Name:

Family Name:

Date of birth:

Department:

Reference:

Date: 18-Feb-22

Type: Control

Nature: Serum

Analysis:

Method Name	Result	Unit	Evaluation	Min	Max
Bilirubin total	3.64	mg/dl		3.59	4.13
Bilirubin direct	2.47	mg/dl		2.2	3.08
ASAT/GOT	133	U/l		124	144
ALAT/GPT	122	U/l		113	131
Alk Phos AMP IFCC	335	U/l		301	351
Total Protein	9.58	g/dl		8.88	11
Albumin IFCC	5.05	g/dl		4.63	5.97
Urea	15.71	mg/dl		133	163
Creatinine	4.41	mg/dl		3.69	6.47
Uric acid	9.72	mg/dl		8.07	11.79

Comment:

Human

LG

SAMPLE REPORT

File: New

Sample report

Default laboratory
(to be replaced)

**NARAYAN PATHOLOGY & BIOPSY
CENTRE**

ID: serodos plus-00000007

Name:

Reference:

Family Name:

Date: 21-Feb-22

Date of birth:

Type: Control

Department:

Nature: Serum

Analysis:

Method Name	Result	Unit	Evaluation	Min	Max
Bilirubin total	3.73	mg/dl		3.59	4.13
Bilirubin direct	2.96	mg/dl		2.2	3.08
ASAT/GOT	142	U/L		124	144
ALAT/GPT	117	U/L		113	131
Ask. Phos. AMP /GCC	311	U/L		301	361
Total Protein	9.49	g/dl		8.88	11
Albumin /GCC	5.06	g/dl		4.63	6.97
Urea	161.59	mg/dl		133	163
Creatinine	4.57	mg/dl		3.69	6.47
Uric acid	9.75	mg/dl		8.07	11.79
Phosphorus	9.27	mg/dl		7.99	9.22
Glucose	151.9	mg/dl		186	208

Human

LG

15:01:05

Handwritten notes on a piece of paper:

- DEEP - NCD
- APT - NCD
- PT - NCD
- BRID - NCD
- APTT - NCD
- PT

Other handwritten notes and numbers:

- 18 - Feb
- 19 - Feb
- 20 - Feb
- 21 - Feb
- 22 - Feb
- 23 - Feb
- 24 - Feb
- 344
- 2.77
- 1.8
- 9.0
- 9.01 - 20
- 9.01 - 20.5
- 612
- 131
- 3.0
- 4.2
- Kalpang
- 13
- 2.83

SAMPLE REPORT

Find | Next

HS200

Sample report

Default laboratory (to be replaced)

NARAYAN PATHOLOGY & BIOPSY CENTRE

ID:

Name:

Reference:

Family Name:

Date:

Date of birth:

Type:

Department:


Nature:

Analysis:

Method Name	Result	Unit	Evaluation	Min	Max
Bilirubin total	3.60	mg/dl		3.59	4.13
Bilirubin direct	2.58	mg/dl		2.2	3.08
ASAT/GOT	136	U/l		124	144
ALAT/GPT	118	U/l		113	131
Alk.Phos AMP IFCC	317	U/l		301	351
Total Protein	8.94	g/dl		8.88	11
Albumin IFCC	5.02	g/dl		4.63	5.97
Urea	134.76	mg/dl		133	153
Creatinine	4.24	mg/dl		3.69	6.47
Unc acid	9.94	mg/dl		8.07	11.79
Phosphorus	8.79	mg/dl		7.98	9.22
Glucose	200.0	mg/dl		186	208

Humana

LG





REAGENTS



MONITOR



SAMPLES



WORKLIST



METHODS



Save

Load

Clear



Worksheets

Stack sheets

001

calibration 19-02-2022

QC 19-02-2022

sample

Remove test

Execute test

Inspect test

Recalculate test

Test n°	Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
80142	Autocal-00000017	S1	1	CreaA	Completed	0.0921	0.0000	0.0910	2022-02-19 11:03
80143	Autocal-00000017	S2	1	CreaA	Completed	0.0909	0.0000	0.0898	2022-02-19 11:03
80144	Autocal-00000017	S3	1	CreaA	Completed	0.0931	0.0000	0.0920	2022-02-19 11:03
80145	Autocal-00000017	S1	1	Glu	Completed	0.5131	0.0284	0.4745	2022-02-19 11:04
80146	Autocal-00000017	S2	1	Glu	Completed	0.3748	-0.0714	0.4360	2022-02-19 11:05
80147	Autocal-00000017	S3	1	Glu	Completed	0.4819	0.0267	0.4450	2022-02-19 11:05
80148	Autocal-00000017	S1	1	GOT	Completed	-0.0521	0.0000	-0.0518	2022-02-19 11:06
80149	Autocal-00000017	S2	1	GOT	Completed	-0.0548	0.0000	-0.0545	2022-02-19 11:07
80150	Autocal-00000017	S3	1	GOT	Completed	-0.0538	0.0000	-0.0535	2022-02-19 11:08
80151	Autocal-00000017	S1	1	GPT	Completed	-0.0582	-0.0009	-0.0570	2022-02-19 11:08
80152	Autocal-00000017	S2	1	GPT	Completed	-0.0593	-0.0008	-0.0582	2022-02-19 11:09
80153	Autocal-00000017	S3	1	GPT	Completed	-0.0574	-0.0005	-0.0566	2022-02-19 11:10
80154	Autocal-00000017	S1	1	Phos	Completed	0.6938	0.0161	0.5923	2022-02-19 11:10
80218	Autocal-00000017	S2	1	Phos	Completed	0.6234	0.0147	0.5233	2022-02-19 11:42
80156	Autocal-00000017	S3	1	Phos	Completed	0.6764	0.0158	0.5752	2022-02-19 11:12
80225	Autocal-00000017	S1	1	Prot	Completed	0.4112	0.0340	0.3055	2022-02-19 12:15
80158	Autocal-00000017	S2	1	Prot	Completed	0.4184	0.0618	0.2849	2022-02-19 11:12

Test

serodos pl
 40806 juli
 40804 anja
 geeta
 ramraj
 40822 shya
 40813 kama
 40793 ramra
 40823
 40827 sarita

Method gr


- | Clin1 | Clin2 |
|--|-------------------------------------|
| <input checked="" type="checkbox"/> Alb | <input checked="" type="checkbox"/> |
| <input checked="" type="checkbox"/> CA | <input checked="" type="checkbox"/> |
| <input checked="" type="checkbox"/> GGT | <input checked="" type="checkbox"/> |
| <input checked="" type="checkbox"/> GOTM | <input checked="" type="checkbox"/> |
| <input checked="" type="checkbox"/> Phos | <input checked="" type="checkbox"/> |
| <input checked="" type="checkbox"/> UA | <input checked="" type="checkbox"/> |



Human



Save Load Clear



Worksheets

Stack sheets

001 calibration 19-02-2022 QC 19-02-2022 sample

Remove test Execute test Inspect test Recalculate test

Test n°	Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
80225	Autocal-00000017	S1	1	Prot	Completed	0.4112	0.0340	0.3055	2022-02-19 12:15
80158	Autocal-00000017	S2	1	Prot	Completed	0.4184	0.0618	0.2849	2022-02-19 11:12
80159	Autocal-00000017	S3	1	Prot	Completed	0.4252	0.0644	0.2891	2022-02-19 11:13
80160	Autocal-00000017	S1	1	Trig	Completed	0.5815	0.0040	0.5099	2022-02-19 11:14
80161	Autocal-00000017	S2	1	Trig	Completed	0.5805	0.0032	0.5097	2022-02-19 11:14
80162	Autocal-00000017	S3	1	Trig	Completed	0.5655	0.0049	0.4930	2022-02-19 11:15
80163	Autocal-00000017	S1	1	UA	Completed	0.2185	0.0343	0.1630	2022-02-19 11:15
80164	Autocal-00000017	S2	1	UA	Completed	0.2082	0.0247	0.1623	2022-02-19 11:15
80165	Autocal-00000017	S3	1	UA	Completed	0.2131	0.0262	0.1657	2022-02-19 11:16
80166	Autocal-00000017	S1	1	UreaUV	Completed	-0.1033	0.0000	-0.0981	2022-02-19 11:16
80167	Autocal-00000017	S2	1	UreaUV	Completed	-0.0974	0.0000	-0.0922	2022-02-19 11:16
80223	Autocal-00000017	S3	1	UreaUV	Completed	-0.1120	0.0000	-0.1068	2022-02-19 11:59
80195	geeta	U	1	Alb	Completed	0.9083	0.0973	2.77	2022-02-19 11:25
80196	geeta	U	1	Glu	Completed	0.6282	0.0601	243.2	2022-02-19 11:25
80197	geeta	U	1	Prot	Completed	0.2624	0.0519	4.31	2022-02-19 11:25
80199	ramraj	N	1	Glu	Completed	0.3903	0.0582	140.3	2022-02-19 11:33
80200	ramraj	N	1	Prot	Completed	0.1565	0.0541	0.95	2022-02-19 11:45



Human



Save

Load

Clear

Worksheets

Stack sheets

001

001

001

001

001

001

001

001

001

001

001

001

Remove test

Execute test

Inspect test

Recalculate
test

Test n°	Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date	
82010	42605 yashmeen	N	1	APAMP	Completed	0.0236	0.0000	90	2022-03-01 07:12	
82011	42605 yashmeen	N	1	Prot	Completed	0.3642	0.0359	8.08	2022-03-01 07:13	
82012	42605 yashmeen	N	1	Alb	Completed	1.1239	0.1182	3.82	2022-03-01 07:13	
82013	42605 yashmeen	N	1	CreaA	Completed	0.0212	0.0000	0.89	2022-03-01 07:13	
82014	serodos plus-00000007	QC2	1	Bilta	Completed	0.0760	0.0049	3.68	2022-03-01 07:19	
82015	serodos plus-00000007	QC2	1	Bilda	Completed	0.3031	0.2271	2.25	2022-03-01 07:20	
82016	serodos plus-00000007	QC2	1	GOT	Completed	-0.0514	0.0000	132	2022-03-01 07:20	
82017	serodos plus-00000007	QC2	1	GPT	Completed	-0.0585	-0.0014	119	2022-03-01 07:20	
82018	serodos plus-00000007	QC2	1	APAMP	Completed	0.0618	0.0000	335	2022-03-01 07:21	
82019	serodos plus-00000007	QC2	1	Prot	Completed	0.4289	0.0606	9.32	2022-03-01 07:21	
82020	serodos plus-00000007	QC2	1	Alb	Completed	1.4938	0.1597	5.49	2022-03-01 07:21	
82021	serodos plus-00000007	QC2	1	UreaUV	Completed	-0.1062	0.0000	145.26	2022-03-01 07:22	
82022	serodos plus-00000007	QC2	1	CreaA	Completed	0.0959	0.0000	5.10	2022-03-01 07:22	
82023	serodos plus-00000007	QC2	1	UA	Completed	0.2201	0.0192	11.08	2022-03-01 07:24	
82024	serodos plus-00000007	QC2	1	Phos	Completed	0.6460	0.0108	8.13	2022-03-01 07:24	
82025	serodos plus-00000007	QC2	1	Glu	Completed	0.4932	0.0272	200.4	2022-03-01 07:25	
82026	jaibun	N	1	Alb	Completed	0.7404	0.0836	2.04	2022-03-01 07:25	

Nature: Serum

Cup: n/a

Scheduled test: 0

Programmed test: 36

Patient:

Remove test

Execute test

Inspect test

Recalculate test

Nº	Meth. code	Dilution	OD1	OD2	Result	Test status	Date
82214	Bilta	1	0.0741	0.0043	3.61	Completed	2022-03-02 07:23
82215	Bilda	1	0.2708	0.1879	2.22	Completed	2022-03-02 06:58
82216	GOT	1	-0.0514	0.0000	130	Completed	2022-03-02 06:59
82217	GPT	1	-0.0605	-0.0014	121	Completed	2022-03-02 06:59
82218	APAMP	1	0.0582	0.0000	315	Completed	2022-03-02 06:59
82219	Prot	1	0.4139	0.0488	9.19	Completed	2022-03-02 07:00
82220	Alb	1	1.4822	0.1546	5.48	Completed	2022-03-02 07:00
82221	UreaUV	1	-0.1149	0.0000	142.82	Completed	2022-03-02 07:01
82222	CreaA	1	0.1031	0.0000	5.48	Completed	2022-03-02 07:01
82223	UA	1	0.2264	0.0208	11.45	Completed	2022-03-02 07:02
82276	Trig	1	0.2855	0.0032	244	Completed	2022-03-02 11:07

Human

Worksheets

Stack sheets

001 001 001 001 001

Remove test

Execute test

Inspect test

Recalculate test

Test n°	Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
82437	42934	N	1	Glu	Completed	0.3108	0.0142	127.3	2022-03-03 06:56
82439	serodos plus-00000007	QC2	1	Bilta	Completed	0.0888	0.0152	3.67	2022-03-03 07:08
82440	serodos plus-00000007	QC2	1	Bilda	Completed	0.3103	0.2009	2.40	2022-03-03 07:08
82441	serodos plus-00000007	QC2	1	GOT	Completed	-0.0487	0.0000	126	2022-03-03 07:09
82442	serodos plus-00000007	QC2	1	GPT	Completed	-0.0558	-0.0007	113	2022-03-03 07:09
82443	serodos plus-00000007	QC2	1	APAMP	Completed	0.0494	0.0000	310	2022-03-03 07:09
82444	serodos plus-00000007	QC2	1	Prot	Completed	0.4191	0.0450	9.58	2022-03-03 07:10
82445	serodos plus-00000007	QC2	1	Alb	Completed	1.4775	0.1595	5.40	2022-03-03 07:10
82446	serodos plus-00000007	QC2	1	UreaUV	Completed	-0.1032	0.0000	140.79	2022-03-03 07:11
82447	serodos plus-00000007	QC2	1	CreaA	Completed	0.0992	0.0000	5.32	2022-03-03 07:13
82448	serodos plus-00000007	QC2	1	UA	Completed	0.2191	0.0215	10.85	2022-03-03 07:13
82449	serodos plus-00000007	QC2	1	Phos	Completed	0.6516	0.0122	8.38	2022-03-03 07:13
82450	serodos plus-00000007	QC2	1	Glu	Completed	0.5522	0.0375	202.5	2022-03-03 07:13
82451	Diluent	BLANK	1	Trig	Completed	0.0606	0.0040	0.0566	2022-03-03 07:08
82452	serodos plus-00000007	QC2	1	Trig	Completed	0.2768	0.0085	221	2022-03-03 07:14
82453	Diluent	BLANK	1	GOT	Completed	-0.0003	1.4951	-0.0003	2022-03-03 07:22
82455	42956 radha	N	1	CreaA	Completed	0.0167	0.0000	0.67	2022-03-03 08:00

Test

Calibration

Sample list

ID	Pos.	M
Autocal-00000018	59	S
serodos plus-00000007	58	S
42980 shubham	12	S
saryu	13	S
saurabh	14	S
42991	15	S
42988	16	S
42989 r.k	17	S
42986 pramila	18	S
42999 narrendra	19	S

Method groups

Method

Clin1	Clin2	Turbi
<input checked="" type="checkbox"/> Alb	<input checked="" type="checkbox"/> APAMP	<input checked="" type="checkbox"/> Bilda
<input checked="" type="checkbox"/> CA	<input checked="" type="checkbox"/> Chol	<input checked="" type="checkbox"/> CHOLE
<input checked="" type="checkbox"/> GGT	<input checked="" type="checkbox"/> GGTER	<input checked="" type="checkbox"/> Glu
<input checked="" type="checkbox"/> GOTM	<input checked="" type="checkbox"/> GPT	<input checked="" type="checkbox"/> HDL
<input checked="" type="checkbox"/> Phos	<input checked="" type="checkbox"/> Prot	<input checked="" type="checkbox"/> Trig
<input checked="" type="checkbox"/> UA	<input checked="" type="checkbox"/> UreaUV	

Human

Save

Load

Clear

Worksheets

Stack sheets

001

001

001

001

Remove test

Execute test

Inspect test

Recalculate
test

Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
Diluent	BLANK	1	Glu	Completed	0.0057	-0.0007	0.0064	2022-03-04 05:35
serodos plus-00000007	QC2	1	Bilta	Completed	0.0805	0.0095	3.67	2022-03-04 05:36
serodos plus-00000007	QC2	1	Bilda	Completed	0.3081	0.1970	2.44	2022-03-04 05:37
serodos plus-00000007	QC2	1	GOT	Completed	-0.0496	0.0000	129	2022-03-04 05:38
serodos plus-00000007	QC2	1	GPT	Completed	-0.0562	-0.0010	113	2022-03-04 05:38
serodos plus-00000007	QC2	1	APAMP	Completed	0.0647	0.0000	351	2022-03-04 05:39
serodos plus-00000007	QC2	1	Prot	Completed	0.4303	0.0654	8.95	2022-03-04 05:39
serodos plus-00000007	QC2	1	Alb	Completed	1.4396	0.1554	5.26	2022-03-04 05:39
serodos plus-00000007	QC2	1	UreaUV	Completed	-0.1037	0.0000	142.09	2022-03-04 05:40
serodos plus-00000007	QC2	1	CreaA	Completed	0.0881	0.0000	4.69	2022-03-04 05:41
serodos plus-00000007	QC2	1	UA	Completed	0.2156	0.0254	10.18	2022-03-04 05:42
serodos plus-00000007	QC2	1	Phos	Completed	0.6806	0.0139	8.85	2022-03-04 05:42
serodos plus-00000007	QC2	1	Glu	Completed	0.5225	0.0649	196.7	2022-03-04 05:43
43119 saroj	N	1	Bilta	Completed	0.0167	0.0072	0.39	2022-03-04 05:43

Test Calibration

Remove test Execute test Inspect test Recalculate test

Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
43317 chandan	N	1	Glu	Completed	0.1227	0.0067	46.8	2022-03-05 06:44
anupriya	N	1	Glu	Completed	0.3307	0.0553	116.3	2022-03-05 06:44
43328	N	1	Glu	Completed	0.7729	-0.0173	340.7	2022-03-05 06:51
43326 p.l	N	1	CreaA	Completed	0.0217	0.0000	1.00	2022-03-05 06:51
erodos plus-00000007	QC2	1	Bilta	Completed	0.1188	0.0124	3.79	2022-03-05 06:51
erodos plus-00000007	QC2	1	Bilda	Completed	0.3900	0.2155	2.39	2022-03-05 06:52
erodos plus-00000007	QC2	1	GOT	Completed	-0.0510	0.0000	130	2022-03-05 06:52
erodos plus-00000007	QC2	1	GPT	Completed	-0.0592	-0.0015	118	2022-03-05 06:52
erodos plus-00000007	QC2	1	APAMP	Completed	0.0633	0.0000	343	2022-03-05 06:53
erodos plus-00000007	QC2	1	Prot	Completed	0.4346	0.0626	9.40	2022-03-05 06:53
erodos plus-00000007	QC2	1	Alb	Completed	1.5075	0.1636	5.52	2022-03-05 06:54
erodos plus-00000007	QC2	1	UreaUV	Completed	-0.1011	0.0000	138.48	2022-03-05 06:56
erodos plus-00000007	QC2	1	CreaA	Completed	0.0886	0.0000	4.76	2022-03-05 06:56
erodos plus-00000007	QC2	1	UA	Completed	0.2246	0.0327	10.53	2022-03-05 06:57
erodos plus-00000007	QC2	1	Phos	Completed	0.6969	0.0178	9.13	2022-03-05 06:57
erodos plus-00000007	QC2	1	Glu	Completed	0.5436	0.0381	198.8	2022-03-05 06:57
soni	N	1	Alb	Completed	1.1363	0.1189	3.85	2022-03-05 06:57

Sample list

ID	Pos
Autocal-00000018	59
serodos plus-00000007	58
42980 shubham	12
saryu	13
saurabh	14
42991	15
42988	16
42989 r.k	17
42986 pramila	18
42999 narrendra	19

Method groups

Clin1	Clin2	Turbi
<input checked="" type="checkbox"/> Alb	<input checked="" type="checkbox"/> APAMP	<input checked="" type="checkbox"/> Bil
<input checked="" type="checkbox"/> CA	<input checked="" type="checkbox"/> Chol	<input checked="" type="checkbox"/> CHC
<input checked="" type="checkbox"/> GGT	<input checked="" type="checkbox"/> GGTER	<input checked="" type="checkbox"/> G
<input checked="" type="checkbox"/> GOTM	<input checked="" type="checkbox"/> GPT	<input checked="" type="checkbox"/> H
<input checked="" type="checkbox"/> Phos	<input checked="" type="checkbox"/> Prot	<input checked="" type="checkbox"/> Tr
<input checked="" type="checkbox"/> UA	<input checked="" type="checkbox"/> UreaUV	

Human

Remove test

Execute test

Inspect test

Recalculate test

Nº	Meth. code	Dilution	OD1	OD2	Result	Test status	Date
3262	Bilta	1	0.1162	0.0121	3.70	Completed	2022-03-06 09:12
3263	Bilda	1	0.3483	0.1846	2.24	Completed	2022-03-06 09:12
3264	GOT	1	-0.0506	0.0000	131	Completed	2022-03-06 09:13
3266	APAMP	1	0.0650	0.0000	351	Completed	2022-03-06 09:14
3267	Prot	1	0.4241	0.0602	9.15	Completed	2022-03-06 09:14
3268	Alb	1	1.4674	0.1581	5.33	Completed	2022-03-06 09:15
3269	UreaUV	1	-0.1106	0.0000	137.23	Completed	2022-03-06 09:15
3270	CreaA	1	0.0962	0.0000	5.22	Completed	2022-03-06 09:16
3271	UA	1	0.2126	0.0269	10.28	Completed	2022-03-06 09:17
3272	Phos	1	0.7010	0.0207	9.06	Completed	2022-03-06 09:17
3273	Glu	1	0.5334	0.0347	196.1	Completed	2022-03-06 09:11
3295	GPT	1	-0.0571	-0.0001	119	Completed	2022-03-06 10:27

Remove test

Execute test


Inspect test

Recalculate test

N°	Meth. code	Dilution	OD1	OD2	Result	Test status	Date
83380	CreaA	1	0.0962	0.0000	5.23	Completed	2022-03-07 12:13
83381	UA	1	0.2163	0.0289	10.18	Completed	2022-03-07 12:13
83383	Bilta	1	0.1162	0.0155	3.59	Completed	2022-03-07 12:14
83384	Bilda	1	0.3378	0.1793	2.48	Completed	2022-03-07 12:29
83385	GOT	1	-0.0487	0.0000	125	Completed	2022-03-07 12:14
83386	GPT	1	-0.0562	0.0000	116	Completed	2022-03-07 12:14
83387	APAMP	1	0.0639	0.0000	345	Completed	2022-03-07 12:15
83388	Prot	1	0.4101	0.0374	9.43	Completed	2022-03-07 12:15
83389	Alb	1	1.4774	0.1614	5.37	Completed	2022-03-07 12:16
83390	Glu	1	0.5311	0.0361	195.8	Completed	2022-03-07 12:16
83395	UreaUV	1	-0.0967	0.0000	144.70	Completed	2022-03-07 12:32
83362	Bilta	1	0.1162	0.0155	3.59	Completed	2022-03-06 09:13

Human

8	Diluent	BLANK	1	Glu	Completed	0.0074	0.0043	0.0031	2022-03-08	06:16
9	serodos plus-00000007	QC2	1	Bilta	Completed	0.1224	0.0240	3.75	2022-03-08	06:18
9	serodos plus-00000007	QC2	1	Bilda	Completed	0.3992	0.2244	2.72	2022-03-08	06:19
0	serodos plus-00000007	QC2	1	GOT	Completed	-0.0511	0.0000	130	2022-03-08	06:19
1	serodos plus-00000007	QC2	1	GPT	Completed	-0.0568	-0.0007	115	2022-03-08	06:20
2	serodos plus-00000007	QC2	1	APAMP	Completed	0.0636	0.0000	339	2022-03-08	06:20
3	serodos plus-00000007	QC2	1	Prot	Completed	0.4275	0.0642	9.18	2022-03-08	06:21
4	serodos plus-00000007	QC2	1	Alb	Completed	1.4635	0.1619	5.33	2022-03-08	06:21
5	serodos plus-00000007	QC2	1	UreaUV	Completed	-0.1112	0.0000	138.66	2022-03-08	06:23
6	serodos plus-00000007	QC2	1	CreaA	Completed	0.0914	0.0000	4.91	2022-03-08	06:24
7	serodos plus-00000007	QC2	1	UA	Completed	0.2218	0.0330	10.31	2022-03-08	06:24
8	serodos plus-00000007	QC2	1	Phos	Completed	0.7314	0.0204	9.17	2022-03-08	06:25
9	serodos plus-00000007	QC2	1	Glu	Completed	0.5450	0.0439	199.8	2022-03-08	06:25
30	Diluent	BLANK	1	Chol	Completed	0.0475	0.0082	0.0393	2022-03-08	06:17



HUM



001

001

001

001

001

001

001

001

001

001

001

001

Remove test

Execute test

Inspect test

Recalculate
test

Test n°	Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
83635	serodos plus-00000007	QC2	1	Bilta	Completed	0.1138	0.0114	3.75	2022-03-09 12:58
83636	serodos plus-00000007	QC2	1	Bilda	Completed	0.3507	0.1887	2.52	2022-03-09 12:58
83637	serodos plus-00000007	QC2	1	GOT	Completed	-0.0483	0.0000	126	2022-03-09 12:59
83638	serodos plus-00000007	QC2	1	GPT	Completed	-0.0562	-0.0001	115	2022-03-09 13:00
83639	serodos plus-00000007	QC2	1	APAMP	Completed	0.0598	0.0000	323	2022-03-09 13:00
83640	serodos plus-00000007	QC2	1	Prot	Completed	0.4274	0.0523	9.51	2022-03-09 13:01
83641	serodos plus-00000007	QC2	1	Alb	Completed	1.4669	0.1601	5.30	2022-03-09 13:01
83642	serodos plus-00000007	QC2	1	UreaUV	Completed	-0.0957	0.0000	143.90	2022-03-09 13:02
83643	serodos plus-00000007	QC2	1	CreaA	Completed	0.0858	0.0000	4.68	2022-03-09 13:03
83644	serodos plus-00000007	QC2	1	UA	Completed	0.2100	0.0269	9.95	2022-03-09 13:03
83646	serodos plus-00000007	QC2	1	Glu	Completed	0.4555	0.0249	205.3	2022-03-09 13:04
83648	serodos plus-00000007	QC2	1	Chol	Completed	0.3182	-0.0859	291	2022-03-09 13:04
83650	serodos plus-00000007	QC2	1	Trig	Completed	0.1707	-0.0834	230	2022-03-09 13:04
83651	43692 arjun	N	1	Glu	Completed	0.2604	0.0066	97.8	2022-03-09 13:05
83652	43692 arjun	N	1	CreaA	Completed	0.0515	0.0000	2.75	2022-03-09 13:05
83653	43682 prem	N	1	Glu	Completed	0.5797	0.0200	220.5	2022-03-09 13:06

83668	serodos plus-00000007	QC2	1	Glu	Completed	0.0234	0.0183	0.0051	2022-03-10	13:48
83669	serodos plus-00000007	QC2	1	Bilta	Completed	0.1171	0.0171	3.90	2022-03-10	13:49
83670	serodos plus-00000007	QC2	1	Bilda	Completed	0.3691	0.2056	2.54	2022-03-10	13:49
83671	serodos plus-00000007	QC2	1	GOT	Completed	-0.0510	0.0000	133	2022-03-10	13:49
83672	serodos plus-00000007	QC2	1	GPT	Completed	-0.0582	-0.0004	120	2022-03-10	13:50
83673	serodos plus-00000007	QC2	1	APAMP	Completed	0.0660	0.0000	324	2022-03-10	13:50
83674	serodos plus-00000007	QC2	1	Prot	Completed	0.4145	0.0354	9.73	2022-03-10	13:50
83675	serodos plus-00000007	QC2	1	Alb	Completed	1.5117	0.1644	5.49	2022-03-10	13:51
83676	serodos plus-00000007	QC2	1	UreaUV	Completed	-0.1098	0.0000	135.41	2022-03-10	13:51
83677	serodos plus-00000007	QC2	1	CreaA	Completed	0.0995	0.0000	5.42	2022-03-10	13:51
83678	serodos plus-00000007	QC2	1	UA	Completed	0.2202	0.0297	10.41	2022-03-10	13:53
83679	serodos plus-00000007	QC2	1	Phos	Completed	0.6793	0.0166	8.70	2022-03-10	13:54
83680	serodos plus-00000007	QC2	1	Glu	Completed	0.5455	0.0361	202.3	2022-03-10	13:54
83681	43700	N	1	CreaA	Completed	0.0287	0.0000	1.44	2022-03-10	14:01



Human



001 001 001 001 001 001 001 001 001 001 001 001

Remove test

Execute test

Inspect test

Recalculate
test

Test n°	Sample ID	Type	Dil.	Method	Test status	OD1	OD2	Result	Date
83727	Diluent	BLANK	1	Phos	Completed	0.1001	0.0004	0.0997	2022-03-11 15:59
83728	Diluent	BLANK	1	Glu	Completed	0.0056	-0.0016	0.0072	2022-03-11 15:59
83729	serodos plus-00000007	QC2	1	Bilta	Completed	0.1165	0.0180	3.98	2022-03-11 16:00
83730	serodos plus-00000007	QC2	1	Bilda	Completed	0.3820	0.2204	2.53	2022-03-11 16:00
83731	serodos plus-00000007	QC2	1	GOT	Completed	-0.0515	0.0000	135	2022-03-11 16:01
83732	serodos plus-00000007	QC2	1	GPT	Completed	-0.0585	0.0000	119	2022-03-11 16:01
83733	serodos plus-00000007	QC2	1	APAMP	Completed	0.0644	0.0000	315	2022-03-11 16:31
83734	serodos plus-00000007	QC2	1	Prot	Completed	0.4189	0.0350	9.80	2022-03-11 16:02
83735	serodos plus-00000007	QC2	1	Alb	Completed	1.5485	0.1673	5.69	2022-03-11 16:03
83736	serodos plus-00000007	QC2	1	UreaUV	Completed	-0.1155	0.0000	144.12	2022-03-11 16:31
83737	serodos plus-00000007	QC2	1	CreaA	Completed	0.0929	0.0000	5.03	2022-03-11 16:31
83738	serodos plus-00000007	QC2	1	UA	Completed	0.2192	0.0276	10.51	2022-03-11 16:32
83739	serodos plus-00000007	QC2	1	Phos	Completed	0.6847	0.0166	8.76	2022-03-11 16:32
83740	serodos plus-00000007	QC2	1	Glu	Completed	0.5411	0.0327	201.0	2022-03-11 16:32
83741	muzibul	N	1	Alb	Not scheduled			n/a	
83742	muzibul	N	1	Glu	Not scheduled			n/a	
83743	muzibul	N	1	Prot	Not scheduled			n/a	

Remove test


Execute test

Inspect test

Recalculate test

Nº	Meth. code	Dilution	OD1	OD2	Result	Test status	Date
83851	Bilta	1	0.1182	0.0210	3.90	Completed	2022-03-12 11:18
83852	Bilda	1	0.3683	0.2096	2.50	Completed	2022-03-12 11:18
83853	GOT	1	-0.0493	0.0000	129	Completed	2022-03-12 11:19
83854	GPT	1	-0.0538	-0.0010	110	Completed	2022-03-12 11:20
83855	APAMP	1	0.0608	0.0000	297	Completed	2022-03-12 11:20
83856	Prot	1	0.4104	0.0531	8.92	Completed	2022-03-12 11:20
83857	Alb	1	1.4526	0.1584	5.27	Completed	2022-03-12 11:21
83858	UreaUV	1	-0.1130	0.0000	135.80	Completed	2022-03-12 11:22
83859	CreaA	1	0.0942	0.0000	5.11	Completed	2022-03-12 11:22
83861	UA	1	0.2158	0.0283	10.29	Completed	2022-03-12 11:23
83863	Phos	1	0.3097	0.0267	0.64	Completed	2022-03-12 11:24
83865	Glu	1	0.4921	0.0300	182.2	Completed	2022-03-12 11:24

Activate Windows
Go to Settings to activate Windows



OK



SAMPLE INSPECTION

Information

ID Serodios plus 00000
 Reference _____
 Department _____
 Type Control _____
 Nature Serum _____
 Patient _____

Sample status Completed
 Position 58
 Date of insertion 13-Mar-22 11:08 AM
 Urgent No
 Container Cup
 Scheduled test 0
 Programmed test 12



Remove test

Execute test

Inspect test

Recalculate test

Nº	Meth. code	Dilution	OD1	OD2	Result	Test status	Date
84047	Bilta	1	0.1190	0.0214	3.93	Completed	2022-03-13 11:26
84048	Bilda	1	0.3697	0.2064	2.55	Completed	2022-03-13 11:26
84049	GOT	1	-0.0494	0.0000	128	Completed	2022-03-13 11:26
84049	GOT	1	-0.0563	-0.0003	115	Completed	2022-03-13 11:27
84050	GPT	1	0.0692	0.0000	337	Completed	2022-03-13 11:27
84051	APAMP	1	0.4365	0.0562	9.74	Completed	2022-03-13 11:28
84052	Prot	1	0.4365	0.0562	5.30	Completed	2022-03-13 11:29
84052	Prot	1	1.4593	0.1604	5.16	Completed	2022-03-13 11:31
84053	Alb	1	0.0964	0.0000	10.33	Completed	2022-03-13 11:31
84055	CreaA	1	0.2188	0.0319	195.7	Completed	2022-03-13 11:32
84056	UA	1	0.5304	0.0366	8.47	Completed	2022-03-13 12:22
84058	Glu	1	0.8117	0.0197	137.49	Completed	2022-03-13 12:22
84065	Phos	1	-0.1115	0.0000			
84066	UreaUV	1					



OK

Sample list
 I-00000018
 arti
 alzah
 arjun
 nisha
 s plus-00000007
 kamruddin
 radha
 anwari

M.P.C.C.
 Manager: 7004255100
 Manager: 8370001002
 Manager: 7004027100
 Manager: 7004217000
 Manager: 8730201713
 800222000
 800222000

Reagent name / Reagenzbezeichnung / Parámetro / Paramètre Method / Methode / Método/ Méthode	applicable for / anwendbar für / applicable pour / applicable pour	SI Unit SI Einheit Unitad SI Unité SI	Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs	Unit Einheit Unitad Unité	Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs
GOT (ASAT) IFCC mod. IliqUV Enzymatic, 37°C, IFCC without PSP	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.23	2.10 - 2.38	U/l	134	126 - 143
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.05	1.92 - 2.18	U/l	123	115 - 133
GPT (ALAT) IFCC mod. IliqUV Enzymatic, 37°C, IFCC without PSP	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.08	1.82 - 2.33	U/l	125	109 - 140
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.05	1.88 - 2.20	U/l	123	113 - 132
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	1.98	1.77 - 2.20	U/l	119	106 - 132
HDL CHOLESTEROL Iliqicolor Homogenous enzymatic assay	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	3.15	2.69 - 3.62	mg/dl	101	86.8 - 116
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.61	2.24 - 3.00	mg/dl	107	94.1 - 120
HDL CHOLESTEROL Precipitation method	HumaStar analyzer	mmol/l	3.31	2.92 - 3.72	mg/dl	128	113 - 144
	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	52.1	37.8 - 66.2	µg/dl	291	211 - 370
IRON Iliqicolor LAB	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	50.8	46.2 - 55.7	µg/dl	284	258 - 311
	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	39.9	27.2 - 52.6	µg/dl	223	152 - 294
IRON TPTZ Iliqicolor TPTZ	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	54.4	49.9 - 59.1	µg/dl	304	279 - 330
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	53.2	49.9 - 56.6	µg/dl	297	279 - 326
IMMUNOGLOBULINS direct IgA Immunoturbidimetry	HumaStar analyzer, Humalyzer photometer	g/l	2.58	2.28 - 2.88	mg/dl	258	228 - 288
	HumaStar analyzer, Humalyzer photometer	g/l	12.3	9.83 - 14.8	mg/dl	1234	983 - 1484
IMMUNOGLOBULINS direct IgG Immunoturbidimetry	HumaStar 600, Humalyzer 4000	g/l	0.86	0.78 - 0.93	mg/dl	85.6	78.2 - 93.0
	HumaStar 100/200/300SR, Humalyzer 2000/3000	g/l	1.16	0.90 - 1.43	mg/dl	116	89.7 - 143
LDH SCE mod. IliqUV Substrate Pyruvate, 37°C, SCE	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	11.6	10.4 - 12.8	U/l	656	625 - 767
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	11.3	10.3 - 12.2	U/l	676	619 - 733
LDL CHOLESTEROL Iliqicolor Homogenous enzymatic assay	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	4.22	2.82 - 5.64	mg/dl	163	109 - 218
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	4.58	3.05 - 6.10	mg/dl	177	118 - 236
LIPASE Iliqicolor Substrate Methylresorufin, 37°C	LOT 21001 and higher: AUTOCAL LOT 0017	µkat/l	1.39	1.25 - 1.53	U/l	83.5	75.0 - 92.0
	LOT 21001 and higher: AUTOCAL LOT 0018	µkat/l	1.21	1.05 - 1.37	U/l	72.5	62.9 - 82.1
MAGNESIUM Iliqicolor Xylylidiol blue	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	1.30	1.16 - 1.43	mg/dl	3.15	2.83 - 3.47
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	1.24	1.07 - 1.41	mg/dl	3.01	2.59 - 3.43
PANCREAS-AMYLASE Iliqicolor EPS-G7, 37°C	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.98	2.80 - 3.15	U/l	179	168 - 189
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.57	2.35 - 2.77	U/l	154	141 - 166
PHOSPHORUS Iliqicolor Molybdate (UV)	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	2.78	2.62 - 2.94	mg/dl	8.61	8.10 - 9.11
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.85	2.63 - 3.07	mg/dl	8.83	8.14 - 9.52
Potassium ISE direct	HumaStar 600, Humalyte	mmol/l	2.77	2.44 - 3.11	mg/dl	8.59	7.55 - 9.62
POTASSIUM IliqUV Enzymatic	HumaStar analyzer, Humalyzer photometer	mmol/l	6.30	5.91 - 6.69	mval/l	6.30	5.91 - 6.69
SODIUM Iliqicolor Enzymatic	HumaStar analyzer, Humalyzer photometer	mmol/l	6.46	5.57 - 7.34	mval/l	6.46	5.57 - 7.34
SODIUM ISE direct	HumaStar 600, Humalyte	mmol/l	160	155 - 165	mval/l	160	155 - 165
SODIUM RAPID Precipitation method	HumaStar analyzer, Humalyzer photometer	mmol/l	151	131 - 171	mval/l	151	131 - 171
TIBC Fe saturation Fe determination using IRON TPTZ Iliqicolor	HumaStar analyzer	µmol/l	134	106 - 161	mval/l	134	106 - 161
TOTAL PROTEIN Iliqicolor Biuret	HumaStar analyzer AUTOCAL LOT 0017	g/l	105	87.0 - 123	µg/dl	588	486 - 689
	HumaStar analyzer AUTOCAL LOT 0018	g/l	100.0	94.4 - 106	g/dl	10.0	9.44 - 10.6
TRIGLYCERIDES Iliqicolor ^{smm} POD	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	mmol/l	93.9	85.3 - 102	g/dl	9.39	8.53 - 10.2
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.68	2.29 - 3.07	mg/dl	235	201 - 269
UREA Iliqicolor Berthelot mod.	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	mmol/l	2.74	2.51 - 2.98	mg/dl	240	220 - 261
UREA IliqUV rease (UV)	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	mmol/l	23.6	20.0 - 27.3	mg/dl	142	120 - 164
URIC ACID Iliqicolor ^{plaa} Uricase	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	mmol/l	23.8	22.1 - 25.6	mg/dl	143	133 - 154
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	22.3	20.8 - 24.0	mg/dl	134	125 - 144
URIC ACID Iliqicolor ^{plaa} Uricase	HumaStar analyzer AUTOCAL LOT 0017 and 0018	µmol/l	636	520 - 755	mg/dl	10.7	8.74 - 12.7
	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	607	556 - 660	mg/dl	10.2	9.35 - 11.1
URIC ACID Iliqicolor Uricase	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	569	491 - 648	mg/dl	9.57	8.25 - 10.9
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	613	504 - 720	mg/dl	10.3	8.47 - 12.1
	HumaStar analyzer	µmol/l	613	544 - 678	mg/dl	10.3	9.14 - 11.4

001 | valid of 13.03.2021

725102 - Version 3/07 - 2021

Huma

Human Gesellschaft für Biochemie und Diagnostik
Max-Planck-Ring 21 · 65205 Wiesbaden · Germany
Telefon +49 6122-9988-0 · Telefax +49 612-9988-100 · e-Mail human@human

Reagent name / Reagenzbezeichnung / Paramètre / Paramètre / Methode / Methode / Método / Methode	Applicable for / anwendbar für / aplicable por / applicable pour	SI Unit / SI Einheit / Unidad SI / Unité SI	Mean / Mittelwert / Media / Moyenne	Range / Bereich / Rango / Plage de valeurs	Unit / Einheit / Unidad / Unité	Mean / Mittelwert / Media / Moyenne	Range / Bereich / Rango / Plage de valeurs
HUMAN reagent kits							
ACID PHOSPHATASE α-Naphthylphosphate, Hillmann mod., 37°C	AUTOCAL LOT 0017	µkat/l	0.26	0.19 - 0.32			
	AUTOCAL LOT 0018	µkat/l	0.28	0.24 - 0.33	U/l	15.3	13.6 - 18.9
ALBUMIN Iliquicolor Bromocresol green	HumaStar analyzer AUTOCAL LOT 0017 and 0018	g/l	54.3	49.4 - 59.3	g/dl	5.43	4.94 - 5.91
	Humalyzer photometer	g/l	50.2	45.1 - 55.2	g/dl	5.02	4.51 - 5.52
ALKALINE PHOSPHATASE Iliquicolor AMP buffer, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	5.43	5.02 - 5.87	U/l	326	303 - 352
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	4.82	4.28 - 5.33	U/l	289	257 - 320
	Humalyzer photometer	µkat/l	4.68	4.15 - 5.22	U/l	281	249 - 313
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	6.35	5.65 - 7.07	U/l	381	339 - 424
ALKALINE PHOSPHATASE opt. Iliquicolor DEA Buffer, 37°C, GSKC/DGKC	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	6.70	6.18 - 7.22	U/l	402	371 - 433
	Humalyzer photometer	µkat/l	6.85	5.83 - 7.87	U/l	411	350 - 472
	Humalyzer 4000, Setting 004H	µkat/l	5.70	5.00 - 6.40	U/l	342	300 - 384
	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	µkat/l	4.53	4.07 - 5.02	U/l	272	244 - 301
α-AMYLASE Iliquicolor CNFC3, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	4.70	4.25 - 5.13	U/l	282	255 - 308
	Humalyzer photometer	g/l	1.65	1.25 - 2.06	mg/dl	165	125 - 206
APOLOPROTEIN A1 Immunoturbidimetry	HumaStar analyzer	g/l	1.95	1.74 - 2.16	mg/dl	195	174 - 216
	Humalyzer photometer	g/l	1.35	0.95 - 1.74	mg/dl	135	95.3 - 174
auto-BILIRUBIN-D Iliquicolor DPD method	AUTOCAL LOT 0017	µmol/l	45.2	37.6 - 52.7	mg/dl	2.64	2.20 - 3.08
	AUTOCAL LOT 0018	µmol/l	40.7	38.8 - 42.6	mg/dl	2.38	2.27 - 2.49
	Humalyzer 4000	µmol/l	40.2	26.7 - 53.5	mg/dl	2.35	1.56 - 3.13
auto-BILIRUBIN-T Iliquicolor DPD method	AUTOCAL LOT 0017, Humalyzer photometer 4000	µmol/l	66.0	61.4 - 70.6	mg/dl	3.86	3.59 - 4.13
	AUTOCAL LOT 0018	µmol/l	68.4	63.3 - 73.4	mg/dl	4.00	3.70 - 4.29
BILIRUBIN Iliquicolor DCA method	Humalyzer photometer	µmol/l	62.3	50.8 - 73.7	mg/dl	3.64	2.97 - 4.31
BILIRUBIN DIRECT/TOTAL Iliquicolor Determination of Bilirubin total Jendrassik-Gróf	Humalyzer photometer w/o Humalyzer 4000	µmol/l	65.3	55.2 - 75.4	mg/dl	3.82	3.23 - 4.41
	Humalyzer 4000	µmol/l	60.5	54.0 - 67.0	mg/dl	3.54	3.16 - 3.92
BILIRUBIN DIRECT/TOTAL Iliquicolor Determination of Bilirubin direct Jendrassik-Gróf	Humalyzer photometer	µmol/l	43.3	34.0 - 52.3	mg/dl	2.53	1.99 - 3.06
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	3.03	2.83 - 3.23	mg/dl	12.1	11.3 - 12.9
CALCIUM Iliquicolor Ortho-cresolphthalein	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	3.03	2.85 - 3.23	mg/dl	12.1	11.4 - 12.9
	Humalyzer photometer	mmol/l	3.10	2.95 - 3.23	mg/dl	12.4	11.8 - 12.9
Chloride ISE direct	HumaStar 600, Humalyte	mmol/l	116	107 - 124	mg/dl	411	379 - 440
CHLORIDE Iliquicolor TPTZ method	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	124	117 - 131	mg/dl	440	415 - 464
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	119	108 - 129	mg/dl	422	383 - 457
	Humalyzer photometer Primus only	mmol/l	121	105 - 137	mg/dl	429	372 - 486
	Humalyzer photometer 4000 only	mmol/l	131	116 - 146	mg/dl	464	411 - 518
CHOLESTEROL Iliquicolor CHOD-PAP	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	8.15	7.03 - 9.26	mg/dl	315	272 - 358
	HumaStar analyzer AUTOCAL LOT 0018, Humalyzer photometer	mmol/l	8.48	7.06 - 9.88	mg/dl	328	273 - 382
CHOLINESTERASE Iliquicolor Butyrylthiocholine, 37°C, GSKC/DGKC	HumaStar analyzer	µkat/l	131	123 - 140	U/l	7880	7359 - 8400
	Humalyzer 4000	µkat/l	140	126 - 154	U/l	8400	7566 - 9234
CK NAC activated Enzymatic, 37°C	Humalyzer photometer	µkat/l	11.9	9.50 - 14.2	U/l	712	570 - 854
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	13.0	11.9 - 14.1	U/l	782	715 - 848
CK NAC IliquIV Enzymatic, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0018, Humalyzer 4000/Primus	µkat/l	12.0	10.9 - 13.0	U/l	718	656 - 779
	Humalyzer 2000/3000	µkat/l	10.8	8.50 - 13.1	U/l	649	510 - 788
	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	µmol/l	449	350 - 547	mg/dl	5.08	3.96 - 6.19
CREATININE Iliquicolor Jaffé	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	443	419 - 467	mg/dl	5.01	4.74 - 5.28
	Humalyzer photometer	µmol/l	419	371 - 466	mg/dl	4.74	4.20 - 5.27
CREATININE (enzym) Iliquicolor Enzymatic	HumaStar analyzer AUTOCAL LOT 0018, w/o HumaStar 600 AUTOCAL LOT 0017	µmol/l	468	429 - 507	mg/dl	5.29	4.85 - 5.73
	HumaStar 600 AUTOCAL LOT 0017	µmol/l	425	395 - 454	mg/dl	4.81	4.47 - 5.14
	Humalyzer 4000	µmol/l	418	349 - 487	mg/dl	4.73	3.95 - 5.51
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.25	1.78 - 2.72	U/l	135	107 - 163
γ-GT Iliquicolor Gamma-Glutamyl-3-carboxy-4-nitroanilide, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.37	2.13 - 2.58	U/l	142	128 - 155
	Humalyzer photometer	µkat/l	2.05	1.72 - 2.37	U/l	123	103 - 142
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	11.4	10.7 - 12.0	mg/dl	205	192 - 217
GLUCOSE Iliquicolor OD	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	11.0	10.3 - 11.7	mg/dl	198	185 - 211
	Humalyzer photometer	mmol/l	11.5	10.4 - 12.5	mg/dl	207	188 - 225
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	11.5	10.5 - 12.5	mg/dl	208	190 - 225
GLUCOSE IliquIV ⁶⁰⁰⁰ okinase	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	11.0	10.3 - 11.8	mg/dl	199	186 - 212
	Humalyzer photometer	mmol/l	12.4	10.8 - 14.0	mg/dl	223	194 - 252

Target Values / Sollwerte / Valores Asignados/ Valeurs Cibles

Reagent name / Reagensbezeichnung / Paramètre / Paramètre / Methode / Methode / Método / Méthode	*Applicable for / anwendbar für / applicable pour / applicable pour	SI Unit / Einheit / Unidad SI / Unidade SI	Mean / Mittelwert / Media / Moyenne	Range / Bereich / Rango / Plage de valeurs	Unit / Einheit / Unidad / Unidade	Mean / Mittelwert / Media / Moyenne	Range / Bereich / Rango / Plage de valeurs	
HUMAN reagent kits								
ACID PHOSPHATASE p-Naphthylphosphate, Hillmann mod., 37°C	AUTOCAL LOT 0017	µkat/l	0.26	0.19 - 0.32	U/l	15.3	11.6 - 18.9	
	AUTOCAL LOT 0018	µkat/l	0.28	0.24 - 0.31	U/l	16.5	14.5 - 18.5	
ALBUMIN Iliqicolor Bromocresol green	HumaStar analyzer AUTOCAL LOT 0017 and 0018	g/l	54.3	49.4 - 59.1	g/dl	5.43	4.94 - 5.91	
	Humalyzer photometer	g/l	50.2	45.1 - 55.2	g/dl	5.02	4.51 - 5.52	
ALKALINE PHOSPHATASE Iliqicolor AMP buffer, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	5.43	5.02 - 5.87	U/l	326	301 - 352	
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	4.82	4.28 - 5.33	U/l	289	257 - 320	
	Humalyzer photometer	µkat/l	4.68	4.15 - 5.22	U/l	281	249 - 313	
ALKALINE PHOSPHATASE opt. Iliqicolor DEA Buffer, 37°C, GSKC/DGKC	Humalyzer photometer	µkat/l	6.35	5.65 - 7.07	U/l	381	339 - 424	
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	6.70	6.18 - 7.22	U/l	402	371 - 433	
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	6.85	5.83 - 7.87	U/l	411	350 - 472	
	Humalyzer photometer	µkat/l	5.70	5.00 - 6.40	U/l	342	300 - 384	
alpha-AMYLASE Iliqicolor CNPG3, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	µkat/l	4.53	4.07 - 5.02	U/l	272	244 - 301	
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	4.70	4.25 - 5.13	U/l	282	255 - 308	
APOLOPROTEIN A1 Immunoturbidimetry	HumaStar analyzer	g/l	1.65	1.25 - 2.06	mg/dl	165	125 - 206	
	Humalyzer photometer	g/l	1.95	1.74 - 2.16	mg/dl	195	174 - 216	
APOLOPROTEIN B Immunoturbidimetry	HumaStar analyzer, Humalyzer photometer		g/l	1.35	0.95 - 1.74	mg/dl	135	95.3 - 174
	AUTOCAL LOT 0017	µmol/l	45.2	37.6 - 52.7	mg/dl	2.64	2.20 - 3.08	
auto-BILIRUBIN-D Iliqicolor DPD method	AUTOCAL LOT 0018	µmol/l	40.7	38.8 - 42.6	mg/dl	2.38	2.27 - 2.49	
	Humalyzer 4000	µmol/l	40.2	26.7 - 53.5	mg/dl	2.35	1.56 - 3.13	
	AUTOCAL LOT 0017, Humalyzer photometer 4000	µmol/l	66.0	61.4 - 70.6	mg/dl	3.86	3.59 - 4.13	
auto-BILIRUBIN-T Iliqicolor DPD method	AUTOCAL LOT 0018	µmol/l	68.4	63.3 - 73.4	mg/dl	4.00	3.70 - 4.29	
	Humalyzer photometer	µmol/l	62.3	50.8 - 73.7	mg/dl	3.64	2.97 - 4.31	
BILIRUBIN Iliqicolor DCA method	Humalyzer photometer	µmol/l	62.3	50.8 - 73.7	mg/dl	3.64	2.97 - 4.31	
BILIRUBIN DIRECT/TOTAL Iliqicolor Determination of Bilirubin total Jendrassik-Gróf	Humalyzer photometer w/o Humalyzer 4000	µmol/l	65.3	55.2 - 75.4	mg/dl	3.82	3.23 - 4.41	
	Humalyzer 4000	µmol/l	60.5	54.0 - 67.0	mg/dl	3.54	3.16 - 3.92	
BILIRUBIN DIRECT/TOTAL Iliqicolor Determination of Bilirubin direct Jendrassik-Gróf	Humalyzer photometer		µmol/l	43.3	34.0 - 52.3	mg/dl	2.53	1.99 - 3.06
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	3.03	2.83 - 3.23	mg/dl	12.1	11.3 - 12.9	
CALCIUM Iliqicolor Ortho-cresolphthalein	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	3.03	2.85 - 3.23	mg/dl	12.1	11.4 - 12.9	
	Humalyzer photometer	mmol/l	3.10	2.95 - 3.23	mg/dl	12.4	11.8 - 12.9	
	HumaStar 600, Humalyte	mmol/l	116	107 - 124	mg/dl	411	379 - 440	
Chloride ISE direct	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	124	117 - 131	mg/dl	440	415 - 464	
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	119	108 - 129	mg/dl	422	383 - 457	
	Humalyzer photometer Primus only	mmol/l	121	105 - 137	mg/dl	429	372 - 486	
	Humalyzer photometer 4000 only	mmol/l	131	116 - 146	mg/dl	464	411 - 518	
CHOLESTEROL Iliqicolor CHOD-PAP	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	8.15	7.03 - 9.26	mg/dl	315	272 - 358	
	HumaStar analyzer AUTOCAL LOT 0018, Humalyzer photometer	mmol/l	8.48	7.06 - 9.88	mg/dl	328	273 - 382	
	HumaStar analyzer	µkat/l	131	123 - 140	U/l	7880	7359 - 8400	
CHOLINESTERASE Iliqicolor Butyrylthiocholine, 37°C, GSKC/DGKC	Humalyzer 4000	µkat/l	140	126 - 154	U/l	8400	7566 - 9234	
	Humalyzer photometer	µkat/l	11.9	9.50 - 14.2	U/l	712	570 - 854	
CK NAC activated Enzymatic, 37°C	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	13.0	11.9 - 14.1	U/l	782	715 - 848	
CK NAC IliqUV Enzymatic, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0018, Humalyzer 4000/Primus	µkat/l	12.0	10.9 - 13.0	U/l	718	656 - 779	
	Humalyzer 2000/3000	µkat/l	10.8	8.50 - 13.1	U/l	649	510 - 788	
	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	µmol/l	449	350 - 547	mg/dl	5.08	3.96 - 6.19	
auto-CREATININE Iliqicolor Iaffé	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	443	419 - 467	mg/dl	5.01	4.74 - 5.28	
	Humalyzer photometer	µmol/l	419	371 - 466	mg/dl	4.74	4.20 - 5.27	
CREATININE Iliqicolor Iaffé	HumaStar analyzer AUTOCAL LOT 0018, w/o HumaStar 600 AUTOCAL LOT 0017	µmol/l	468	429 - 507	mg/dl	5.29	4.85 - 5.73	
	HumaStar 600 AUTOCAL LOT 0017	µmol/l	425	395 - 454	mg/dl	4.81	4.47 - 5.14	
	Humalyzer 4000	µmol/l	418	349 - 487	mg/dl	4.73	3.95 - 5.51	
gamma-GT Iliqicolor Gamma-Glutamyl-3-carboxy-4-nitroanilide, 37°C, IFCC	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.25	1.78 - 2.72	U/l	135	107 - 163	
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.37	2.13 - 2.58	U/l	142	128 - 155	
	Humalyzer photometer	µkat/l	2.05	1.72 - 2.37	U/l	123	103 - 142	
GLUCOSE Iliqicolor GOD	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	11.4	10.7 - 12.0	mg/dl	205	192 - 217	
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	11.0	10.3 - 11.7	mg/dl	198	185 - 211	
	Humalyzer photometer	mmol/l	11.5	10.4 - 12.5	mg/dl	207	188 - 225	
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	11.5	10.5 - 12.5	mg/dl	208	190 - 225	
GLUCOSE IliqUV ^{mono} Hexokinase	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	11.0	10.3 - 11.8	mg/dl	199	186 - 212	
	Humalyzer photometer	mmol/l	12.4	10.8 - 14.0	mg/dl	223	194 - 252	

Reagent name / Reagenzbzeichnung / Parámetro / Paramètre Method / Methode / Método / Méthode	applicable for / anwendbar für / aplicable por / applicable pour	SI Unit SI Einheit Unidad SI Unité SI	Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs	Unit Einheit Unidad Unité	Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs
GOT (ASAT) IFCC mod. IliqUV Enzymatic, 37°C, IFCC without PSP	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.23	2.10 - 2.38	U/l	134	126 - 143
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.05	1.92 - 2.18	U/l	123	115 - 131
	Humalyzer photometer	µkat/l	2.08	1.82 - 2.33	U/l	125	109 - 140
GPT (ALAT) IFCC mod. IliqUV Enzymatic, 37°C, IFCC without PSP	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.05	1.88 - 2.20	U/l	123	113 - 132
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	1.98	1.77 - 2.20	U/l	119	106 - 132
	Humalyzer photometer	µkat/l	1.92	1.60 - 2.25	U/l	115	96.1 - 135
HDL CHOLESTEROL Iliqcolor Homogenous enzymatic assay	AUTOCAL LOT 0017	mmol/l	3.15	2.69 - 3.62	mg/dl	122	104 - 140
	AUTOCAL LOT 0018	mmol/l	2.61	2.24 - 3.00	mg/dl	101	86.8 - 116
	for Kit calibrator only	mmol/l	2.77	2.43 - 3.10	mg/dl	107	94.1 - 120
HDL CHOLESTEROL Precipitation method	Humalyzer photometer	mmol/l	3.31	2.92 - 3.72	mg/dl	128	113 - 144
IRON Iliqcolor CAB	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	52.1	37.8 - 66.2	µg/dl	291	211 - 370
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	50.8	46.2 - 55.7	µg/dl	284	258 - 311
	Humalyzer photometer	µmol/l	39.9	27.2 - 52.6	µg/dl	223	152 - 294
IRON TPTZ Iliqcolor TPTZ	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	54.4	49.9 - 59.1	µg/dl	304	279 - 330
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	53.2	49.9 - 56.6	µg/dl	297	279 - 316
	Humalyzer photometer	µmol/l	51.7	45.3 - 58.2	µg/dl	289	253 - 325
IMMUNOGLOBULINS direct IgA Immunoturbidimetry	HumaStar analyzer, Humalyzer photometer	g/l	2.58	2.28 - 2.88	mg/dl	258	228 - 288
IMMUNOGLOBULINS direct IgG Immunoturbidimetry	HumaStar analyzer, Humalyzer photometer	g/l	12.3	9.83 - 14.8	mg/dl	1234	983 - 1484
IMMUNOGLOBULINS direct IgM Immunoturbidimetry	HumaStar 600, Humalyzer 4000	g/l	0.86	0.78 - 0.93	mg/dl	85.6	78.2 - 93.0
	HumaStar 100/200/300SR, Humalyzer 2000/3000	g/l	1.16	0.90 - 1.43	mg/dl	116	89.7 - 143.
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	11.6	10.4 - 12.8	U/l	676	625 - 767
LDH SCE mod. IliqUV Substrate Pyruvate, 37°C, SCE	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	11.3	10.3 - 12.2	U/l	677	619 - 733
	Humalyzer photometer	µkat/l	11.3	9.42 - 13.2	U/l	677	565 - 790
	AUTOCAL LOT 0017	mmol/l	4.22	2.82 - 5.64	mg/dl	163	109 - 218
LDL CHOLESTEROL Iliqcolor Homogenous enzymatic assay	AUTOCAL LOT 0018	mmol/l	4.58	3.05 - 6.10	mg/dl	177	118 - 236
	Kit calibrator	mmol/l	4.78	3.62 - 5.95	mg/dl	185	140 - 230
	LOT 21001 and higher: AUTOCAL LOT 0017	µkat/l	1.39	1.25 - 1.53	U/l	83.5	75.0 - 92.0
LIPASE Iliqcolor Substrate Methylresorufin, 37°C	LOT 21001 and higher: AUTOCAL LOT 0018	µkat/l	1.21	1.05 - 1.37	U/l	72.5	62.9 - 82.1
	up to LOT 20010: AUTOCAL LOT 0017 and 0018	µkat/l	1.09	0.81 - 1.38	U/l	65.5	48.4 - 82.6
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	1.30	1.16 - 1.43	mg/dl	3.15	2.83 - 3.47
MAGNESIUM Iliqcolor Xylydyl blue	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	1.24	1.07 - 1.41	mg/dl	3.01	2.59 - 3.43
	Humalyzer photometer	mmol/l	1.34	1.21 - 1.47	mg/dl	3.25	2.93 - 3.57
	AUTOCAL LOT 0017	µkat/l	2.98	2.80 - 3.15	U/l	179	168 - 189
PANCREAS-AMYLASE Iliqcolor EPS-G7, 37°C	AUTOCAL LOT 0018	µkat/l	2.57	2.35 - 2.77	U/l	154	141 - 166
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	2.78	2.62 - 2.94	mg/dl	8.61	8.10 - 9.11
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.85	2.63 - 3.07	mg/dl	8.83	8.14 - 9.52
PHOSPHORUS Iliqrapid Molybdate (UV)	Humalyzer photometer	mmol/l	2.77	2.44 - 3.11	mg/dl	8.59	7.55 - 9.62
	HumaStar 600, Humalyte	mmol/l	6.30	5.91 - 6.69	mval/l	6.30	5.91 - 6.69
POTASSIUM ISE direct	HumaStar analyzer, Humalyzer photometer	mmol/l	6.46	5.57 - 7.34	mval/l	6.46	5.57 - 7.34
POTASSIUM IliqUV Enzymatic	HumaStar 600, Humalyte	mmol/l	160	155 - 165	mval/l	160	155 - 165
SODIUM ISE direct	HumaStar analyzer, Humalyzer photometer	mmol/l	151	131 - 171	mval/l	151	131 - 171
SODIUM Iliqcolor Enzymatic	Humalyzer photometer	mmol/l	134	106 - 161	mval/l	134	106 - 161
SODIUM RAPID Precipitation method	HumaStar analyzer	µmol/l	105	87.0 - 123	µg/dl	588	486 - 689
TIBC Fe saturation Fe determination using IRON TPTZ Iliqcolor	HumaStar analyzer AUTOCAL LOT 0017	g/l	91.2	86.1 - 96.2	g/dl	9.12	8.61 - 9.62
TOTAL PROTEIN Iliqcolor Biuret	HumaStar analyzer AUTOCAL LOT 0018	g/l	100.0	94.4 - 106	g/dl	10.0	9.44 - 10.6
	Humalyzer photometer	g/l	93.9	85.3 - 102	g/dl	9.39	8.53 - 10.2
	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	mmol/l	2.68	2.29 - 3.07	mg/dl	235	201 - 269
TRIGLYCERIDES Iliqcolor ^{mnso} POD	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.74	2.51 - 2.98	mg/dl	240	220 - 261
	Humalyzer photometer	mmol/l	23.6	20.0 - 27.3	mg/dl	142	120 - 164
UREA Iliqcolor Berthelot mod.	HumaStar analyzer AUTOCAL LOT 0017, Humalyzer photometer	mmol/l	23.8	22.1 - 25.6	mg/dl	143	133 - 154
UREA IliqUV Urease (UV)	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	22.3	20.8 - 24.0	mg/dl	134	125 - 144
	Humalyzer photometer	µmol/l	636	520 - 755	mg/dl	10.7	8.74 - 12.7
URIC ACID Iliqcolor ^{plus} Uricase	HumaStar analyzer AUTOCAL LOT 0017 and 0018	µmol/l	607	556 - 660	mg/dl	10.2	9.35 - 11.1
	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	569	491 - 648	mg/dl	9.57	8.25 - 10.9
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	613	504 - 720	mg/dl	10.3	8.47 - 12.1
URIC ACID Iliqcolor Uricase	Humalyzer photometer	µmol/l	613	544 - 678	mg/dl	10.3	9.14 - 11.4

CS-5P

Rev. 001 | valid of 13.03.2021

Rev. 13.11.02 | Version 1/07 - 2021

HUMAN

Human Gesellschaft für Biochemica und Diagnostica m
Max-Planck-Ring 21 65205 Wiesbaden - Gem
Telefon +49 6122-9988-0 Telefax +49 612-9988-100 e-Mail human@human

Reagent name / Reagenzbezeichnung / Parameter / Paramètre Method / Methode / Método/ Método	applicable for / Anwendung für / applicable for / applicable pour	SI Unit SI Einheit Unidad SI Unite SI	Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs	Unit Einheit Unidad Unite	Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs
COT (ASAT) FCC mod. liquidUV Enzymatic, 37°C, FCC without PVP	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.25	2.10 - 2.38	U/l	134	126 - 143
	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	2.05	1.92 - 2.18	U/l	123	115 - 131
	HumaStar photometer	µkat/l	2.08	1.82 - 2.33	U/l	123	113 - 132
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	2.05	1.88 - 2.20	U/l	119	106 - 132
CPT (ASAT) FCC mod. liquidUV Enzymatic, 37°C, FCC without PVP	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	1.98	1.77 - 2.20	U/l	115	96.1 - 135
	HumaStar photometer	µkat/l	1.92	1.60 - 2.25	U/l	122	104 - 140
	AUTOCAL LOT 0017	mmol/l	3.15	2.69 - 3.62	mg/dl	107	86.8 - 116
	AUTOCAL LOT 0018	mmol/l	2.61	2.34 - 3.00	mg/dl	107	94.1 - 120
HDL CHOLESTEROL liquidcolor Heterogeneous enzymatic assay	for Kit calibration only	mmol/l	2.77	2.43 - 3.10	mg/dl	128	113 - 144
	HumaStar photometer	mmol/l	3.31	2.92 - 3.72	mg/dl	291	211 - 370
HDL CHOLESTEROL Precipitation method	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	52.1	37.8 - 66.3	µg/dl	284	258 - 311
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	50.8	46.2 - 55.7	µg/dl	223	152 - 294
HDL CHOLESTEROL CAS	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	39.9	27.2 - 52.6	µg/dl	304	279 - 330
	HumaStar photometer	µmol/l	54.4	49.9 - 59.1	µg/dl	297	279 - 316
HDL TPTZ liquidcolor TPTZ	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	53.2	49.9 - 56.6	µg/dl	289	251 - 325
	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	51.7	45.3 - 58.2	µg/dl	289	251 - 325
IMMUNOGLOBULINS direct IBA (immunoturbidimetry)	HumaStar analyzer, HumaStar photometer	g/l	2.58	2.28 - 2.88	mg/dl	258	228 - 288
	HumaStar analyzer, HumaStar photometer	g/l	12.3	9.83 - 14.8	mg/dl	1234	983 - 1484
IMMUNOGLOBULINS direct IBC (immunoturbidimetry)	HumaStar analyzer, HumaStar photometer	g/l	0.86	0.78 - 0.93	mg/dl	85.6	78.2 - 93.0
	HumaStar 600, HumaStar 4000	g/l	1.16	0.90 - 1.43	mg/dl	116	89.7 - 143
IMMUNOGLOBULINS direct IGM (immunoturbidimetry)	HumaStar 100/200/100SR, HumaStar 2000/3000	µkat/l	11.6	10.4 - 12.8	U/l	696	625 - 767
	HumaStar analyzer AUTOCAL LOT 0017	µkat/l	11.3	10.3 - 12.2	U/l	677	565 - 790
LDH SCE mod. liquidUV Spectrofluorimetry, 37°C, SCE	HumaStar analyzer AUTOCAL LOT 0018	µkat/l	11.3	9.42 - 13.2	U/l	165	109 - 218
	HumaStar photometer	mmol/l	4.22	2.82 - 5.64	mg/dl	177	118 - 236
LDL CHOLESTEROL liquidcolor Heterogeneous enzymatic assay	AUTOCAL LOT 0017	mmol/l	4.58	3.05 - 6.10	mg/dl	185	140 - 230
	AUTOCAL LOT 0018	mmol/l	4.78	3.62 - 5.95	mg/dl	185	150 - 92.0
LDL CHOLESTEROL liquidcolor Homogeneous enzymatic assay	Kit calibration	µkat/l	1.39	1.25 - 1.53	U/l	83.5	75.0 - 92.0
	LOT 22001 and Higher- AUTOCAL LOT 0017 LOT 22001 and Higher- AUTOCAL LOT 0018 up to LOT 20013- AUTOCAL LOT 0017 and 0018	µkat/l	1.21	1.05 - 1.37	U/l	72.5	62.9 - 82.1
LDL CHOLESTEROL liquidcolor Substrate Methylresorufin, 37°C	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	1.09	0.81 - 1.38	U/l	65.5	48.4 - 82.6
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	1.30	1.16 - 1.43	mg/dl	3.15	2.83 - 3.47
LDL CHOLESTEROL liquidcolor Substrate Methylresorufin, 37°C	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	1.24	1.07 - 1.41	mg/dl	3.01	2.59 - 3.43
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	1.34	1.21 - 1.47	mg/dl	3.25	2.93 - 3.57
MAGNESIUM liquidcolor Hydride flux	HumaStar photometer	µkat/l	2.98	2.80 - 3.16	U/l	154	141 - 166
	AUTOCAL LOT 0017	µkat/l	2.57	2.35 - 2.77	U/l	179	168 - 189
MAGNESIUM liquidcolor Hydride flux	AUTOCAL LOT 0018	µkat/l	2.78	2.62 - 2.94	mg/dl	8.61	8.10 - 9.11
	HumaStar analyzer AUTOCAL LOT 0017	mmol/l	2.85	2.63 - 3.07	mg/dl	8.83	8.34 - 9.52
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.77	2.44 - 3.11	mg/dl	8.59	7.55 - 9.62
	HumaStar photometer	mmol/l	2.85	2.63 - 3.07	mg/dl	8.59	7.55 - 9.62
MAGNESIUM liquidcolor Hydride flux	HumaStar 600, HumaStar	mmol/l	6.30	5.91 - 6.69	mval/l	6.30	5.91 - 6.69
	HumaStar analyzer, HumaStar photometer	mmol/l	6.46	5.57 - 7.34	mval/l	6.46	5.57 - 7.34
MAGNESIUM liquidcolor Hydride flux	HumaStar 600, HumaStar	mmol/l	160	155 - 165	mval/l	160	155 - 165
	HumaStar analyzer, HumaStar photometer	mmol/l	151	151 - 171	mval/l	151	131 - 171
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer, HumaStar photometer	mmol/l	134	106 - 161	mval/l	134	106 - 161
	HumaStar photometer	mmol/l	105	87.0 - 123	µg/dl	588	486 - 689
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer	µmol/l	105	87.0 - 123	µg/dl	588	486 - 689
	Fe saturation Fe determination using IRON TPTZ liquidcolor	HumaStar analyzer AUTOCAL LOT 0017	g/l	91.2	86.1 - 96.2	g/dl	9.12
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0018	g/l	100.0	94.4 - 106	g/dl	10.0	9.44 - 10.6
	HumaStar photometer	g/l	93.9	85.3 - 102	g/dl	9.39	8.53 - 10.2
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0017, HumaStar photometer	mmol/l	2.68	2.29 - 3.07	mg/dl	235	201 - 269
	HumaStar analyzer AUTOCAL LOT 0018	mmol/l	2.74	2.51 - 2.98	mg/dl	240	220 - 261
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0017, HumaStar photometer	mmol/l	23.8	22.1 - 25.6	mg/dl	143	133 - 154
	HumaStar photometer	mmol/l	22.3	20.8 - 24.0	mg/dl	134	125 - 144
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	636	520 - 755	mg/dl	10.7	8.74 - 12.7
	HumaStar photometer	µmol/l	607	556 - 660	mg/dl	10.2	9.35 - 11.1
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0017 and 0018	µmol/l	569	491 - 648	mg/dl	9.57	8.25 - 10.9
	HumaStar analyzer AUTOCAL LOT 0017	µmol/l	613	504 - 720	mg/dl	10.3	8.47 - 12.1
MAGNESIUM liquidcolor Hydride flux	HumaStar analyzer AUTOCAL LOT 0018	µmol/l	613	544 - 678	mg/dl	10.3	9.14 - 11.4
	HumaStar photometer	µmol/l	613	544 - 678	mg/dl	10.3	9.14 - 11.4

13-1P
Rev. 0011 04/01/19 (P.13.01.2022)
HW 1221010 - Human 907 - 0022

Human

Human Gesellschaft für Biochemie und Diagnostica mbH
Max-Planck-Ring 21 · 65205 Wiesbaden, Germany
Telefon +49 6122-9988-0 · Telefax +49 6122-9988-100 · e-Mail human@human.de