

CORRECTIVE ACTIONS

Month: November 2021

Date of Issue Of PT report : 03/12/2021

Explanation:

The results of November 2021 were published on 03/12/2021. The following parameters were unacceptable.

Sr no	Analyte	Method	Analyser name	No of participants	DV	CV	SD	Your value	SDI	U
1	GLUCOSE	GOP-POD	Transasia / erba	775	144.88	10.50	15.21	97.5 mg/dl	-3.12	1.09
2	CREATININE	Enzymatic colorimetric	Transasia / erba	150	10.42	8.83	0.92	1.11 mg/dl	-10.12	0.15

Corrective actions:

We have processed the EQAS sample of November 2021 on 12/11/2021. The actual value of Glucose was 157.4 mg/dl and Creatinine 11.15 mg/dl (Please refer to the attached document 1).

The mean value of EQAS for November 2021 was 144.88 mg/dl for Glucose and 10.42 mg/dl for Creatinine. Our actual values were satisfactory. There was an typing error while entering the above results in the EQAS online.

Result Reprint

Report Type : Patients

12-Nov-2021

Sr #	Sample ID	Test	Result Unit	Flag	Curve #	Result Date
1	1	CRENZ	11.15 mg/dl	H	114232	12-Nov-2021 14:13:29
2	1	GLURX(X)	157.4 mg/dl	H	114237	12-Nov-2021 14:14:59