



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - DECEMBER 2021



PC-1024

Lab Name **KING MICRO LAB** Lab No **13792**
Constituent Group **Chemistry II** Date of Result Entered : **18/12/2021**
PT item **Lyophilized Serum** Date of Report Published : **03/01/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	969	215.80	9.25	19.96	198.5 mg/dL	-0.87	1.28
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	533	107.40	11.34	12.17	109.4 mg/dL	0.16	1.05
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	725	2.24	11.51	0.26	2.05 mg/dL	-0.74	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	789	3.00	13.68	0.41	2.79 mg/dL	-0.51	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	738	5.21	9.17	0.48	5.3 g/dL	0.19	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	728	3.19	10.00	0.32	3.19 g/dL	0.00	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	776	4.39	19.07	0.84	3.99 mg/dL	-0.48	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	975	108.50	11.43	12.40	108.2 mg/dL	-0.02	0.79
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	803	96.54	15.20	14.68	103.3 mg/dL	0.46	1.04

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

KING MICRO LAB
PONNUSAMY COMPLEX, AUTOSTAND OPP , CHINNATHIRUPATHI BUS STOP
SALEM
TAMILNADU636008

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
******* End of Report *******