

CORRECTIVE ACTION

Month: February

Date of Issue of PT report: 15/03/2022

Explanation:

The results of February 2022 were published on 15/03/2022. The following parameters were unacceptable.

Sr no	Analyte	Method	Analyser name	No of participants	DV	CV	SD	Your value	SDI	U
1	Albumin	BCG-Colorimetric	Transasia / erba	597	3.40	8.90	0.30	4.63 g/dl	4.1	0.02

Corrective actions for QC:

The sample of February EQAS was preserved in the lab. We have calibrated the parameters and after rerunning the sample for albumin it is found to be 3.51 g/dl. The mean value of EQAS for Feb 2021 3.40 g/dl. So our value is Satisfactory.

External Quality Control :

We have performed an external QC (General Diagnostics, NABL Accredited) for the following parameters:

PARAMETERS	1 Pramod Sawant		2 Santosh Kulkarni		3 Pooja Mundra		4 Sujata Sohani		5 Pravin Raparkar	
	Lab Value	EQC	Lab Value	EQC	Lab Value	EQC	Lab Value	EQC	Lab Value	EQC
CHLOESTROL	143.0	135.2	123.0	125.7	212.0	218.0	222.0	209.9	180.0	186.0
CREATININE	1.01	1.10	0.84	0.90	0.46	0.50	0.50	0.50	0.72	0.70
ALBUMIN	4.53	4.44	3.97	4.10	4.10	4.05	4.22	4.01	4.68	4.33