

Date: 6th Aug. 2021

Calibration Certificate

This is to certify that Automated Haematology Analyzer **XN-10, Serial No: 39934** installed at "**Central Clinical Laboratory, Pravara Rural Hospital, Loni 413736**" has been calibrated successfully.

Calibration Done on: 5 Aug 2021

Calibration Validity: 5 Aug 2021 to 4 Aug 2022

Calibration Due on: 4 Aug 2022

Material: XN-CAL (Lot No.-11932101, Expiry-2021-08-15)

| Run | WBC-N | WBC-D | RBC | HGB | HCT | PLT |
|--------------|-------|-------|-------|------|-------|------|
| 1 | 6.90 | 6.96 | 4.30 | 11.8 | 34.3 | 248 |
| 2 | 6.80 | 6.93 | 4.29 | 11.8 | 34.4 | 249 |
| 3 | 6.78 | 6.90 | 4.28 | 11.8 | 34.2 | 242 |
| 4 | 6.74 | 7.00 | 4.28 | 11.7 | 34.3 | 236 |
| 5 | 6.56 | 6.91 | 4.32 | 11.9 | 34.7 | 239 |
| 6 | 6.83 | 6.78 | 4.32 | 11.7 | 34.6 | 239 |
| 7 | 6.61 | 6.89 | 4.27 | 11.8 | 34.2 | 244 |
| 8 | 6.73 | 6.90 | 4.32 | 11.7 | 34.6 | 244 |
| 9 | 6.72 | 6.99 | 4.30 | 11.7 | 34.4 | 237 |
| 10 | 6.78 | 6.89 | 4.31 | 11.7 | 34.6 | 241 |
| Mean | 6.75 | 6.92 | 4.30 | 11.8 | 34.4 | 242 |
| Target Value | 6.833 | 6.997 | 4.353 | 11.9 | 34.78 | 243 |
| CV% | 1.5% | 0.9% | 0.4% | 0.6% | 0.5% | 1.8% |
| Criteria | 3.0% | 3.0% | 1.5% | 1.0% | 1.5% | 4.0% |

Calibration accepted for above mentioned values.

For Sysmex India Pvt. Ltd.

Authorized Signatory



Sysmex India Pvt Ltd (CIN: U33120MH1998PTC115943)

Office 1002, Damji Shamji Business Galleria, 10th Floor, LBS Marg, Kanjurmarg (West), Mumbai – 400078, Maharashtra, India Tel. +91 (22) 6112 6666

Factory Village Malpur, Pargana Dharampur, Nalagarh Road, Baddi – 173205, Dist. Solan, Himachal Pradesh, India Tel. +91-9218422282 / 9816672282

www.sysmex.co.in