

**IN-HOUSE CALIBRATION CERTIFICATE**

<b>Name of the Equipment :-</b> Centrifuge	
<b>Equipment I.D. :-</b> PPTCT/DWH/CENTRI/01	
<b>Location :-</b> Room No. 104, PPTCT District Women Hospital, Gorakhpur	<b>Working Range:-</b> 5000
<b>Manufacture:-</b> Oscar	
<b>Calibrated on:-</b> 01/02/2022	<b>Next Calibration Due:-</b> 01/07/2022

**Method:- Traceability Certificate**

MICRO/HDRL/Tach-01 SRF No. & Date: 210915A/NN-105 Dt. 15/09/2021  
Standard: Line Seiki/TM5010EK-039, GODREJ, M-200925-25-1 Dt. 19/09/2021

**Test condition: Working****Room Temperature : 26 °C****RH: 74%**

Instrument	Reading					Mean
	1	2	3	4	5	
<b>Tachometer</b> (BRDMC/Micro/HDRL/Tachometer/01/SRL)	3008	3010	3005	3015	3009	3009
<b>Centrifuge 01</b> (PPTCT/DWH/DEGITALTERM./01)	3000	3000	3000	3000	3000	3000
<b>Acceptable variation: +/- 10% with standard Tachometer</b>						

**Remarks: Instrument has passed the acceptance criteria (Yes/No)..... YES**

Checked by

**Jyoti Kumar**  
Technical Manager  
H D R L

Verified by

Lab Director/Quality Manager  
H D R L