

**IN-HOUSE CALIBRATION CERTIFICATE**

Name of the Equipment :- Digital Thermometer	
Equipment I.D. :- PPTCT/DWH/THERMO/01	
Location :- Room No. 104, PPTCT District Women Hospital, Gorakhpur	Working Range:- -50°C
Manufacture:- Aptechdeals	
Calibrated on:- 01/02/2022	Next Calibration Due:- 01/07/2022

Method:- Traceability Certificate

MICRO/HDRL/Tach-01 SRF No. & Date: 210915A/NN Dt. 15/09/2021
Standard: Line AE/ACS/DTIS/01, ACS Noida, 201210 A/N-101 Dt. 10/12/2021

Test condition: Working**Room Temperature : 26°C****RH: 74%**

Instrument	Reading					Mean
	1	2	3	4	5	
Digital Thermometer SRL (BRDMC/Micro/HDRL/DegitalTerm./01/SRL)	3.8	3.3	3.5	3.7	3.4	3.54
Thermometer (HCTS) (PPTCT/DWH/DEGITALTERM./01)	3.2	3.5	3.9	3.6	3.7	3.58
Acceptable variation: +/- 10% with standard Tachometer						

Remarks: Instrument has passed the acceptance criteria (Yes/No)..... YES

Checked by

Jyoti Kumar
Technical Manager
HDRL

Verified by

Lab Director/Quality Manager
HDRL