

Analyte	Instrument	Result Value	Standard Unit
X pH	TransAsia Laura urine analyser	6.5	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

VISHWARAJ HOSPITAL LABORATORY (568)

Outlier And Analyte Summary Report

Result Details For Cycle No 210410 and Sample No 01


Report Date : 02/11/2021

Instrument : TransAsia Laura urine analyser

Analyte	Standard Unit	Result Value	Accepted Value
X pH	-	6.5	5,5,5,6
@ Specific Gravity	-	1.030	1.020,1.025,1.030
@ Glucose (urine)	-	3+	+2,+3
@ Protein (Urine)	-	Trace	Neg,Tr
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	1+	+1,+2,+3,+4
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
@ Nitrite (urine)	-	Absent	Absent

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit
X Specific Gravity	TransAsia Laura urine analyser	1.030	-
# Blood (urine)	TransAsia Laura urine analyser	1+	-
# Nitrite (urine)	TransAsia Laura urine analyser	Absent	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

VISHWARAJ HOSPITAL LABORATORY (568)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01


Report Date : 30/12/2021

Instrument : TransAsia Laura urine analyser

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6.5	5.5,6,6.5
X Specific Gravity	-	1.030	1.015,1.020,1.025
@ Glucose (urine)	-	3+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,Tr
# Blood (urine)	-	1+	
# Nitrite (urine)	-	Absent	
@ Ketone (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	2
Evaluated Parameters	7
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	85.71 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

End of Report

T-11

Proficiency Testing outliers and corrective action records

Sr. No	Date	Test Name	Lab Reported Result	Peer / Expected Value	Details of Root Cause analysis and Corrective action taken	Done by	Sign
1.	30/12/2021	Specific gravity (urine)	1.030	1.015-1.025	Random error observe in next cycle	Amar	A. Vishnise
2.							
3.							
4.							
5.							
6.							
7.							

Issue No. 1	Issue Date : 15-07-2021	Amend No. 00	Amend Date. 00	Page 1 of 1
Prepared and Issued By : <i>Dr. Dipit Galikwad</i>		Reviewed & Approved By : <i>Dr. Aniruddha Garud</i>		

CONTROLLED COPY