



Date 06-08-2021

CALIBRATION CERTIFICATE

This is to certify that the Hematology analyzer **Sysmex XN-350** bearing serial number 12783 installed at **Reliance Pathology lab Medical Center, Shahdol** as calibrated on 06-08-2021

Calibrator Used: XNL Calibrator

Lot No: 11932101

Expiry Date: 15-08-2021

The report of Blank Cycle, Repeatability and Calibration values were all found in acceptable range & Calibration data is attached with this certificate.

Next Calibration Cycle is due on 05-08-2022

For Transasia Bio Medicals Ltd


Abhishek Choudhary

Sr. Service Engineer