



**King George's Medical University Lucknow (U.P.)**  
**Department of Microbiology**

**CERTIFICATE OF CALIBRATION**

Name of equipment: Digital Thermometer

Equipment ID: ICTC-DWH-HDI-28

Location: Room No. 15

Working range: 2°C to 8°C

Manufacturer: NA

Calibrated on: 15/02/22

Next Calibration due on: 14/02/23

Method: Traceability Certificate

KG/Micro/HIV/DTHM-01/SRL Certificate No. 211219A/N-102R

Digital Temp Indicator with sensor AE/ACS/DTIS/01 ACS Noida 211210D/NN-101 Dt.  
09/12/2022

Test Conditions: Room Temperature: 27.8°C RH: 38%

Instrument	Readings (°C)					Mean
	1	2	3	4	5	
Refrigerator						
Master Thermometer reading	4.4	4.5	6	7.8	5	5.5
Laboratory Thermometer reading	6.8	5.9	6.8	7.9	5.8	6.5
Acceptable Variation: $\pm 2$ °C of the mean master thermometer reading						

Remarks: Instrument has passed the acceptance criteria (YES/NO)

  
Checked & Verified by