



# GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207/ 9028888728



CC-2957

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- Avior Clinical Diagnostic Centre MULUND (E), MUMBAI	Page No. :- 1 of 1 Certificate No. :- GTS/220224/01- 004 Date of Received :- 24.02.2022 Date of Calibration :- 24.02.2022 Next Calibration Due On :- 23.02.2023 Issue Date :- 28.02.2022 Calibration method No. :- MECH-WI-07 ULR No :- CC295722000000947F
Ambient Temp. (°C) :- 28.6 Relative Humidity (%RH) :- 52 Location of calibration :- On Site Condition of Item :- Ok	

<b>2. Description of Item</b>	
Name :- Centrifuge Id No :- ACDC/CEN/01 Make :- Remi C854/6 Type :- Knob	Range :- 2500 RPM Least Count :- - RPM Location :- LAB Sr No :- HLLC 5718 Dept :- PATHOLOGY


<b>3.Details of Equipment used for calibration</b>				
Name	Certificate No.	Certified By	ID No.	Calibration Validity
Digital Tachometer	NI/2106/002/003	Nishitronics Instrumentation	GTS/TM/01	31.05.2022

<b>*Mechanical Calibration</b>				
<b>4.Calibration Results</b> :-				
Calibration Points RPM	Standard Reading RPM	UUC Reading RPM	Error in RPM	Expanded Uncertainty in ± %
2500	2507	2500	-7	1.20

**Note:**

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By  
U.G  
Calibration Engineer  
Umesh.G



Approved By  
  
Technical Manager  
Swapnil Bhagwat

RF-55/01

End of Certificate