



CERTIFICATE OF CALIBRATION

Certificate No : CC-2223-S-334/01		Page. No : 1 of 1	
Customer Name & Address	ULR No.	CC272522000004641F	
M/s. Naalvar Diagnostics	SRF No.	VCI-2223-S-334	
New No.22, Old No. 23, Neelakidangu St, Karaikal-609 602	Date of Receipt	26.03.2022	
	Date of Calibration	26.03.2022	
	Due Date of Calibration	25.03.2023	
	Date of Issue	28.03.2022	
Details of Unit Under Calibration			
D 1 11			

Description	Centrifuge	Make	Labtech Medico
Range	100 to 3000 rpm	Model	80-1
Resolution	NA	Condition of UUC	Good
Serial Number	6029	Instrument Location	Bio Chemistry Lab
ID Number	ND/QSP/13	Cal. At	On Site

Environmental Condition		Calibration Method Used	
Temperature (°C)	23±5	National / International Standard	SANAS TR 45-02
Humidity (%RH)	40 - 70	Cal Procedure No	VCI-L03-WI-SPE-002
Standard Hand		Car i locedule No	VCI-L03-WI-SPE-002

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Digital Tachometer	VCI-M-TM-011	CC-22-23-L-015/08	Lutron / DT-2234C	National Standards	08.01.2023

Results of Calibration			
Sl. No.	UUC Set Point	Standard Reading (rpm)	Measurement Uncertainty (±) (%)
01	1	500.3	0.62
02	2	1027	0.19
03	3	1606	0.21
04	4	3005	0.26

Remarks:

- 1. UUC is Defined as the Unit Under Calibration.
- 2. This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- 3. The reported results are valid at the time of under stated condition of measurement.
- 4. Calibration of the UUC are traceable to National / International Standards.
- 5. The recalibration interval shall be decided by the user.
- 6. Nabl-133 guidelines are adopted for use of NABL symbol.
- 7. The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k=2.

Checked & Issued by:

M.Manju

(Engineer-Calibration)



Authorised by:

K.Sathish (Deputy Quality Manager)