



CC-2725

**CERTIFICATE OF CALIBRATION**

Certificate No: CC-2223-L-244/02

Page No : 1 of 1

Customer Name & Address M/s. NAALVAR DIAGNOSTICS New No.22, Old No.23, Neelakidangu St, Karaikal - 609 602.	ULR No.	CC2725220000004645F
	SRF No.	VCI-2223-L-244
	Date of Receipt	28.03.2022
	Date of Calibration	28.03.2022
	Due Date for Calibration	27.03.2023
	Date of Issue	29.03.2022

Details of Unit Under Calibration

Description	Micro Pipette	Make	Dragon Lab
Range	100 - 1000µl	Model	--
Resolution	5µl	Material	pvc
Serial Number	YE17BAG0056715	Operating Range	100 - 1000µl
ID Number	ND/QSP/08	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	--

Environmental Condition				Calibration Method Used	
Temperature (°C)	23±2	Humidity (%RH)	55±15	National / International Standard	ISO 8655-6:2002, ISO 8655-2:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	VCI-L03-WI-MASS-003

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	VCI-M-WB-017 / 55492	CC-21-22-S-994/01	Wensar / MAB 250	National Standards	14.10.2022

Z Factor: 1.00319

Result of Calibration in µl											
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.98	99.96	99.94	99.93	99.89	99.93	-0.07	0.04	0.20	
		99.97	99.95	99.91	99.92	99.85					
2	500	500.05	500.07	500.11	500.15	500.19	500.12	0.12	0.06	0.88	
		500.02	500.09	500.13	500.17	500.22					
3	1000	1001.25	1001.29	1001.33	1001.37	1001.41	1001.34	1.34	0.06	0.88	
		1001.27	1001.32	1001.35	1001.39	1001.43					

Note: Conformity Statement

Pass: Fail:

Remarks

- UUC is Defined as the Unit Under Calibration.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor $k=2$.
- This Calibration Certificate relates only to the above UUC & Reported results are valid at the time of and under the stated conditions of measurements.
- Calibration of the UUC are traceable to National / International Standards.
- This report shall not be reproduced in full/ part without prior permission of Venus Calibration and Instruments.
- The recalibration interval shall be determined on the user requirements.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.
- Our Nabl Scope is Upto 10µl to 10ml.
- The Deviation is within MPE, So after calibration is not required.

Checked & Issued by:

M. Manju
M.Manju

(Engineer - Calibration)



Authorised by:

C.Dhinakaran

(Technical / Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com, info@venuscal.com | www.venuscal.com