



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :- Avior Clinical Diagnostic Centre MULUND (E), MUMBAI		Page No.	:- 1 of 1
		Certificate No.	:- GTS/220224/01- 002
		Date of Received	:- 24.02.2022
		Date of Calibration	:- 24.02.2022
		Next Calibration Due On	:- 23.02.2023
		Issue Date	:- 28.02.2022
Ambient Temp. (°C)	:- 24.3	Calibration method No.	:- MECH-WI-06
Relative Humidity (%RH)	:- 61	ULR No	:- CC295722000000945F
Barometric Pressure (mbar)	:- 942.1		
Location of calibration	:- In Lab		
Condition of Item	:- Ok		

2. Description of Item			
Name	:- Micropipette	Range	:- 20 TO 200 µl
Id No	:- ACDC/PIP/02	Least Count	:- 0.2 µl
Make	:- Finnpiette	Location	:- LAB
Type	:- Variable	Sr No	:- LH 79338
		Dept	PATHOLOHY

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI/GTS/010621/001	Nishitronics Instrumentation	GTS/WB-01	31.05.2022

*Mechanical Calibration


4.Calibration Results :-				
Calibration Points µl	Standard Reading µl	Set Value on UUC µl	Error in µl	Expanded Uncertainty in ± µl
20	19.9263	20	0.0737	1.50
100	99.6451	100	0.3549	1.50
200	199.2707	200	0.7293	1.50

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

U.G
Calibration Engineer
Umesh.G



Approved By

SwB
Technical Manager
Swapnil Bhagawat

End of Certificate

RF-51/00