



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - DECEMBER 2021



PC-1024

Lab Name	ANKERITE PUSHPA SRIRAM HOSPITAL	Lab No	14688
Constituent Group	Chemistry II	Date of Result Entered :	20/12/2021
PT item	Lyophilized Serum	Date of Report Published :	03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	735	220.59	10.06	22.19	136.2 mg/dL	-3.8	1.64
2	UREA II	Urease UV / GLDH II	Transasia / Erba	555	110.52	10.22	11.30	45.4 mg/dL	-5.76	0.96
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	488	2.27	9.44	0.21	2 mg/dL	-1.26	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	649	2.93	11.30	0.33	1.9 mg/dL	-3.12	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	665	5.34	9.22	0.49	8.9 g/dL	7.24	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	650	3.21	9.20	0.30	4.9 g/dL	5.73	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	637	4.08	14.87	0.61	5.6 mg/dL	2.50	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	716	104.99	12.01	12.61	98 mg/dL	-0.55	0.94
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	605	85.49	17.53	14.98	198 mg/dL	7.51	1.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

ANKERITE PUSHPA SRIRAM HOSPITAL
C.R.P.F CHAURAH, MATI ROAD, MATI, BIJNOUR
LUCKNOW
UTTAR PRADESH 226002

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****