



AVM LABS PVT. LTD.

No. 49, 3rd Street | Moorthy Nagar | Chettiar Agaram | Porur | Chennai - 600 077.
Mob.: +91 770 893 2211 | 735 837 6115
Email : avmlabsales@gmail.com | avmlabs21@gmail.com | www.avmgroup.biz



CC208821000023044F

CALIBRATION CERTIFICATE

ULR No.: CC208821000023044F

CERT No.	: C-21/22-0735-01	DATE OF ISSUE	: 4-Oct-21	PAGE	: 1 OF 1
JOB No.	: J-21/22-0735	CALIBRATED ON	: 1-Oct-21	SUGG.NEXT CAL	: 1-Oct-22

DETAILS OF CUSTOMER :

NAME & ADDRESS	: M/s. EL-EVEN LABORATORIES NO. 37/B2, KIZHAKKEKARA LANE, CHANGAMPUZHA NAGAR P.O, SOUTH KALAMASSERY, ERNAKULAM - 682 033.	CALIBRATED AT	: IN LAB
		DATE OF RECEIPT	: 1-Oct-21
		CONDITION OF RECEIPT	: SATISFACTORY
		CALIB. PROCEDURE	: AVM-WI-M13

DETAILS OF DEVICE UNDER CALIBRATION :

NOMENCLATURE	: MICROPIPETTE	RANGE / SIZE	: 100 ~ 1000 µl
MANUFACTURER	: ERBAPETTE	OPERATING RANGE	: -
MODEL	: -	INPUT / OUTPUT	: -
TYPE	: -	RESOLUTION	: 1µl
SERIAL NUMBER	: 2174173	ACCURACY	: -
ID/TAG NO.	: E-005	LOCATION	: -

ENVIRONMENTAL CONDITIONS:

TEMPERATURE :	(23 ± 3) °C	HUMIDITY :	(50 ± 10) %RH
---------------	-------------	------------	---------------

MEASUREMENT TRACEABILITY

The reference measurement standards used are traceable to National/International Standards.

REFERENCE STANDARDS	SERIAL / TAG NUMBER	CERTIFICATE NUMBER	DUE DATE
ELECTRONIC SEMI MICRO BALANCE	327560/11 / BI-0094	C-AVM-IH-21/22-019	17-Jul-22

CALIBRATION RESULTS

Nominal Capacity	Measured Capacity (Mean)	Error	Expanded Uncertainty (±)
µl	µl	µl	µl
100	99.98	-0.02	0.13
500	499.97	-0.03	0.74
1000	1000.01	0.01	0.74

The reported Expanded Uncertainty is calculated at 95.45% Confidence Level with a coverage factor of k = 2.

END OF CERTIFICATE

CALIBRATED BY:

S.MANO
CALIBRATION ENGINEER

REVIEWED & APPROVED BY :

S.JAGATHEESWARI / AVICHI MEIYAPPAN
AUTHORISED SIGNATORY