



AVM LABS PVT. LTD.

No. 49, 3rd Street | Moorthy Nagar | Chettiar Agaram | Porur | Chennai - 600 077.
Mob.: +91 770 893 2211 | 735 837 6115
Email : avmlabsales@gmail.com | avmlabs21@gmail.com | www.avmgroup.biz



AVM/C/202088

CALIBRATION CERTIFICATE

ULR No.: CC208821000023045F

CERT No. : C-21/22-0735-02	DATE OF ISSUE : 4-Oct-21	PAGE : 1 OF 1
JOB No. : J-21/22-0735	CALIBRATED ON : 1-Oct-21	SUGG.NEXT CAL : 1-Oct-22

DETAILS OF CUSTOMER :

NAME & ADDRESS : M/s. EL-EVEN LABORATORIES NO. 37/B2, KIZHAKKEKARA LANE, CHANGAMPUZHA NAGAR P.O, SOUTH KALAMASSERY, ERNAKULAM - 682 033.	CALIBRATED AT : IN LAB DATE OF RECEIPT : 1-Oct-21 CONDITION OF RECEIPT : SATISFACTORY CALIB. PROCEDURE : AVM-WI-M13
---	--

DETAILS OF DEVICE UNDER CALIBRATION :

NOMENCLATURE : MICROPIPETTE	RANGE / SIZE : 100 ~ 1000 µl
MANUFACTURER : ERBAPETTE	OPERATING RANGE : -
MODEL : -	INPUT / OUTPUT : -
TYPE : -	RESOLUTION : 1µl
SERIAL NUMBER : 11614816	ACCURACY : -
ID/TAG NO. : E-006	LOCATION : -

ENVIRONMENTAL CONDITIONS:

TEMPERATURE : (23 ± 3) °C	HUMIDITY : (50 ± 10) %RH
---------------------------	--------------------------

MEASUREMENT TRACEABILITY

The reference measurement standards used are traceable to National/International Standards.

REFERENCE STANDARDS	SERIAL / TAG NUMBER	CERTIFICATE NUMBER	DUE DATE
ELECTRONIC SEMI MICRO BALANCE	327560/11 / BI-0094	C-AVM-IH-21/22-019	17-Jul-22

CALIBRATION RESULTS

Nominal Capacity	Measured Capacity (Mean)	Error	Expanded Uncertainty (±)
µl	µl	µl	µl
100	99.93	-0.07	0.13
500	499.66	-0.34	0.74
1000	999.89	-0.11	0.74

The reported Expanded Uncertainty is calculated at 95.45% Confidence Level with a coverage factor of k = 2.

END OF CERTIFICATE

CALIBRATED BY:

S MANOJ
CALIBRATION ENGINEER

REVIEWED & APPROVED BY :

S.JAGATHEESWARI / AVICHI MEIYAPPAN
AUTHORISED SIGNATORY