


AVM LABS PVT. LTD.

No. 49, 3rd Street | Moorthy Nagar | Chettiar Agaram | Porur | Chennai - 600 077.
Mob.: +91 770 893 2211 | 735 837 6115
Email : avmlabsales@gmail.com | avmlabs21@gmail.com | www.avmgroup.biz



C G VIM / 2 / 018 8

CALIBRATION CERTIFICATE

ULR.NO: CC208821000023049F

CERT No.	: C-21/22-0735-06	DATE OF ISSUE	: 4-Oct-21	PAGE	: 1 OF 1
JOB No.	: J-21/22-0735	CALIBRATED ON	: 1-Oct-21	SUGG.NEXT CAL	: 1-Oct-22

DETAILS OF CUSTOMER :

NAME & ADDRESS	: M/s. EL-EVEN LABORATORIES NO. 37/B2, KIZHAKKEKARA LANE, CHANGAMPUZHA NAGAR P.O, SOUTH KALAMASSERY, ERNAKULAM - 682 033.	CALIBRATED AT	: ON SITE
		DATE OF RECEIPT	: -
		CONDITION OF RECEIPT	: SATISFACTORY
		CALIB. PROCEDURE	: AVM-WI-T07

DETAILS OF DEVICE UNDER CALIBRATION :

NOMENCLATURE	: DRY BATH	RANGE / SIZE	: Amb ~ 100 °C
MANUFACTURER	: LABTECH MEDICO	OPERATING RANGE	: -
MODEL	: C-POULO	INPUT / OUTPUT	: -
TYPE	: -	RESOLUTION	: 0.1°C
SERIAL NUMBER	: 4392	ACCURACY	: -
ID/TAG NO.	: E-010	LOCATION	: -

ENVIRONMENTAL CONDITIONS:

TEMPERATURE :	(25 ± 10) °C	HUMIDITY :	(50 ± 20) %RH
---------------	--------------	------------	---------------

MEASUREMENT TRACEABILITY

The reference measurement standards used are traceable to National/International Standards.

REFERENCE STANDARDS	TAG NUMBER	CERTIFICATE NUMBER	DUE DATE
RTD TEMPERATURE SENSOR (PT-100) 4 WIRE	10570 / BL-0168	TL/021/224.1.1	01-Mar-22
TEMPERATURE CALIBRATOR	16091609 / BL-0312	TSC/21-22/14644-12	13-Jan-23

CALIBRATION RESULTS

TEST VALUE (DUC)	STD VALUE (REF) (MEAN)	ERROR	EXPANDED UNCERTAINTY (±)
°C	°C	°C	°C
37.0	37.9	-0.9	0.67

The reported Expanded Uncertainty is calculated at 95.45% Confidence Level with a coverage factor of k = 2.

END OF CERTIFICATE

CALIBRATED BY:

S. Manoj
S MANO

CALIBRATION ENGINEER

REVIEWED & APPROVED BY :

S. Jagatheeswarthi
S.JAGATHEESWARTI / AVICHI MEIYAPPAN

AUTHORISED SIGNATORY