



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF BLOOD MIXER

Certificate No:	21900002460	Issued On:	21/07/2021
ULR No:	CC214621900002460F	Calibration Date:	19/07/2021
Job Identification No:	ICS/C/AS/07/2460	Next Calibration Date:	18/07/2022
Ref. No:	SRF, Dated - 16/07/2021		

CALIBRATED FOR: M/s Arth Diagnostics (P) Ltd.
4C Apex Chamber, Behind Lok Kala Mandal,
Opp. Canara Bank, Madhuban, Udaipur, (Rajasthan-313001).

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Blood Mixer		
Make/Model	Top Tech	Visual Inspection	Ok
Serial No.	0931	I.D. No.	ADPL/MIXER/01
Range	At 30 RPM	Least Count	---
Location	Hematology	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National Standard (NPL, New Delhi)

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	Lutron	C-210615-2-1	Godrej, Chennai	16/06/2021	16/06/2023

USED EQUIPMENT DETAILS Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

U.U.C. SPEED KNOB' POSITION	CORRECTED VALUE OF TACHOMETER READING (AVG.)	Uncertainty at approx 95% C.L and coverage factor k=2
Reading # 1	30.3	± 2
Reading # 2	30.2	
Reading # 3	29.8	
Reading # 4	29.9	
Reading # 5	30.1	

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.

2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.

3. Total liability of our Organisation is limited to the invoiced amount.

4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.

5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.

6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory
RAJESH KAPOOR
Operation-Head