



Calibration Certificate

This is to certify that the **Cobas E411** is a fully automated analyzer bearing the Serial Number **0936 -02** at Likhitha's Diagnostics and Speciality Lab, ZenRise, Madinaguda, Hyderabad – 500049, has been calibrated on **19.08.2021**.

The calibration includes:

Adjustments: Checked and adjusted sample/Reagent Rotor, Incubator, Sample/Reagent probe, Bead Mixer and their respective home positions.

Incubation Disk: Range: 36.8°C to 37.2°C
Adjusted: 37°C

Detection Unit: Target: 28.0°C
Adjusted: 28.0°C

PC/CC: Target: 28°C
Adjusted: 28°C

Adding to this all the temperatures, PMT Voltage and volumes drawn by all pumps were checked and found they are OK.

Next calibration due on 18.08.2022

For Roche Diagnostics India Pvt Ltd.,

T R Siddardha,
Sr.Technical Service Specialist,
Hyderabad.

cobas[®]
Life needs answers

cobas[®] e411 instrument

Qualification Service
Installation Qualification / Operation Qualification (v.1.0)





cobas[®] e411 instrument



General Information

Country: INDIA

Customer Name: LIKHITHA'S DIAGNOSTICS & SPECIALITY LAB

Customer Address: Likhithas Diagnostics and Speciality Lab, ZenRise CRO, Plot No. 201, Mythrinagar, NH – 65, Madinaguda, Miyapur, Hyderabad - 500049.

Person Responsible for Quality Assurance: Mr. Hanmant Jadhav

System Information

Cobas e411 Disk

	S/N	IP Address
cobas e411	0936 -02	182.18.38. 220

cobas link: SCL NA

Host provider: NA

User Software Version: 02 - 08.

Installation Information

Installation Start Date: 16.08.2021

First Installation: YES

Relocation: From: To:

Roche Responsible Representative : Mr. Siddartha TR, Sr. Technical Service Engineer





Installation Qualification:

This document forms the basis of the Qualification Services Certificate. It certifies that the instrument is installed according to the manufacturer's specifications. The report presents and documents the test procedures, the documentation, reference and acceptance criteria used to verify that the system is installed according specifications. The report demonstrated that all installation qualification criteria have been met satisfactorily.

Notice: The following tests are to be carried out by trained Roche personnel only.

Purpose: The purpose of this test is to confirm that the instrument was delivered undamaged and installed correctly.

Test #	Test	Pass Fail	Signature Date
IQ.1.1	Operator's Manual available	Pass	 D. Thompson
IQ 1.2	Environmental parameters met	Pass	
IQ 1.3	Instrument delivered undamaged and complete	Pass	
IQ 1.4	Transport locking successfully removed	Pass	
IQ 1.5	All connections correctly installed	Pass	
IQ 1.6	Instrument positioned according to Installation Manual	Pass	
IQ 1.7	Instrument boot process successfully	Pass	
IQ 1.8	Checksum according to specification	Pass	
IQ 1.9	Mechanical adjustments complete	Pass	
IQ 1.10	Auxiliary components positioned	Pass	
IQ 1.11	Instrument installation check	Pass	
IQ 1.12	Host communication settings checked	not applicable	

Test #	Test	Pass Fail	Signature Date
IQ.3	Installation Qualification for all cobas <e411>	yes	 [Signature]



Deviation Report: Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1	
NA	
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify

Deviation #2	
NA	
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify

Deviation #3	
NA	
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify



Operational Qualification:

This document is the basis of the Qualification Service Certificate. It certifies that the instrument is operating according to the manufacture's specifications. This report presents and documents the test procedures, documentation, references and acceptance criteria used to verify that the specified system is operating according the specifications. The report demonstrates that all operational qualification criteria have been met satisfactorily.

Notice: The following tests are to be carried out by trained Roche personnel only.

Purpose: The purpose of this test is to check that the modules are operating in accordance with the

Test #	Test	Pass Fail	Signature Date
OQ.1	Calibration successfully	Pass	
OQ.2	Quality Control successfully	Pass	
OQ.3	Accuracy check successfully	Pass	

Deviation Report: Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1
NA
Investigation
Action taken
Deviation resolved satisfactorily? specify

Deviation #2
NA
Investigation
Action taken
Deviation resolved satisfactorily? specify



Conclusion

All test results are acceptable. yes

Any deviation or non-conformances observed have been recorded as a deviation and the relevant forms completed. no

All acceptance criteria have been met. This equipment is deemed acceptable and the unit is approved for its intended use. yes

Comments

All available parameter calibrations were passed. QC values were within the range.

Completed by Roche Representative Date 08 / 16/ 2021

Print Name Mr. Bhavani Prasad Varupula Signature 

Reviewed by Customer Contact Date 08 / 16/ 2021

Print Name Mr. HANMANT JADHAV Signature 

Reviewed by Customer Quality Assurance Date 08/16/2021

Print Name Mr. HANMANT JADHAV Signature 

cobas[®]
Life needs answers

cobas[®] e411 instrument

Qualification Service
Installation Qualification





Installation Qualification for cobas e411®

Description

IQ.3.1	Function check of e411 according to specifications	
	The e411 is installed according to the installation manual and using official tools	Pass
IQ.3.2	Mechanical adjustments complete	
	All mechanical adjustments for the different e411 mechanical parts are carried out	Pass
	Adjustment check during mechanical check function	Pass
	Sample LLD adjusted to spec.	Pass
IQ.3.3	Auxiliary components positioned	
	Sample probe, reagent probe, beads mixer paddle and measuring cell installed	yes
	Waste liner, CC/PC cups, CleanCell, ProCell positioned	yes
	System prime and system air purge for syringes and reagents	Pass
IQ 3.4	Instrument installation check	
	Incubator temperature 37°C ± 0.3°C	Pass
	No alarms during check	Pass
	System Volume Check (attached printout)	Pass
	PMT HV adjustment (attached printout)	Pass
	Assay Performance Check (AM+TSH, attached printout)	Pass
	Blank Cell calibration (attached printout)	Pass
	Mechanical adjustments parameter backed up	Pass



Deviation Report: Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1	
NA	
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify

Deviation #2	
NA	
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify

Deviation #3	
NA	
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify

cobas[®]
Life needs answers

cobas[®] e411 instrument

Qualification Service
Operation Qualification (v.1.0)





Operational Qualification:

Notice:

The steps described in OQ.1 have to be carried out after a new system installation and after any repair action which requires additional calibration.

If the service action does not affect the measurement performance, only apply steps OQ.2 and OQ.3 of the Operation Qualification.

Description

OQ 1	Calibration
	Calibration of all Immuno parameters successful (attached printout) yes
OQ.2	Quality Control
	Specify the type of control used: 3rd Party control (Bio-Rad)
	QC of Immuno parameters within acceptable range (attached printout) yes
OQ.3	Accuracy check for Immunology Assays
	Perform test with analytical reagents
	Any non-infectious disease assay (e.g. TSH) Number of det. 21
	Sample solution: Applicable control material Fill 21 Hitachi cups with control material and perform 21 determinations of each
	Accuracy check for immunoassays was within acceptable range Yes

Deviation Report: Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1	NA
Investigation	
Action taken	
Deviation resolved satisfactorily?	specify



Attachments

1. Calibration Printout
2. QC Printout
3. Precision Data

cobas[®] e411 instrument

Qualification Service
Attachments



Likhitha's Diagnostics and Speciality Lab, ZenRise, Madinaguda, Hyderabad

Cobas E411 - Intrassay Precision Study

Date : 09.02.2022

Sample ID - Pooled Human Serum Sample


Analyte - Thyroid Stimulating Hormone

Replicate No.	Results
1	5.55
2	5.51
3	5.63
4	5.76
5	5.67
6	5.58
7	5.44
8	5.66
9	5.78
10	5.52
11	5.63
12	5.51
13	5.53
14	5.35
15	5.51
16	5.69
17	5.69
18	5.65
19	5.67
20	5.72
21	5.4
Mean	5.59
SD	0.11
%CV	2.03

Processed By : Lab Technician

Mr. Sivashankar

Verified By : Quality Manager


02/09/2022

Result Report

Operator-ID: NELIMA

09/02/2022, 18:03

Sample/Control ID	Seq.	Pos.	Test	Dil	RPos	Result	Flag
000015142304	3523	0-21	TSH 0		3	5.55	H
000015142304	3522	0-20	TSH 0		3	5.59	H
000015142304	3521	0-19	TSH 0		3	5.63	H
000015142304	3520	0-18	TSH 0		3	5.76	H
000015142304	3519	0-17	TSH 0		3	5.67	H
000015142304	3518	0-16	TSH 0		3	5.58	H
000015142304	3517	0-15	TSH 0		3	5.44	H
000015142304	3516	0-14	TSH 0		3	5.66	H
000015142304	3515	0-13	TSH 0		3	5.78	H
000015142304	3514	0-12	TSH 0		3	5.52	H
000015142304	3513	0-11	TSH 0		3	5.63	H
000015142304	3512	0-10	TSH 0		3	5.51	H
000015142304	3511	0-9	TSH 0		3	5.53	H
000015142304	3510	0-8	TSH 0		3	5.35	H
000015142304	3509	0-7	TSH 0		3	5.51	H
000015142304	3508	0-6	TSH 0		3	5.69	H
000015142304	3507	0-5	TSH 0		3	5.69	H
000015142304	3506	0-4	TSH 0		3	5.65	H
000015142304	3505	0-3	TSH 0		3	5.67	H
000015142304	3504	0-2	TSH 0		3	5.72	H
000015142304	3503	0-1	TSH 0		3	5.40	H
000926896900	3502	0-4	T3 0 T4 1 TSH 0		1 2 3	1.09 6.43 2.45	
000926897000	3501	0-3	TSH 0		3	9.60	H
000910308103	3500	0-2	A-HCVII 0		14	0.050	
000014942704	3499	0-1	IGE 0		13	94.92	
000014942904	3498	0-20	TSH 0		3	1.72	
000910308503	3497	0-19	TSH 0		3	6.79	H

P-7
 PKO-10P
 Cortisol
 progesterone
 HAV-IgM
 Insulin
 A19.9
 Trop-T
 IPTH
 BeAg
 BeAb
 -peptide
 IV
 BeAg
 EV

Calibration

QC

Quality

04/04/2022 (Mon)

18:56

Calibration Trace


Test: **HVCOMPT 0**

Scale Mode: **Automatic**

Period: **29/10/2021** - **08/03/2022**

Calibrator	No.	Signal	Alarm	Date	Time
Std(Min.)	8	2391		08/03/2022	11:49
Std(Max.)	8	28298		08/03/2022	11:49

Reagent Lot No.	RP No.
00561758	009950



Navigation controls: back, left, right, forward arrows.

Scale

Comment

Close

Stop
 Logout
 S.Stop
 Stat Mode
 Alarm
 Print
 Start

Calibration

QC

Utility

Method

Stop

Calibration Result

Test	Calibration Type	Unit	Date Time	Calibrator Lot	Reagent Lot	RP No.
HIVCOMPT 0	Cutoff	COI	08/03/2022 11:49:43	00561758	00561758	032543

L-Calib. was generated!

	Level1	Level2
Signal1	2418	28014
Signal2	2365	28582
Dupl.	---	---
Sys.Err.	---	---
Slope	---	
Diff	---	
Cutoff	1402	
Border Area	0.900 - 1.00	

Close

Logoff

S.Stop

Stat
Mode

Alarm

Print

Start

P07
 PKO - NP
 Cortisol
 progesteron
 HAV-IgM
 Insulin
 CA199
 Trop-T
 IPTH
 HBeAg
 HBeAb
 c-peptide
 HIV
 HBSAg
 HCV

Calibration

GC

Unit

Calibration Trace


Test

HBSA,all 0

Scale Mode Automatic

Period 06/01/2021 - 14/12/2021

Calibrator	No.	Signal	Alarm	Date	Time	Reagent Lot No.	RP No.
Std(Min.)	23	964.5		14/12/2021	15:18	00548342	025277
Std(Max.)	23	4315		14/12/2021	15:18		



Scale

Comment

Close

Start

Stop

Logout

S.Stop

Stat Mode

Alarm

Print

Calibration

QC

Utility

Calibration Result

Test	Calibration Type	Unit	Date Time	Calibrator Lot	Reagent Lot	RP No.
HBSAGII 0	Cutoff	COI	15/11/2021 11:54:01	00559215	00559215	015023

L-Calib. was generated!

	Level1	Level2
Signal1	972.1	4142
Signal2	986.9	4369
Dupl.	----	----
Sys.Err.	----	----
Slope	---	
Diff	----	
Cutoff	339.9	
Border Area	0.900 - 1.00	

Close

Stop

Logout

S.Stop

Stat Mode

Alarm

Print

Start

P-7

PKO-ENP

Corn'sol
progesterone

HAV-IgM

Insult.

CA19.9

Trop-T

IPTH

HBeAg

HBeAb

C-peptide

HIV

HBSAg

HCV

Calibration

QC

Utility

Stop

Calibration Trace

Test

A-HCVII 0

Scale Mode Automatic

Period 31/03/2018 - 08/03/2022



Calibrator	No.	Signal	Alarm	Date	Time
Std(Min.)	50	1097		08/03/2022	11:37
Std(Max.)	50	60810		08/03/2022	11:37

Reagent Lot No.

RP No.

00578749

003030



Scale

Comment

Close

Stat
Mode

Alarm

Print

Start

bmserv

04/04/2022 (Mon)

18:55

Calibration

QC

Utility

INSTALL

Stop

Calibration Result

Test	Calibration Type	Unit	Date Time	Calibrator Lot	Reagent Lot	RP No.
A-HCVI0	Cutoff	COI	08/03/2022 11:37:49	00578749	00578749	044100

L-Calib. was generated!

	Level1	Level2
Signal1	1113	60553
Signal2	1081	61067
Dupl.	----	----
Sys.Err.	----	----
Slope	----	
Diff	----	
Cutoff	9731	
Border Area	0.900 - 1.00	

Close

Logout

S.Stop

Stat Mode

Alarm

Print

Start

P. 7
 PKO - NP
 Corh'501
 progettemou
 HAV-3gm
 Insultu
 CA19-9
 Trop-T
 IPTH
 HBeAg
 HBeAb
 c-peptide
 HIV
 HBsAg
 HCV