

## Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,  
Andheri East, Mumbai - 400059  
T : +91 22 6787 9300  
F : +91 22 6787 9333

# Calibration Certificate

The below mentioned instrument has been calibrated and tests performed to check the system performance.

Instrument : VITROS V5600

Serial no : 5600 0703

Customer Name : Central Clinical Laboratory Pravara Rural Hospital-Loni

Calibration performed on : 29/11/2021

The system's calibration includes optics calibration and checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.

Next Calibration will be performed in **MAY 2022**.

Ortho Clinical Diagnostics India Pvt Ltd.



Mitesh Shah

Date: 29/11/2021

Sr.Zonal Service Manager-Ortho Care  
Pune

# Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,  
Andheri East, Mumbai - 400059  
T: +91 22 6787 9300  
F: +91 22 6787 9333

## Calibration Certificate

The below mentioned instrument has been calibrated and tests performed to check the system performance.



Instrument : VITROS V5600  
Serial No : 5600 1041  
Customer Name : Central Clinical Laboratory Pravara Rural Hospital-Loni  
Calibration performed on : 26/11/2021

The system's calibration includes optics calibration and checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.



Next Calibration will be performed in **MAY 2022**.

For **Ortho Clinical Diagnostics India Pvt Ltd.**

  
**Mitesh Shah**

**Date: 26/11/2021**

**Sr.Zonal Service Manager-Ortho Care  
Pune**

# Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,  
Andheri East, Mumbai - 400059  
T: +91 22 6787 9300  
F: +91 22 6787 9333

## Calibration Certificate

The below mentioned instrument has been calibrated and tests performed to check the system performance.



Instrument : VITROS V5600  
Serial No : 5600 0703  
Customer Name : PRAVARA MEDICAL TRUST, LONI  
(BIOCHEMISTRY DEPT).  
Calibration performed on : 29/11/2021

The system's calibration includes optics calibration and checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.

Next Calibration will be performed in **MAY 2022**.


For **Ortho Clinical Diagnostics India Pvt Ltd.**

  
Mitesh Shah

Sr.Zonal Service Manager-Ortho Care  
Pune

Date: 29/11/2021

Approved By

  
S. S. J.

Incharge  
CCL- Biochemistry  
PMT, RMC, Loni

# Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,  
Andheri East, Mumbai - 400059  
T: +91 22 6787 9300  
F: +91 22 6787 9333

## Calibration Certificate

The below mentioned instrument has been calibrated and tests performed to check the system performance.



Instrument : VITROS V5600  
Serial No : 5600 1041  
Customer Name : PRAVARA MEDICAL TRUST, LONI  
(BIOCHEMISTRY DEPT.)  
Calibration performed on : 26/11/2021

The system's calibration includes optics calibration and checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.

Next Calibration will be performed in **MAY 2022**.

For **Ortho Clinical Diagnostics India Pvt Ltd.**

**Mitesh Shah**

**Sr.Zonal Service Manager-Ortho Care  
Pune**

**Date: 26/11/2021**

*Approved By*  
*S.P. J...*

Incharge  
CCL- Biochemistry  
PMT, RMC, Loni

# Service Report

Ortho Clinical Diagnostics

INDIA Pvt. Ltd.

SOP-OCDIN-019/F02  
Revision-02

Report No: 3624

Customer's Name: <u>Pravara Medical Trust</u> Contact Person: <u>Dr. Jangle (Biochemistry Lab)</u> Address: <u>At post Loni, Tal. Rahata,</u> <u>Dist. Ahmednagar,</u>	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%; padding: 5px;"> <b>SOLD</b> <input type="checkbox"/>  <input type="checkbox"/> Warranty  <input type="checkbox"/> AMC/CMC  <input type="checkbox"/> Charge Call                         </td> <td style="width: 50%; padding: 5px;"> <b>RRC</b> <input checked="" type="checkbox"/> </td> </tr> </table>	<b>SOLD</b> <input type="checkbox"/> <input type="checkbox"/> Warranty <input type="checkbox"/> AMC/CMC <input type="checkbox"/> Charge Call	<b>RRC</b> <input checked="" type="checkbox"/>
<b>SOLD</b> <input type="checkbox"/> <input type="checkbox"/> Warranty <input type="checkbox"/> AMC/CMC <input type="checkbox"/> Charge Call	<b>RRC</b> <input checked="" type="checkbox"/>		

Nature of Call:

<input type="checkbox"/> Service Call	<input type="checkbox"/> Follow up	<input checked="" type="checkbox"/> Modification	<input type="checkbox"/> Application	<input checked="" type="checkbox"/> Preventive Maintenance
<input type="checkbox"/> Telephone	<input checked="" type="checkbox"/> eConn Proactive	<input type="checkbox"/> UPS	<input type="checkbox"/> Printer	<input type="checkbox"/> Other _____

Observed damage before service:	Instrument Particulars:
NA	Model <u>Vitros 5600</u>
	J Number <u>J56001041</u>
	Serial Number <u>56001041</u>

Problem reported:	Call Details : Status after call :
Error Code : <u>xpm; Ecompres; Eintellicheck; xmf6; xmf6</u> Description : <u>Preventive maintenance;</u> <u>Econ alert for compressor &amp; intellicheck.</u> <u>modification # FG software update 3.7.1</u> <u>modification # EG wear pads.</u>	Functioning well <input checked="" type="checkbox"/> Incomplete but operative <input type="checkbox"/> Needs spares <input type="checkbox"/> Complaint forwarded to L.S. <input type="checkbox"/> Requires follow-up <input type="checkbox"/> Complaint escalated to _____ <input type="checkbox"/>

Call Details :	METER COUNT
Complaint Forwarded by : <u>SELF</u> Complaint Call Date : <u>22/11/2021</u> Time: <u>20:15</u> Travel Duration (To & Fro): <u>2</u> Hr. Call Attended Date : <u>25/11/2021</u> Time: <u>17:10</u> Start Time : <u>25/11/2021 17:15</u> End Time : <u>26/11/2021 19:00</u> Actual Work Hours: <u>14 Hr. 15 mins. + 2 Hrs.</u> <input type="checkbox"/> Solved over Phone. Date : _____ Time: _____	SO No. <u>98191413</u>

**Diagnosis :**

software update v.3.7.1, mod# EG, replace air filter (convistor) & preventive maintenance.

**Action taken / Action required:**

Done backup. Updated slw version to 3.7.1. Cleaned the system. Cleaned all the modules. Done adjustments & performance tests. Run maintenance pack. Requested user to run control & samples.

For Customer Only :	For Engineer Only :
Job complete <input type="checkbox"/> Job incomplete <input type="checkbox"/> Satisfied <input type="checkbox"/> Unsatisfied <input type="checkbox"/> Comments: _____ Customer's Signature & Date: <u>[Signature] 29/11/21</u>	Engineer's Name : <u>Mitesh C. Shah / Kedar More</u> Engineer's Code : <u>AINFE 09 / AINFE 10</u> Comments : <u>System working ok.</u> Engineer's Signature & Date: <u>[Signature] 26/11/2021</u>

Approved By [Signature]  
**Incharge**  
**CCL- Biochemistry**  
**PMT, RMC, Loni**

Ortho Clinical Diagnostics


**INSTALLATION QUALIFICATION**

For

**VITROS® 5600 INTEGRATED SYSTEM**

---

VITROS<sup>®</sup> System  
Integrated | 5600



Manufactured by:

Ortho Clinical Diagnostics, Inc., US


## Table of Contents

<b>Sr. No.</b>	<b>Contents</b>	<b>Page No.</b>
I	Approval of the IQ procedure	3
II	Instructions	4
III	Scope	5
IV	Ancillary Information	6
V	Installation Qualification	8
VI	Installation Procedure	10
VII	Installation Report	15
VIII	Comments	16
IX	System Certification	17
	Appendix	
	I. Installation Certificate	


**I. APPROVAL OF THE IQ PROCEDURE:**

Both Central Clinical Laboratory Pravara Rural Hospital and Ortho Clinical Diagnostics are responsible for installation of VITROS® 5600 Immunodiagnostic System bearing Sr. No **J56001041** in Central Clinical Laboratory, Pravara Rural Hospital as per the attached protocol.


Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : *Mitesh Shah* Signature:   
Designation : Sr. Zonal Manager Date: 05/12/2020  
Company : Ortho Clinical Diagnostics

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : *Mr. K. N. Hotkar* Signature:   
Designation : *Asso. Prof.* Date: *05/12/2020*

Department : *Biochemistry*

Name : *Mr. Bhawar S. R.* Signature:   
Designation : *Lab Tech.* Date: *05/12/2020*

Department : *Biochemistry*

Customer Authorizations:

Name : *Dr. S. N. Jangle*  
Designation : *Professor and Head, Biochemistry*  
Signature : *S. N. Jangle*  
Date : *05/12/2020*

**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LON-413736, (M.S.) INDIA**



**II. INSTRUCTIONS:**

1. This document is to be completed at the time the system is installed to its location and set up for operation.
2. An authorized Ortho Clinical Diagnostics representative will check the system and enter the specific data as outlined in the appropriate Installation Qualification.
3. Employees of **Central** Clinical Laboratory Pravara Rural Hospital will verify each result and sign in the each page. The member of the validation team will carry out this procedure.
4. ~~ALL deviations from normal specification during installation will be noted under COMMENTS.~~  
All resolution to such problems will also be noted in the COMMENTS section. Additional space is provided at the end of this protocol for the same.

### III. SCOPE

This Installation Qualification protocol will be performed on the VITROS® 5600 Immunodiagnostic System, and the Sr. No. **J56001041** located at Central Clinical Laboratory Pravara Rural Hospital.

This protocol will define the documentation that will be used to evaluate the instrument's installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage.

~~Installation checks will also be performed to verify that the instrument has been installed with proper connections and utilities.~~

Trained, knowledgeable personnel will perform Installation Qualification studies.

Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated, and the appropriate course of action determined. All documents will be initialed and dated.

**IV. ANCILLARY INFORMATION**

a. Certification of Purchase Order Compliance

I certify to the best of my knowledge, the instrument VITROS® 5600 Immunodiagnostic System and Sr. No. **J56001041** installed on 05/12/2020.

Verified By: Mitesh Shah

Date:05/12/2020

b. Utilities

Sr. No	Utility	Verified by & date
1.	<b>Environmental condition:</b> As per requirement (To be free from Dust, Electrical & magnetic Interferences and free from vibration)	Yes / No Mitesh Shah 05/12/2020
2.	<b>Adequate space for installation:</b> (Length 110 inches x Width 35 inches x Height 84 inches)	Yes / No Mitesh Shah 05/12/2020
3.	<b>Electrical Outlets:</b> Actual Voltage on site [200 Vac – 240 Vac] <b>Electrical Input:</b> Voltage supplied through ON LINE UPS (232Vac @ 50Hz frequency, Earthing < 2.0Vac)	Yes / No Mitesh Shah 05/12/2020
4.	<b>Capacities:</b> <ul style="list-style-type: none"> <li>• 90 samples (80 Routine positions &amp; 10 STAT positions are available)</li> <li>• 150 Reagent Positions are available.</li> </ul>	Yes / No Mitesh Shah 05/12/2020
5.	<b>Temperature:</b> 15° C to 30° C 15% to 75% relative humidity	Yes / No Mitesh Shah 05/12/2020

## Ortho Clinical Diagnostics

The instrument has been verified for the following:

Sr. No.	Verification	Yes / No	Verified by & date
1.	Equipment is identified	Yes / No	Mitesh Shah 05/12/2020
2.	Manufacturer's specifications are included	Yes / No	Mitesh Shah 05/12/2020
3.	Accessories / Consumables are listed	Yes / No	Mitesh Shah 05/12/2020
4.	Equipment manual from the manufacturer is documented	Yes / No	Mitesh Shah 05/12/2020
5.	Manufacturer's Certificate of compliance attached	Yes / No	Mitesh Shah 05/12/2020

## V INSTALLATION QUALIFICATION:

### A. Equipment Description

The VITROS® 5600 Integrated System is a Random access, walk away system intended for use in the *in vitro* quantitative, semi-quantitative, and qualitative measurement of a variety of analytes of clinical interest, using VITROS Chemistry Products MicroSlides, VITROS Chemistry Products MicroTip reagents and VITROS Immunodiagnostic Products Reagents.

Instrument Identification		Verified by	Date
Equipment Name :	<b>Automated Integrated System</b>	Mitesh Shah	05/12/2020
Model :	VITROS® 5600	Mitesh Shah	05/12/2020
Manufacturer :	Ortho Clinical Diagnostics, Inc., US	Mitesh Shah	05/12/2020
Marketed by :	Ortho Clinical Diagnostics India Pvt. Ltd.	Mitesh Shah	05/12/2020
Serial Number :	56001041	Mitesh Shah	05/12/2020
Lab Id :	Biochemistry Analyzer No. 2	Mitesh Shah	05/12/2020
Software Name :	QNX	Mitesh Shah	05/12/2020
Software Version :	V	Mitesh Shah	05/12/2020
Size (in inches) :	Adequate for installation: (Length 170 x Width 83 x Height 84).	Mitesh Shah	05/12/2020
Power :	1600W@ 50Hz of 220Vac – 240Vac	Mitesh Shah	05/12/2020

## Ortho Clinical Diagnostics

### B. Accessories/Consumables

The following accessories were supplied with the instrument. Check (√) 'verified by' in case they are found to be in order.

Description	Quantity	Verified	Date
User Training Manual	1	√	05/12/2020
Application Software-Revelation	1	√	05/12/2020
Universal Sample Tray	9	√	05/12/2020
Backup DVD R/w	3	√	05/12/2020
Printer Cable	1	√	05/12/2020
Printer Software	1	√	05/12/2020
Power Cords	3	√	05/12/2020
Printer	1	√	05/12/2020
Air filter	1	√	05/12/2020
Waste can 5L	1	√	05/12/2020

### C. List of Manuals :

Ortho Clinical Diagnostics has supplied following manual.

**D. Maintenance :**

The instrument listed within this document will be placed under the control of the purchasing institution with respect to proper maintenance procedures as detailed in the Training and Reference Guide. The maintenance procedures will be filed separately.

A trained analyst using the manuals provided with the instrumentation can perform simple maintenance. Upon expiration of the warranty period Ortho Clinical Diagnostics offers several levels of Maintenance Agreements and Performance Testing services to assist you in maintaining GLP/GMP compliance. Contacting your local representative and requesting the additional Service Agreement can supply additional information.

## VI. INSTALLATION PROCEDURE

(The following steps Performed at the time of original installation at the initial location)

1. Locating & unpacking the instrument.
2. Rearfix/verify the circuits boards & CPU
3. Nominal Line voltage frequency selection of transformer.
4. Load Supply & Power ON the system.
5. System Configuration.
6. Systems Tests & Adjustments.
7. Subsystems Performance Verification & calibration
8. Setting and installing printer.

The Above-mentioned steps has completed successfully by trained field Engineer as described below.

### VI.1 Locating & unpacking the VITROS® 5600 instrument:

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- Check the Tip & Tell Label.
- Verify the serial no / J number of the system match those indented for delivery.
- Place the pallet in a position with a minimum of 7 m (24 ft.) of clearance in front of the end with the LABELS
- Remove the STRAPPING and the RAMPS from the pallet..
- Assemble the 2 piece RAMP and hok the RAMPS to the end of the pallet
- Move the system down the ramps. Do the same for the second half of the system.



- Place the instrument in the lab leveled floor.
- Join the two halves of the system as per manufacturer's instructions.
- Remove the packing material from
  - VERSATIP supply carousel
  - Under side of sample supply cover
  - SR metering Nozzle
  - Well Wash Nozzles
  - Beneath SR Pumps
  - Luminometer & Micro ImmunoAssay VERSATIP ring
  - Supply 4 load doors & Reagent Well shuttle
  - SR carousel
- Remove the wire Tie, tape & Foam from UIA REAGENT METERING ARM.
- Remove the moisture separators behind the compressor installed on it bracket.
- Remove the Foam supporting from the compressor.

**VI.2 Reaffix/verify the circuits boards & CPU:**

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- Open the right side front door and open the card rack metal cover.
- Remove the RC labeled DSP boards and reaffix it back.
- Open the right side rear panel and open the card rack metal cover.
- Remove the UC labeled DSP boards and reaffix it back.
- Open the Middle front door and open the CPU top cover & verify the boards.

**VI.3 Nominal Line voltage / frequency selection of transformer:**

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- Connect the Primary T2\_H6 on Label no H6 for 0Vac.
- Connect the Primary T2\_H# on label no H2 for 230Vac.
- Connect the Secondary T2\_X1 on label no X1 for 230Vac.
- Connect the Secondary T2\_X# on label no X2 for 200Vac.

- Connect the Secondary T2\_X3 & T2\_X3 on label no X3 for 0Vac.

#### VI.4 Load Supply & Power ON the system:

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- VersaTip & Sample Trays.
- Signal Reagent & Universal Reagent.
- Check the supply and Earthing voltage.
- Switch ON the Instrument.

#### VI.5 System Configuration.: (\6902906\_3600-RefGd\_Ltr-EN.pdf)

Click on Set Access Level from Status Menu and type password

Option -> Configure System ->

- **Configure Current Date & Time:**

Select the format and set the **Date and Time**.

- **Configure System Name & J Number**

Enter the System Name & J number.

- **Configure the Screen Saver**

Set the Screen saver delay time.

- **Configure the Site Temperature**

Set the site temperature tolerance for the nominal site temperature.

Status -> Diagnostics -> Select the required task

- **Touch Screen Calibration:**

Touch "Calibrate Touch screen" at the bottom of the DIAGNOSTICS Menu.

Touch center of the target appear on the screen.

When you have finished, touch "Save Calibration"

- **Country Code Selection:**

Touch "Diagnostics" then Select "V-Docs"

Press [Alt] and [S] to Access the service Scripts.

Select "Configure Country Code"

Select the appropriate country from the List.

Select "Set country code"

Touch "Return".

Touch "Shut Down". Configure the Language from the System Menu button.

Touch "Final Shutdown". Then reset the system.

#### **VI.6 System Tests and Adjustments: (\6902906\_3600-RefGd\_Ltr-EN.pdf)**

Adjustments are diagnostic functions used to fine-tune or define various system Parameters to ensure proper system performance. With the exception of the IRS Calibration, all other adjustments are available only to trained service personnel.

- MicroImmunoassay ( $\mu$ IA) Metering
- MicroSensor
- MicroWell Reagent Metering
- MicroWell Wash Metering
- Signal Reagent Metering
- Luminometer
- Scrap Run

#### **VI.7 Subsystems Performance Verification & Calibration : (\6902906\_3600-RefGd\_Ltr-EN.pdfpdf)**

- Well Wash Dispense & Aspiration Calibration.
- Signal reagent Dispense calibration.
- 30PSI & 10PSI calibration.
- Soak Volume Verification
- IRS Calibration

**VI.8 Setting and installing printer:** (\.6902906\_3600-RefGd\_Ltr-EN.pdf)

- Remove the Packing material form the printer and assemble the accessories.
- Connect the USB cable and Switch ON the Printer.
- Set report control and print the test page.

**VII. Installation Report :**

Activity	Observation	Remarks	Verified By / Date
Locating & unpacking the instrument.	Instrument was located and unpacked	Ok	Mitesh Shah 05/12/2020
Reaffix/verify the circuits boards & CPU	Reaffixed/verified the circuits boards & CPU	Ok	Mitesh Shah 05/12/2020
Nominal Line voltage frequency selection of transformer.	Nominal Line voltage frequency was sated.	Ok	Mitesh Shah 05/12/2020
Load Supply & Power ON the system.	Supply Loaded & Powered ON the system.	Ok	Mitesh Shah 05/12/2020
System Configuration.	System was configured as per the requirement.	Ok	Mitesh Shah 05/12/2020
System Tests and Adjustments	System Tested and Adjustments done.	Ok	Mitesh Shah 05/12/2020
Subsystems Performance Verification & calibration	Subsystems Performance Verified & calibrated	Ok	Mitesh Shah

Ortho Clinical Diagnostics

	successfully.		05/12/2020
Setting and installing printer	Printer was installed and connected to the system.	Ok	Mitesh Shah 05/12/2020

VIII. COMMENTS :

Deviation:

Nil

Impact On Operation:

Nil

Corrective Action:


Nil

**IX. SYSTEM CERTIFICATION**

Study data has determined that the system described in this document either meets all criteria outlined in this Installation Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been / not been addressed. The system is ready for Operation Qualification.


Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Mitesh Shah  
Designation : Sr. Zonal Manager  
Company : Ortho Clinical Diagnostics


Signature:   
Date: 05/12/2020

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K.N. Hottkar  
Designation : ASSO. prof.  
Department : Biochemistry

Signature:   
Date: 05/12/2020

Name : Mr. Bhawar S.P.  
Designation : Sr. Lab. Tech.  
Department : Biochemistry

Signature:   
Date: 05/12/2020

Customer Authorizations:

Name : Dr. S.N. Jangle sir  
Designation : prof & Head. Biochemistry.  
Signature : S.N. Jangle  
Date : 5/12/2020


**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413736, (M.S.) INDIA**

Ortho Clinical Diagnostics

**OPERATIONAL QUALIFICATION**

For

**VITROS® 5600 INTEGRATED SYSTEM**



Manufactured by:  
Ortho Clinical Diagnostics, Inc., US

Table of Contents

<b>Sr. No.</b>	<b>Contents</b>	<b>Page No.</b>
I	Approval of the OQ procedure	3
II	Instructions	4
III	Scope	5
IV	Operational Qualification	6
IV.I	Operational Procedure	7
V	Operational Qualification Report	21
VI	Comments	23
VII	System Certification	24



**I. APPROVAL OF THE OQ PROCEDURE:**

Both Central Clinical Laboratory Pravara Rural Hospital and Ortho Clinical Diagnostics are responsible for installation of VITROS® 5600 Immunodiagnostic System bearing Sr. No **J5600-0703** in Central Clinical Laboratory, Pravara Rural Hospital as per the attached protocol.

Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Rathod Vijaykumar

Signature: 

Designation : Territory Manager

Date: 07/07/2018

Company : Ortho Clinical Diagnostics

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hotkga-


Signature: 

Designation : ASSO. PROF.

Date: 07/07/2018

Department : Biochemistry

Name : Mr. Bhawar S.R.

Signature: 

Designation : Sr. Lab. Technician


Date: 07/07/2018

Department :

Customer Authorizations:

Name : Dr. S. N. Jangle sir

Designation : Prof. & Head Biochemistry

Signature : 

Date : 07/07/2018

**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413736, (M.S.) INDIA**

**II. INSTRUCTIONS:**

1. This document is to be completed at the time the system is installed to its location and set up for operation.
2. An authorized Ortho Clinical Diagnostics representative will check the system and enter the specific data as outlined in the appropriate Installation Qualification.
3. ~~Employees of Central Clinical Laboratory Pravara Rural Hospital will verify each result and sign in the each page. The member of the validation team will carry out this procedure.~~
4. ALL deviations from normal specification during installation will be noted under COMMENTS. All resolution to such problems will also be noted in the COMMENTS section. Additional space is provided at the end of this protocol for the same.

**III. SCOPE**

This Operational Qualification protocol will be performed on the VITROS® 5600 Integrated System, bearing Sr. No. J56000703 located at Department of Biochemistry.

This protocol will define the documentation that will be used to evaluate the instrument's operational check in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been operated in accordance with the intended usage.

Operational checks will also be performed to verify that the Instrument has been operated with proper information / sequence and utilities.

Trained, knowledgeable personnel will perform operational Qualification studies.

Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

IV. OPERATIONAL QUALIFICATION:

A. Instrument Identification

Verified Date

1. Model Name VITROS® 5600 Integrated System 20/6/18

2. Serial Number J56000703 20/6/18

B. Following is a list of tests to be performed and verified:

Test No.	Test Name	Test purpose	Initial / Date
01	System centers Overview	To make the operator to identify the instrument subsystem.	2/7/18
02	Start up & Shutdown	To make the equipment ready for operation.	2/7/18
03	User Inter Face Overview	Different functionality of software utility available for the operator interaction.	2/7/18
04	Sample programming and Analysis	To process samples either by manual assigning or through LIS.	2/7/18
05	Performing Calibration	To calibrate the system for every new lot of assay or after calibration expiry.	2/7/18
06	Maintenance & System clean	To perform maintenance process to keep the system operating properly.	2/7/18
07	Reagent Management & supply	To update & monitor the status of reagents required for assay processing.	3/7/18
08	Performing Quality control	To confirm that systems, reagents and consumables are acceptable and working within specifications for each assay used	3/7/18
09	Result Review	To review the processed results in the system.	3/7/18
10	Result Intellicheck.	To check the Intellicheck function of the system.	3/7/18
11	Option & Configuration	To setup the system as per Laboratory requirement.	4/7/18

## Ortho Clinical Diagnostics

**Test: 1** : System Hardware Overview

**Purpose** : To make the operator to identify the instrument subsystem.

**Reference** : Operator Reference Guide - Pages 4-1 to 4-5

### Summary:

For better understanding purpose, Instrument has been divided into several parts according to its operation mode, so we call this partition as centers. And those centers are named as mention below.

### Procedure:

This will list the available system centers in the instrument and its subsystem contend to operator understanding. The operator has to overview the Service V-Docs to get an idea about the system centers overview.

- Sampling Centers
  - Sample Supply
  - Primary Tip sealer
  - Micro sensor subsystem
- Micro Immunoassay Center
  - Micro Immunoassay Metering & Reagent Metering
  - Micro Immunoassay Versa Tip Ring
  - Micro Well Incubator
  - Micro Well Wash Assembly
  - Signal Reagent Assembly
  - Luminometer
- Command Center
  - Master Computer & Monitor
  - Keyboard & Touch system
- System Frame and Cabinetry

**Test: 2** : **Starting Up and Shutting Down**  
**Purpose** : To make the instrument 'READY' for operation  
**Reference** : Operator Reference Guide - Pages 3-0 to 3-25

**Summary:**

Instrument will check status of different parts of the instrument automatically after booting up to system status screen; if there is an error code posted, initialize the system and follow corrective action instructions provided for the error code.

**Procedure:**

**Starting Up the System**

- Check the room temperature and switch on the Air Conditioner.
- Check the UPS.
- Switch on the Printer & Load the paper.
- Switch on the VITROS® 5600 Integrated System by lift the main switch up and hold it for about 5 sec.
- Wait for the instrument to get ready after initialization.
- The machine is ready for next step if it displays "READY" on the status console.
- If not, initialize by pressing the initialize button on the error code screen.
- Follow the instructions provided for the error codes.

**Shutting Down the system**

- Touch **Shutdown** in the main menu.
- Press 'Y/N' to continue the shutdown process.
- Desire normal shutdown or final shutdown and then select desired menu.
- If, you selected **Final Shutdown**, press the RESET button to restart the instrument, or press the Main power switch down to make it OFF.

Observation	System status console shows "Ready". Instrument is ready for operation	Remarks	Initial/Date
		Pass	2/7/18

**Test: 3** : **User Interface Overview**


**Purpose** : To make the operator to understand the system screens.

**Reference** : Operator Reference Guide - Pages 6-0 to 6-12

**Summary:**

The System status screen will make the operator to understand of different functionality of software utility available for the operator. This will helps operator to check system status as well as to instruct any commands to the system for required operation.

**Procedure:**



The above (as shown in the Previous Page) picture is actual system software screen and the number in the blue circle is to identify the several functionality of the software icons designed for the operator to interact with the system. We call these software icon as mentioned below.,

## Ortho Clinical Diagnostics

1. Status Line
2. Status Console
3. Function Screen
4. Prompt Line
5. Time, Date and Version Display
6. Status Indicators
7. Process Buttons

**Test: 4** : **Sample programming and Analysis**

**Purpose** : To program and process the samples

**Reference** : Operator Reference Guide (pages 9-1 to 9-13)

**Summary:** The operator can process assay by assigning program manually in the Universal Sample tray in 'Sample Program' menu or they can download sample program through 'LIS'. Sample programming is the process of selecting assays and programming characteristics for samples. The system uses the sample program to meter appropriate sample and select the right reagent for the assay, process and then report results with the correct identification.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Sample Programming methods & Overview	Desire the sample Programming method.	Pass	3/7/2018
02	Loading and Processing of samples	Tray cover opened and samples loaded in sample tray placed in sample supply.	Pass	3/7/2018
03	Programming samples	Sample program assigned for selected tray.	Pass	3/7/2018
04	Processing samples	Samples are processed automatically by the system.	Pass	4/7/2018
05	Unloading the samples	Tray cover opened and processed sample tray unloaded.	Pass	4/7/2018
06	Viewing samples in process	Sample under process are displayed on the 'View Sample Status' Screen.	Pass	4/7/2018



## Ortho Clinical Diagnostics

**Test: 5** : **Performing Calibration**

**Purpose** : To calibrate the system for every new lot of assay

**Reference** : Operator Reference Guide (pages 10-1 to 10-9)

**Summary:** The system requires its own calibration for every individual assay to measure the analytic concentration as well as to accept reagent pack status ready for processing. Assay calibration is a process that relates the response of the system to analyte concentration or activities. Calibration is performed periodically to adjust for changes in the system, assay protocols, or assay reagent lots.

The system requires calibration for individual assays when:

- A new assay is uploaded to the system
- The calibration expires (up to 28 days after it is processed, depending on the assay; refer to the package insert for expiration information)
- An assay reagent lot number changes
- Government regulations specified
- An assay's protocol changes

You also may need to perform calibration when:

- Certain service procedures are performed
- Quality control performance is out of range

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Load New ADD via CD or Downloaded file (New Gen Lot, Protocol , reagent lot calibration , Diluents lot information).	Calibrator identified and updated for protocol & master calibration data.	Pass	Vijay 3/7/2018
02	Preparing calibrators.	Calibrators are ready for processing.	Pass	3/7/2018
03	Performing Calibration with Bar code label.	System recognizes the barcode and processing the calibration automatically.	Pass	3/7/2018
04	Performing Calibration with Sample processing screen.	In the sample programming, the calibration program is assigned and processed the calibration assigned for each assay.	Pass	3/7/2018
05	Calibration report.	Calibration completed successfully. Report printed.	Pass	4/7/2018

## Ortho Clinical Diagnostics

**Test: 6** : **Maintenance & System Clean**

**Purpose** : Clean appropriate modules to maintain Accuracy and precision.

**Reference** : Operator Reference Guide 16-13 to 16-17

**Summary:**

Maintenance procedures are tasks that are performed to keep the system operating properly. Maintenance protocols to be performed according to the recommended schedule (daily, weekly, monthly, or as required). Ensure that we need to use 70% Isopropyl alcohol to disinfect the appropriate module to keep cleanliness and maintain the accuracy & precision.

The Maintenance is classified into four-category ie.,

1. Daily Maintenance
2. Weekly Maintenance
3. Monthly Maintenance
4. As required Maintenance

**Daily Maintenance: Pages from 16-14 to 16-15**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Perform Metering Maintenance	Metering Maintenance Performed	Pass	4/7/18
02	Empty Solid and Liquid waste container	Solid & Liquid waste containers are emptied.	Pass	4/7/18
03	Remove outdated or empty reagent packs, Signal Reagent packs and Universal Wash Buffer	Outdated empty Reagent packs, SR packs and UWR bottles are removed & discarded.	Pass	4/7/18
04	Inspect sample trays and adaptors	Sample Trays are cleaned.	Pass	4/7/18
05	Clean the SR Probe assembly	SR Probes are cleaned.	Pass	4/7/18
06	Load required reagent packs, Signal Reagent pack and Universal Wash Buffer.	All the required reagent packs are loaded and updated in the system as required.	Pass	4/7/18
07	Run Q.C fluids	Q.C samples are processed successfully.	Pass	4/7/18

## Ortho Clinical Diagnostics

Weekly Maintenance: Pages from 16-15 to 16-16

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Clean the Micro well Incubator.	Micro well incubator – Inner ring, outer ring, middle ring, shuttle weight, drop holes, Luminometer FOB, Wash reagent and signal reagent probes are cleaned.	Pass	5/7/18
02	Clean the Primary tip sealer.	Primary Tip Sealer cleaned.	Pass	5/7/18
03	Clean the Secondary Tip Sealer.	Secondary Tip Sealer cleaned.	Pass	5/7/18
04	Clean the Sample Supply and Cap Retainer.	Cap Retainer & Sample Supply cleaned.	Pass	5/7/18
05	Clean the Touch Screen Monitor & Keyboard.	Touch Screen Monitor & Key board Clean Done.	Pass	5/7/18
06	Run the maintenance Pack for Subsystem cleaning.	Maintenance pack is loaded; subsystem cleaning done automatically by the system.	Pass	5/7/18
07	Run QC Fluids	Q.C Processed successfully.	Pass	5/7/18

Ortho Clinical Diagnostics

Monthly Maintenance: Pages from 16-15 to 16-16

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Clean Micro sensor Cover & Ring Area.	Micro Sensor Cover & Ring Surface cleaned.	Pass	6/7/18
02	Inspect/Clean Micro Immuno Assay reagent Supply top Cover.	Micro Immuno Assay Reagent Supply top Cover Inspected and Cleaned.	Pass	6/7/18
03	Clean VITROS Versa Tip supply Registration Rail.	Versa tip Supply Registration Rail cleaned.	Pass	6/7/18
04	Inspect Reagent cooler filter for cleanliness.	Reagent cooler filter removed & cleaned.	Pass	6/7/18
05	Replace Vapors adsorption cartridge for every two months.	Every two months once, VAC replaced.	Pass	6/7/18
06	Make a backup of Q.C, Calibration and Configuration.	Backup of QC, Calibration and Configuration made successfully.	Pass	6/7/18
07	Inspect / Clean Master Computer Filter.	Inspected and Cleaned Master computer Air Filter.	Pass	6/7/18

## Ortho Clinical Diagnostics

**Test: 7** : **Managing reagents Inventory and Supply**

**Purpose** : To Maintain & monitor the status of reagents or supply required for assay processing.

**Reference** : Operator Reference Guide (pages 15-1 to 15-7)

**Summary:** The Reagent Management feature enables you to review current inventory information for the reagents loaded on the system. Using this function, you can load and unload reagents as necessary. To maintain the required reagents in the system for processing, the operator should review the Reagent management screen.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Review the reagent inventory to plan for the day.	The reagent inventory for the day planned.	Pass	3/7/18
02	Loading of Reagent Pack automatically	Required Reagent Packs loaded automatically by software request.	Pass	3/7/18
03	Loading of Reagent Pack with help of Manual Lot Entry button.	Requested Lot Information fed and the reagent pack loaded.	Pass	3/7/18
04	Loading of Signal Reagent automatically.	SR Pack loaded in position 1 & 2 and accepted by barcode reading.	Pass	3/7/18
05	Loading of Signal Reagent with Manual Load Button.	SR packs information fed and loading done.	Pass	3/7/18
06	Loading of Universal Wash Buffer	UWR buffer loading done through Load supply Software icons.	Pass	3/7/18
07	Unloading of Reagents	The entire Empty & expired reagents packs are unloaded by Load/Unload software icons.	Pass	3/7/18

## Ortho Clinical Diagnostics

### Test: 8 : Performing Quality control

**Purpose :** Quality Control (QC) is important in determining the performance and accuracy of the system. To perform Quality Control, QC materials are run with either known, or unknown values along with patient samples to determine whether the system is functioning within the established ranges for your lab.

**Reference :** Operator Reference Guide (Pages 9-6 to 9-8)

**Summary:** Performing quality control procedures is an important part of using or maintaining the system. This section explains:

- When you should perform quality control
- How to choose a control fluid

The recommended frequency for processing quality control fluids is once in every 24 hours. However, the frequency with which you perform quality control procedures may vary, depending on the requirements and regulations for processing control fluids by your national, state, provincial, and local governments. Quality control procedures within your own Laboratory may also require a different frequency. You should also perform quality control procedures when:

- Assays have been calibrated
- Certain service procedures are performed, other than routine maintenance

#### Procedure:

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Choosing the Control fluid	Required control fluid identified.	Pass	4/7/18
02	Preparing Liquid or Lyophilized control fluids	Control fluid prepared and ready for processing.	Pass	4/7/18
03	Creating QC file.	Q.C file created for assay in the system according to control fluid.	Pass.	4/7/18
04	Process QC samples	QC samples are programmed in the sample programming window and the QC samples are loaded and processed automatically	Pass	4/7/18
05	Review Q.C result.	Processed Q.C results are reviewed and found satisfactory.	Pass	4/7/18
06	Display & printing graph.	Q.C graph reviewed and printed.	Pass	4/7/18
07	Managing Quality control Reports	Required reports printed and filed.	Pass	4/7/18

## Ortho Clinical Diagnostics

**Test: 9** : **Result Review.**

**Purpose** : To review the processed results in the system.

**Reference** : Operator Reference Guide (Pages 11-1 to 11-6)

**Summary:** The Results Review function helps to evaluate result records for patient and quality control samples. The results will be displayed along with the Reagents Lot information, Dilution information & if there is any error codes or Flags.

Result records contain the data generated by the system when assays are processed. The system can store up to 25,000 result records. When this limit is reached, new result records overwrite the oldest records. The system permanently deletes the overwritten records from computer memory.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Update List.	Sample under process status displayed with all information.	Pass	3/7/18
02	Monitoring Results.	Completed Recent Assay Results displayed on the screen.	Pass	3/7/18
03	Filter Results.	Processed Assay Results displayed as per the selected criteria.	Pass	3/7/18
04	Edit Patient Data	User can Edit/Add Patient Demography information, but the Patient ID will remain same.	Pass	3/7/18
05	Retrieving and Reviewing Archive Results by Set Report Status.	Archived Results are updated successfully in the CD/Pen Drive. The same Retrieved from the CD/Pen Drive.	Pass	3/7/18
06	Managing Reports by Set Report Status.	Required Reports got printed and filed.	Pass	3/7/18
07	Integrated Codes and Flags	Reported Codes and Flags are referred in Flags and Code chart. Necessary corrective action taken.	Pass	3/7/18

## Ortho Clinical Diagnostics

**Test: 10** : **Result Intellicheck.**

**Purpose** : To check the Integrity of Performed assays.

**Reference** : Operator Reference Guide (Pages 11-8 to 11-9).

**Summary:** The Result Intellicheck screen to view Intellicheck Technology Verifications performed for each sample and assay processed. For the selected result record, you can view verification data and detected exceptions for each analyte.

Example:

- To Analyze Intelli Report, select the sample ID listed on result review screen.
- On Review results screen, touch the 'Result Intellicheck' Icon.
- Result Intellicheck report comes on the screen.
- Check the parameters of Sample Metering, Sample + Reagent volume, Signal reagent volume and well wash verification for their acceptance.
- Take print of the Result intellicheck report.

**Acceptance criteria:**

Sr. No.	Parameter	Acceptance limit	Remarks	Done By
			Pass / Fail	Date
01	<b>Sample Metering</b> <ul style="list-style-type: none"> <li>• Clot</li> <li>• Bubble</li> <li>• Short sample</li> <li>• Viscosity</li> <li>• Thin layer of fluid</li> </ul>	Against all the parameters, "No" should be displayed on screen	Pass	4/7/2018
02	<b>Reagent Metering</b> Sample + Reagent volume	No Exception. Range: 12700 – 19000	Pass	4/7/2018
03	Signal Reagent	No Exception Range: 17500 – 22800	Pass	4/7/2018
04	Well wash verification	No Exception Range: 21300 – 25000	Pass	4/7/2018
05	Luminometer – Self calibration	No Exception	Pass	4/7/2018



## Ortho Clinical Diagnostics

**Test: 9** : Option & Configuration

**Purpose** : To setup the system as per laboratory requirement

**Reference** : Operator Guide (Pages 11-8 to 11-14)

**Summary:** The Options & Configuration function provides many features for customizing your VITROS® 5600 Integrated System. It is organized into three main groups:

- Configure Analyte Data & Review/Edit Calibration Data
- System Setup
- System Services

Selections within these groups allow you to customize analyte parameters, review calibration data, perform user calibrations, configure and set system parameters, review usage inventory, configure e-Connectivity® and network device parameters, configure printer, laboratory computer and auxiliary ports, perform backup for quality control, calibration and configuration files and perform an archive procedure for result records.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Configure Assays	Analyte parameters are configured as desired by laboratory.	Pass	6/7/18
02	Review / Edit Calibration Data	Calibration data updated.	Pass	6/7/18
03	Configure System Setup	Required subsystem configuration is done successfully.	Pass	6/7/18
04	Configure Subsystem Setup	The required subsystem can be disabled / deactivated as per needs.	Pass	6/7/18
05	Configure Report Control	System report parameter has set for printer & LIS.	Pass	6/7/18
06	Configure Communication	System interface protocol set for Laboratory Information System (LIS), Ethernet and e-Connectivity® communications,	Pass	6/7/18
07	Configure Demography	Global demographic attributes to be used when configuring age, sex, and normal ranges for specific assay/body fluids is defined.	Pass	6/7/18
08	System Services 1.Datalogger 2.Perform Backup 3.Usage Counters 4.Option Summary 5.Load System Data	Shall be performed and reviewed as & when required.	Pass	6/7/18

**V. Operational procedure:**

a. Certificate of Training

1. Technician Training

This certifies that the technicians have received basic user training in the following categories for the system described in this Operational Qualification.

Mr.Vijaykumar Rathod , Territory Manager from Ortho Clinical Diagnostics has conducted the training.

Sr. No.	Training program	Initials	Date
1.	Instrument Setup		
2.	System Operation		
3.	Basic trouble shooting and Maintenance		

2. Operator Training

The users responsible for the operation of this equipment have been trained in the proper usage of the system. Training focused on the basic operation and maintenance of the system.

Sr. No.	Operators	Department	Initials	Date
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				

Ortho Clinical Diagnostics

b. Customer SOP / Manuals:

Title	Number	Version	Verified by	Date
Vitros 5600 Integrated System Reference Guide			Vijay	2/7/2018
Microslide Instruction for use Manual			Vijay	2/7/2018
Microtip Chemistry Instruction For Use Manual			Vijay	2/7/2018
Micro Well Instruction For Use Manual			Vijay	2/7/2018

**VI. COMMENTS :**

Deviation:

Impact On Operation:


Corrective Action:

**IX. SYSTEM CERTIFICATION**

Study data has determined that the system described in this document either meets all criteria outlined in this Installation Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been / not been addressed. The system is ready for Operation Qualification.


Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Rathod Vijaykumar  
Designation : Territory Manager  
Company : Ortho Clinical Diagnostics


Signature:   
Date: 07/07/2018

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hotkoo  
Designation : ASSO. PROF.  
Department : Biochemistry

Signature:   
Date: 07/07/2018

Name : Mr. Bhawar S. R.  
Designation : SE. Lab Technician  
Department : Biochemistry

Signature:   
Date: 07/07/2018

Customer Authorizations:

Name : Dr. S. N. Jangle Sir.  
Designation : Prof. & Head Biochemistry  
Signature : S. N. Jangle  
Date : 07/07/2018

**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LOMI-413736, (M.S.) INDIA**

0703


Ortho Clinical Diagnostics

## INSTALLATION QUALIFICATION

For

# VITROS® 5600 INTEGRATED SYSTEM

System  
**VITROS® 5600**  
Integrated



Manufactured by:

Ortho Clinical Diagnostics, Inc., US

## Table of Contents

<b>Sr. No.</b>	<b>Contents</b>	<b>Page No.</b>
I	Approval of the IQ procedure	3
II	Instructions	4
III	Scope	5
IV	Ancillary Information	6
V	Installation Qualification	8
VI	Installation Procedure	10
VII	Installation Report	15
VIII	Comments	16
IX	System Certification	17
	Appendix	
	I. Installation Certificate	


Ortho Clinical Diagnostics

**I. APPROVAL OF THE IQ PROCEDURE:**

Both Central Clinical Laboratory Pravara Rural Hospital and Ortho Clinical Diagnostics are responsible for installation of VITROS® 5600 Immunodiagnostic System bearing Sr. No J5600-0703 in Central Clinical Laboratory, Pravara Rural Hospital as per the attached protocol.

Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Mitesh Shah

Signature: 


Designation : Sr. Zonal Manager

Date: 23/06/2018

Company : Ortho Clinical Diagnostics

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hotkar


Signature: 

Designation : Asst. Professor

Date: 23/06/2018

Department : Biochemistry

Name : Mr. Bhawar S. R.

Signature: 

Designation : Sr. Lab. Tech.

Date: 23/06/2018

Department : Biochemistry

Customer Authorizations:

Name : Dr. S. N. Jangle, Sr

Designation : Prof. & Head, Dept. of Biochemistry

Signature : S. N. Jangle

Date : 23/06/2018

**PROFESSOR & HEAD**  
**DEPARTMENT OF BIOCHEMISTRY**  
**RURAL MEDICAL COLLEGE**  
**LONI-413736, (M.S.) INDIA**



**II. INSTRUCTIONS:**

1. This document is to be completed at the time the system is installed to its location and set up for operation.
2. An authorized Ortho Clinical Diagnostics representative will check the system and enter the specific data as outlined in the appropriate Installation Qualification.
3. ~~Employees of Central Clinical Laboratory Pravara Rural Hospital will verify each result and sign in the each page. The member of the validation team will carry out this procedure.~~
4. ALL deviations from normal specification during installation will be noted under COMMENTS. All resolution to such problems will also be noted in the COMMENTS section. Additional space is provided at the end of this protocol for the same.

**III. SCOPE**

This Installation Qualification protocol will be performed on the VITROS® 5600 Immunodiagnostic System, and the Sr. No. ~~56000703~~ located at

This protocol will define the documentation that will be used to evaluate the instrument's installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage.

Installation checks will also be performed to verify that the Instrument has been installed with proper connections and utilities.

Trained, knowledgeable personnel will perform Installation Qualification studies.

Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

**IV. ANCILLARY INFORMATION**

a. Certification of Purchase Order Compliance

I certify to the best of my knowledge, the instrument VITROS® 5600 Immunodiagnostic System and Sr. No. ~~56000703~~ installed on 20/06/2018.

Verified By: Mitesh C. Shah Date: 20/06/2018

b. Utilities

Sr. No	Utility	Verified by & date
1.	<b>Environmental condition:</b> As per requirement (To be free from Dust, Electrical & magnetic Interferences and free from vibration)	✓ Yes / No Mitesh C. Shah 20/06/2018
2.	<b>Adequate space for installation:</b> (Length 110 inches x Width 35 inches x Height 84 inches)	✓ Yes / No Mitesh C. Shah 20/06/2018
3.	<b>Electrical Outlets:</b> Actual Voltage on site [200 Vac – 240 Vac] <b>Electrical Input:</b> Voltage supplied through ON LINE UPS (232Vac @ 50Hz frequency, Earthing < 2.0Vac)	✓ Yes / No Mitesh C. Shah 20/06/2018
4.	<b>Capacities:</b> <ul style="list-style-type: none"> <li>• 90 samples (80 Routine positions &amp; 10 STAT positions are available)</li> <li>• 150 Reagent Positions are available.</li> </ul>	✓ Yes / No Mitesh C. Shah 20/06/2018
5.	<b>Temperature:</b> 15° C to 30° C 15% to 75% relative humidity	✓ Yes / No Mitesh C. Shah 20/06/2018

## Ortho Clinical Diagnostics

The instrument has been verified for the following:

Sr. No.	Verification	Yes / No	Verified by & date
1.	Equipment is identified	✓ Yes / No	20/06/2018
2.	Manufacturer's specifications are included	✓ Yes / No	20/06/2018
3.	Accessories / Consumables are listed	✓ Yes / No	20/06/2018
4.	Equipment manual from the manufacturer is documented	✓ Yes / No	20/06/2018
5.	Manufacturer's Certificate of compliance attached	✓ Yes / No	20/06/2018

## V INSTALLATION QUALIFICATION:

## A. Equipment Description

The VITROS® 5600 Integrated System is a Random access, walk away system intended for use in the *in vitro* quantitative, semi-quantitative, and qualitative measurement of a variety of analytes of clinical interest, using VITROS Chemistry Products MicroSlides, VITROS Chemistry Products MicroTip reagents and VITROS Immunodiagnostic Products Reagents.

Instrument Identification		Verified by	Date
Equipment Name :	<b>Automated Integrated System</b>	Mitesh C. Shah	20/06/2018
Model :	VITROS® 5600	Mitesh C. Shah	20/06/2018
Manufacturer :	Ortho Clinical Diagnostics, Inc., US	Mitesh C. Shah	20/06/2018
Marketed by :	Ortho Clinical Diagnostics India Pvt. Ltd.	Mitesh C. Shah	20/06/2018
Serial Number :	5600 0703	Mitesh C. Shah	20/06/2018
Lab Id :	No. 1	Mitesh C. Shah	20/06/2018
Software Name :	QNX	Mitesh C. Shah	20/06/2018
Software Version :	V. 3.3.1	Mitesh C. Shah	20/06/2018
Size (in inches) :	Adequate for installation: (Length 170 x Width 83 x Height 84).	Mitesh C. Shah	20/06/2018
Power :	1600W@ 50Hz of 220Vac - 240Vac	Mitesh C. Shah	20/06/2018

## Ortho Clinical Diagnostics

### B. Accessories/Consumables

The following accessories were supplied with the instrument. Check (✓) 'verified by' in case they are found to be in order.

Description	Quantity	Verified	Date
User Training Manual	1	✓	20/06/2018
Application Software-Revelation	1	✓	20/06/2018
Universal Sample Tray	9	✓	20/06/2018
Backup DVD R/w	3	✓	20/06/2018
Printer Cable	1	✓	20/06/2018
Printer Software	1	✓	20/06/2018
Power Cords	3	✓	20/06/2018
Printer	1	✓	20/06/2018
Air filter	1	✓	20/06/2018
Waste can 5L	1	✓	20/06/2018

### C. List of Manuals :

Ortho Clinical Diagnostics has supplied following manual.

**D. Maintenance :**

The instrument listed within this document will be placed under the control of the purchasing institution with respect to proper maintenance procedures as detailed in the Training and Reference Guide. The maintenance procedures will be filed separately.

A trained analyst using the manuals provided with the instrumentation can perform simple maintenance. Upon expiration of the warranty period Johnson and Johnson Ltd offers several levels of Maintenance Agreements and Performance Testing services to assist you in maintaining GLP/GMP compliance. Contacting your local representative and requesting the additional Service Agreement can supply additional information.

## VI. INSTALLATION PROCEDURE

(The following steps Performed at the time of original installation at the initial location)

1. Locating & unpacking the instrument.
2. Reaffix/verify the circuits boards & CPU
3. Nominal Line voltage frequency selection of transformer.
4. Load Supply & Power ON the system.
5. System Configuration.
6. Systems Tests & Adjustments.
7. Subsystems Performance Verification & calibration
8. Setting and installing printer.

The Above mentioned steps has completed successfully by trained field Engineer as described below.

### VI.1 Locating & unpacking the VITROS® 5600 instrument:

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- Check the Tip & Tell Label.
- Verify the serial no / J number of the system match those indented for delivery.
- Place the pallet in a position with a minimum of 7 m (24 ft.) of clearance in front of the end with the LABELS
- Remove the STRAPPING and the RAMPS from the pallet..
- Assemble the 2 piece RAMP and hok the RAMPS to the end of the pallet
- Move the system down the ramps. Do the same for the second half of the system.
- Place the instrument in the lab leveled floor.
- Join the two halves of the system as per manufacturer's instructions.
- Remove the packing material from
  - VERSATIP supply carousel
  - Under side of sample supply cover



- SR metering Nozzle
- Well Wash Nozzles
- Beneath SR Pumps
- Luminometer & Micro ImmunoAssay VERSATIP ring
- Supply 4 load doors & Reagent Well shuttle
- SR carousel
- Remove the wire Tie, tape & Foam from UIA REAGENT METERING ARM.
- Remove the moisture separators behind the compressor installed on it bracket.
- Remove the Foam supporting from the compressor.

**VI.2 Reaffix/verify the circuits boards & CPU:**

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- Open the right side front door and open the card rack metal cover.
- Remove the RC labeled DSP boards and reaffix it back.
- Open the right side rear panel and open the card rack metal cover.
- Remove the UC labeled DSP boards and reaffix it back.
- Open the Middle front door and open the CPU top cover & verify the boards.

**VI.3 Nominal Line voltage / frequency selection of transformer:**

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- Connect the Primary T2\_H6 on Label no H6 for 0Vac.
- Connect the Primary T2\_H# on label no H2 for 230Vac.
- Connect the Secondary T2\_X1 on label no X1 for 230Vac.
- Connect the Secondary T2\_X# on label no X2 for 200Vac.
- Connect the Secondary T2\_X3 & T2\_X3 on label no X3 for 0Vac.

**VI.4 Load Supply & Power ON the system:**

(\ServicePublications\5600\5600Service\J32845\_Unpack-Install\_Intact\J32845.pdf)

- VersaTip & Sample Trays.
- Signal Reagent & Universal Reagent.
- Check the supply and Earthing voltage.
- Switch **ON** the Instrument.

**VI.5 System Configuration.:** (\6902906\_3600-RefGd\_Ltr-EN.pdf)

Click on Set Access Level from Status Menu and type password

Option -> Configure System ->

- **Configure Current Date & Time:**

Select the format and set the **Date and Time.**

- **Configure System Name & J Number**

Enter the System Name & J number.

- **Configure the Screen Saver**

Set the Screen saver delay time.

- **Configure the Site Temperature**

Set the site temperature tolerance for the nominal site temperature.

Status -> Diagnostics -> Select the required task

- **Touch Screen Calibration:**

Touch "Calibrate Touch screen" at the bottom of the DIAGNOSTICS Menu.

Touch center of the target appear on the screen.

When you have finished, touch "Save Calibration"

- **Country Code Selection:**

Touch "Diagnostics" then Select "V-Docs"

Press [Alt] and [S] to Access the service Scripts.

Select "Configure Country Code"

Select the appropriate country from the List.

Select "Set country code"

Touch "Return".

Touch "Shut Down". Configure the Language from the System Menu button.

Touch "Final Shutdown". Then reset the system.

**VI.6 System Tests and Adjustments:** (\6902906\_3600-RefGd\_Ltr-EN.pdf)

Adjustments are diagnostic functions used to fine-tune or define various system Parameters to ensure proper system performance. With the exception of the IRS Calibration, all other adjustments are available only to trained service personnel.

- MicroImmunoassay ( $\mu$ IA) Metering
- MicroSensor
- MicroWell Reagent Metering
- MicroWell Wash Metering
- Signal Reagent Metering
- Luminometer
- Scrap Run

**VI.7 Subsystems Performance Verification & Calibration :** (\6902906\_3600-RefGd\_Ltr-EN.pdfpdf)

- Well Wash Dispense & Aspiration Calibration.
- Signal reagent Dispense calibration.
- 30PSI & 10PSI calibration.
- Soak Volume Verification
- IRS Calibration

**VI.8 Setting and installing printer:** (\6902906\_3600-RefGd\_Ltr-EN.pdf)

- Remove the Packing material form the printer and assemble the accessories.
- Connect the USB cable and Switch ON the Printer.
- Set report control and print the test page.

## VII. Installation Report :

Activity	Observation	Remarks	Verified By / Date
Locating & unpacking the instrument.	Instrument was located and unpacked	Ok	Mitesh C. Shah 21/06/2018
Reaffix/verify the circuits boards & CPU	Reaffixed/verified the circuits boards & CPU	Ok	Mitesh C. Shah 21/06/2018
Nominal Line voltage frequency selection of transformer.	Nominal Line voltage frequency was sated.	Ok	Mitesh C. Shah 21/06/2018
Load Supply & Power ON the system.	Supply Loaded & Powered ON the system.	Ok	Mitesh C. Shah 21/06/2018
System Configuration.	System was configured as per the requirement.	Ok	Mitesh C. Shah 21/06/2018
System Tests and Adjustments	System Tested and Adjustments done.	Ok	Mitesh C. Shah 21/06/2018
Subsystems Performance Verification & calibration	Subsystems Performance Verified & calibrated successfully.	Ok	Mitesh C. Shah 21/06/2018
Setting and installing printer	Printer was installed and connected to the system.	Ok	Mitesh C. Shah 21/06/2018

**VIII. COMMENTS :**

Deviation:

NA

Impact On Operation:

NA

Corrective Action:


NA

**IX. SYSTEM CERTIFICATION**

Study data has determined that the system described in this document either meets all criteria outlined in this Installation Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been / not been addressed. The system is ready for Operation Qualification.

Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Mitesh Shah  
Designation : Sr. Zonal Manager  
Company : Ortho Clinical Diagnostics

Signature:   
Date: 23/06/2018

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hotkar


Signature: 

Designation : ASSO. PROFESSOR

Date: 23/06/2018

Department : Biochemistry

Name : Mr. Bhawar S. P.

Signature: 

Designation : Sr. Lab. Technician.

Date: 23/06/2018

Department : Biochemistry

Customer Authorizations:

Name : Dr. S. N. Jangle Sr

Designation : Prof. & Head Biochemistry

Signature : S. N. Jangle

Date : **PROFESSOR & HEAD**  
**DEPARTMENT OF BIOCHEMISTRY**  
**RURAL MEDICAL COLLEGE**  
**LONI-413736, (M.S.) INDIA**


Ortho Clinical Diagnostics

**PERFORMANCE QUALIFICATION**

**For**

**VITROS® 5600 INTEGRATED SYSTEM**

System  
**VITROS® 5600**  
Integrated



Manufactured by:

Ortho Clinical Diagnostics, Inc., US

Table of Contents

<b>Sr. No.</b>	<b>Contents</b>	<b>Page No.</b>
I	Approval of the PQ procedure	3
II	Instructions	4
III	Scope	5
IV	Performance Qualification	6
IV.I	Performance Procedure	7
V	Performance Qualification Report	10
VI	Comments	11
VII	System Certification	12



**I. APPROVAL OF THE PQ PROCEDURE:**

Both Central Clinical Laboratory Pravara Rural Hospital and Ortho Clinical Diagnostics are responsible for installation of VITROS® 5600 Immunodiagnostic System bearing Sr. No J5600-0703 in Central Clinical Laboratory, Pravara Rural Hospital as per the attached protocol.

Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Rathod Vijaykumar

Signature: 


Designation : Territory Manager

Date: 07/07/2018

Company : Ortho Clinical Diagnostics

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hottkar


Signature: 

Designation : ASSO. PROF.

Date: 07/07/2018

Department : Biochemistry

Name : Mr. Shewar S. R.

Signature: 

Designation : SE. Lab Tech.

Date: 07/07/2018

Department : Biochemistry

Customer Authorizations:

Name : Dr. S. N. Jangle, Sr.

Designation : Prof & Head, Biochemistry

Signature : S. N. Jangle

Date : 07/07/2018

PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LON-413736, (M.S.) INDIA

**II. INSTRUCTIONS:**

1. This document is to be completed at the time the system is installed to its location and set up for operation.
2. An authorized Ortho Clinical Diagnostics representative will check the system and enter the specific data as outlined in the appropriate Installation Qualification.
3. ~~Employees of Central Clinical Laboratory Pravara Rural Hospital~~ will verify each result and sign in the each page. The member of the validation team will carry out this procedure.
4. ALL deviations from normal specification during installation will be noted under COMMENTS. All resolution to such problems will also be noted in the COMMENTS section. Additional space is provided at the end of this protocol for the same.

**III. SCOPE**

This Performance Qualification protocol will be performed on the VITROS® 5600 Integrated System, and the Sr. No. ~~J 56000705~~ located at

This protocol will define the documentation that will be used to evaluate the instrument's performance check in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been performed in accordance with the intended usage.

Performance checks will also be performed to verify that the Instrument has been operated with proper information/sequence and utilities.

Trained, knowledgeable personnel will perform Performance Qualification studies.

Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

**IV. PERFORMANCE QUALIFICATION**

**A. Instrument Identification**

1. Model Name                      VITROS 5600  
 2. Serial Number                  J56000703

**Verified Date**

2/7/2018  
2/7/2018

**B. Following is a list of tests to be performed and verified:**

Sr.No	Test Name	Test Purpose	Initial / Date
01	QC Run	To see the performance of quality control material on the equipment on selected assay parameters as per the specifications given	2/7/2018
02	Accuracy Study	To compare the obtained value with true values of processed control.	22/6/2018
03	Precision Study	To check the precision performance of the equipment	22/7/2018

**C. Performance Testing:**

**Test I**

**Test Name** : **QC Run**  
**Purpose** : To see the performance of quality control material on the equipment as per the specifications given  
**Method** : Microslide – Rate Chemistry & Endpoint Chemistry  
 Microslide – Potentiometric Chemistry;  
 Microslide – Immunorate Chemistry;  
 Microtip Chemistry  
 Microsensor Chemistry  
 Microwell - Chemiluminescence Immunoassay

**Analysis of controls:**

**Note:** Analyze controls for ALT (Microslide Rate Chemistry);  
 Sodium (Potentiometric Chemistry);  
 BuBc (Microslide End point Chemistry)  
 Phenytoin (Microslide – Immunorate Chemistry)  
 dLDL (Microtip Chemistry)  
 Gentamycin (Microtip Chemistry)  
 IgM (Microtip Chemistry)  
 HIT (Microsensor Chemistry)  
 TSH (Microwell - Immunometric assay) &  
 TT4 (Microwell - Competitive assay).

Sr. No.	Activity	Procedure done as per the protocol defined in Vitros 5600 V Docs – System Operation – Quality Control	Remarks	Done By
			Pass/Fail	Date
01	Preparing Liquid or Lyophilized control fluids	“Instructions for use” of QC material	Pass	Vjpy 3/7/18
02	Creating QC file	V Docs – System Operation – Quality Control – Define control fluids	Pass	3/7/18
03	QC sample programming and analysis	V Docs – System Operation – Quality Control – Process Control fluid samples & Review the Control sample results.	Pass	3/7/18

**Ortho Clinical Diagnostics**

**Test II**

**Test Name** : Accuracy

**Purpose** : To see the accuracy of obtained quality control value in comparison with the expected mean values.

**Method** : Microslide; Microtip; Microsensor and Microwell method as mentioned above

**Analysis of controls:**

**Note:** Analyze controls as mentioned above.

Sr. No.	Activity	Procedure done as per the protocol defined in Vitros 5600 V Docs – System Operation – Quality Control	Remarks	Done By
			Pass/Fail	Date
01	Preparing Liquid or Lyophilized control fluids	'Instructions for use' of QC material	Pass	3/7/2018
02	QC sample programming and analysis	V Docs – System Operation – Quality Control – Process Control fluid samples & Review the Control sample results.	Pass	3/7/2018
03	Accuracy Analysis	Compare the obtained Q.C value with mean of expected value as mentioned in the QC Value chart.	Pass	3/7/2018

## Ortho Clinical Diagnostics

### Test III

**Test Name** : Precision Study  
**Purpose** : To see the precision performance of the equipment  
**Method** : Microslide – Rate Chemistry & Endpoint Chemistry  
 Microslide – Potentiometric Chemistry;  
 Microslide – Immunorate Chemistry;  
 Microtip Chemistry  
 Microsensor Chemistry  
 Microwell - Chemiluminescence Immunoassay

- Analyze Vitros Performance Verifier Level 1 control for the following tests : ALT (5 x 7 times), Na<sup>+</sup> (5 x 8 times), dLDL (6 x 6 times), BuBc (5 x 7 times)
- Analyze TDM Performance Verifier Level 3 for Phenytoin (5 x 7 times).
- Analyze TDM Performance Verifier Level 1 for Gentamycin (6 x 6 times).
- Analyze Protein Performance Verifier Level 1 for IgM (6 x 6 times).
- Analyze Microsensor Check Fluid Level I for Hemolysis, Icterus and Turbidity (20 times each).
- Analyze all the three levels of Vitros Total Thyroid controls for Microwell – Chemiluminescence Immunoassay TT4 and TSH (10 times each).
- Calculate the Mean, SD and CV%.

#### Acceptance Criteria :

Sr. No.	Analyte	Control Level	Precision Limit
01	ALT	PV I	≤ 2.2 SD
02	Sodium	PV I	≤ 0.69% CV
03	dLDL	PV I	SD < 2.4
04	Bu	PV I	≤ 0.3 SD
05	Bc	PV I	≤ 0.6 SD
06	Phenytoin	TDM PV III	≤ 0.75 SD
07	Gentamycin	TDM PV I	SD < 0.11

## Ortho Clinical Diagnostics

08	IgM	Protein PV I	SD < 2.1
09	Hemolysis	MS Check Fluid Level I	SD < 4.4
10	Icterus	MS Check Fluid Level I	SD < 0.4
11	Turbidity	MS Check Fluid Level I	SD < 14.8
12	TT4	Total Thyroid Control Level 1	CV% ≤ 4.9
		Level 2	CV% ≤ 4.6
		Level 3	CV% ≤ 6.0
13	TSH	Total Thyroid Control Level 1	CV% ≤ 10.9
		Level 2	CV% ≤ 4.9
		Level 3	CV% ≤ 4.7

\* Reference: The acceptance SD & CV% has been taken from Manufacturer's recommended limits.



**Table for Record the Obtained Q.C Values :**

**Test Name:**

**Control Level:**

**Control Lot:**

**Expiry Date:**

Sr. No	Obtained Q.C Value	Range of the Comparator	Mean of the Comparator	Remarks	Done By
				Pass / Fail	Date
01					
02					
03					
04					
05					
06					
07					
08					
09					
10					

\* Reference :- Attached QC Report.

**V. COMMENTS:**

**Deviation:**

**Impact On Operation:**


**Corrective Action:**

IX. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outlined in this Installation Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been / not been addressed. The system is ready for Operation Qualification.


Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Rathod Vijaykumar  
Designation : Territory Manager  
Company : Ortho Clinical Diagnostics

Signature:   
Date: 07/07/2018

Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hottker


Signature: 

Designation : ASSO. PROF.

Date: 07/07/2018

Department : Biochemistry

Name : Mr. Bhawar S. R.

Signature: 

Designation : Lab. Technician

Date: 07/07/2018

Department : Biochemistry

Customer Authorizations:

Name : Dr. S. N. Jangle Sir

Designation : PROF & Head Biochemistry

Signature : S. N. Jangle

Date : 07/07/2018

PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413736, (M.S.) INDIA

# Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,  
Andheri East, Mumbai – 400059  
T : +91 22 6787 9300  
F : +91 22 6787 9333

## Calibration Certificate

The below mentioned instrument has been calibrated and tests performed to check the system performance.

Instrument : VITROS V5600  
Serial No : 5600-0703  
Customer Name : PRAVARA MEDICAL TRUST-LONI  
Calibration performed on : 06.12.2018

The system's calibration includes optics calibration and checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.

Next Calibration will be performed in JUNE 2019.

For Ortho Clinical Diagnostics India Pvt Ltd.

  
Mitesh Shah

Date: 06/12/2018

Sr.Zonal Service Manager-Ortho Care

Pune


Ortho Clinical Diagnostics

**PERFORMANCE QUALIFICATION**

For

**VITROS® 5600 INTEGRATED SYSTEM**

**VITROS 5600**  
Integrated



Manufactured by:

Ortho Clinical Diagnostics, Inc., US

## Table of Contents

<b>Sr. No.</b>	<b>Contents</b>	<b>Page No.</b>
I	Approval of the PQ procedure	3
II	Instructions	4
III	Scope	5
IV	Performance Qualification	6
IV.I	Performance Procedure	7
V	Performance Qualification Report	10
VI	Comments	11
VII	System Certification	12


## Ortho Clinical Diagnostics

### I. APPROVAL OF THE PQ PROCEDURE:

Both Central Clinical Laboratory, Pravara Rural Hospital and Ortho Clinical Diagnostics are responsible for Performance check of VITROS® 5600 Integrated System bearing Sr. No J56001041 in Central Clinical Laboratory, Pravara Rural Hospital, Loni as per the attached protocol.

Protocol Performed By: Ortho Clinical Diagnostics representative


Name : Kailas Ghait  
Designation : Sr. Territory Manager  
Company : Ortho Clinical Diagnostics

Signature: 

Date: 10/12/2020


Validation Team from Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hotkar  
Designation : ASSO. prof.  
Department : Biochemistry

Signature: 

Date: 10/12/2020

Name : Mr. Bhawar S. R.  
Designation : Lab Tech.  
Department : Biochemistry

Signature: 

Date: 10/12/2020

Customer Authorizations:

Name : Dr. S. N. Jangle sir  
Designation: prof & Head Biochemistry  
Signature : S. N. Jangle  
Date : 10/12/2020

**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413736, (M.S.) INDIA**

**II. INSTRUCTIONS:**

1. An authorized Ortho Clinical Diagnostics representative will check for the performance of the instrument and enter the specific data as outlined in the Performance Qualification.
2. The concerned lab personnel are responsible for performance checks described in the Performance testing.
3. The concerned employees of Central Clinical Laboratory Pravara Rural Hospital will verify each result and sign in each page. The member of the validation team will carry this out.
4. ALL deviations from the acceptance criteria detailed in this document will be noted in the **COMMENTS** section at the end of each PQ protocol. All resolution to such problems will also be noted in the **COMMENTS** section and must be resolved prior to issuance of a **SYSTEM CERTIFICATION**.



### III. SCOPE

This Performance Qualification protocol will be performed on the VITROS® 5600 Integrated System, and the Sr. No. J56001041 located at Central Clinical Laboratory Pravara Rural Hospital, Loni.

This protocol will define the documentation that will be used to evaluate the instrument's performance check in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been performed in accordance with the intended usage.

~~Performance checks will also be performed to verify that the Instrument has been operated with proper information/sequence and utilities.~~

Trained, knowledgeable personnel will perform Performance Qualification studies.

Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated, and the appropriate course of action determined. All documents will be initialed and dated.

## IV. PERFORMANCE QUALIFICATION

## A. Instrument Identification

		Verified Date
1. Model Name	VITROS 5600	05/12/2020
2. Serial Number	J56001041	05/12/2020

## B. Following is a list of tests to be performed and verified:

Sr.No.	Test Name	Test Purpose	Initial / Date
01	QC Run	To see the performance of quality control material on the equipment on selected assay parameters as per the specifications given	Kailas Ghait 05/12/2020
02	Accuracy Study	To compare the obtained value with true values of processed control.	Kailas Ghait 05/12/2020
03	Precision Study	To check the precision performance of the equipment	Kailas Ghait 05/12/2020

**C. Performance Testing:**

**Test I**

**Test Name** : **QC Run**

**Purpose** : To see the performance of quality control material on the equipment as per the specifications given

**Method** : Microslide – Rate Chemistry & Endpoint Chemistry  
 Microslide – Potentiometric Chemistry;  
 Microslide – Immunorate Chemistry;  
 Microtip Chemistry  
 Microsensor Chemistry  
 Microwell - Chemiluminescence Immunoassay

**Analysis of controls:**

**Note:** Analyze controls for ALT (Microslide Rate Chemistry);  
 Sodium (Potentiometric Chemistry);  
 BuBc (Microslide End point Chemistry)  
 Phenytoin (Microslide – Immunorate Chemistry)  
 dLDL (Microtip Chemistry)  
 Gentamycin (Microtip Chemistry)  
 IgM (Microtip Chemistry)  
 HIT (Microsensor Chemistry)  
 TSH (Microwell - Immunometric assay) &  
 TT4 (Microwell - Competitive assay).

Sr. No.	Activity	Procedure done as per the protocol defined in Vitros 5600 V Docs – System Operation – Quality Control	Remarks	Done By
			Pass/Fail	Date
01	Preparing Liquid or Lyophilized control fluids	“Instructions for use” of QC material	Pass	Kailas Ghait 05/12/2020
02	Creating QC file	V Docs – System Operation – Quality Control – Define control fluids	Pass	Kailas Ghait 05/12/2020
03	QC sample programming and analysis	V Docs – System Operation – Quality Control – Process Control fluid samples & Review the Control sample results.	Pass	Kailas Ghait 05/12/2020

## Ortho Clinical Diagnostics

**Test II**

**Test Name** : Accuracy  
**Purpose** : To see the accuracy of obtained quality control value in comparison with the expected mean values.  
**Method** : Microslide; Microtip; Microsensor and Microwell method as mentioned above

**Analysis of controls:**

**Note:** Analyze controls as mentioned above.

Sr. No.	Activity	Procedure done as per the protocol defined in Vitros 5600 V Docs – System Operation – Quality Control	Remarks	Done By
			Pass/Fail	Date
01	Preparing Liquid or Lyophilized control fluids	'Instructions for use' of QC material	Pass	Kailas Ghait 05/12/2020
02	QC sample programming and analysis	V Docs – System Operation – Quality Control – Process Control fluid samples & Review the Control sample results.	Pass	Kailas Ghait 05/12/2020
03	Accuracy Analysis	Compare the obtained Q.C value with mean of expected value as mentioned in the QC Value chart.	Pass	Kailas Ghait 05/12/2020

**Test III**

**Test Name** : **Precision Study**  
**Purpose** : To see the precision performance of the equipment  
**Method** : Microslide – Rate Chemistry & Endpoint Chemistry  
 Microslide – Potentiometric Chemistry;  
 Microslide – Immunorate Chemistry;  
 Microtip Chemistry  
 Microsensor Chemistry  
 Microwell - Chemiluminescence Immunoassay

- Analyze Vitros Performance Verifier Level 1 control for the following tests :  
 ALT (5 x 7 times), Na<sup>+</sup> (5 x 8 times), dLDL (6 x 6 times), BuBc (5 x 7 times)
- Analyze TDM Performance Verifier Level 3 for Phenytoin (5 x 7 times).
- Analyze TDM Performance Verifier Level 1 for Gentamycin (6 x 6 times).
- Analyze Protein Performance Verifier Level 1 for IgM (6 x 6 times).
- Analyze Microsensor Check Fluid Level I for Hemolysis, Icterus and Turbidity (20 times each).
- Analyze all the three levels of Vitros Total Thyroid controls for Microwell – Chemiluminescence Immunoassay TT4 and TSH (10 times each).
- Calculate the Mean, SD and CV%.

**Acceptance Criteria :**

Sr. No.	Analyte	Control Level	Precision Limit
01	ALT	PV I	≤ 2.2 SD
02	Sodium	PV I	≤ 0.69% CV
03	dLDL	PV I	SD < 1.1
04	Bu	PV I	≤ 0.3 SD
05	Bc	PV I	≤ 0.6 SD

Ortho Clinical Diagnostics

06	Phenytoin	TDM PV III	$\leq 0.75$ SD
07	Gentamycin	TDM PV I	SD < 0.11
08	IgM	Protein PV I	SD < 2.1
09	Hemolysis	MS Check Fluid Level I	SD < 4.4
10	Icterus	MS Check Fluid Level I	SD < 0.4
11	Turbidity	MS Check Fluid Level I	SD < 14.8
12	TT4	Total Thyroid Control Level 1	CV% $\leq 4.9$
		Level 2	CV% $\leq 4.6$
		Level 3	CV% $\leq 6.0$
13	TSH	Total Thyroid Control Level 1	CV% $\leq 10.9$
		Level 2	CV% $\leq 4.9$
		Level 3	CV% $\leq 4.7$

\* Reference: The acceptance SD & CV% has been taken from Manufacturer's recommended limits.

**Table for Record the Obtained Q.C Values :**

**Test Name:**

**Control Level:**

**Control Lot:**

**Expiry Date:**

Sr. No	Obtained Q.C Value	Range of the Comparator	Mean of the Comparator	Remarks	Done By
				Pass / Fail	Date
01					
02					
03					
04					
05					
06					
07					
08					
09					
10					

**V. COMMENTS:**

**Deviation:**

Nil

**Impact On Operation:**

Nil

**Corrective Action:**

Nil




## Ortho Clinical Diagnostics

### VI. System Certification

Study data has determined that the system described in this document either meets all criteria outline in this Performance Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for specified usage.


Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Kailas Ghait  
Designation : Sr. Territory Manager  
Company : Ortho Clinical Diagnostics


Signature:   
Date: 05/12/2020

Validation Team Central Clinical Laboratory, Pravara Rural Hospital:

Name : Mr. K. N. Hattar  
Designation : ASSO. PROF.  
Department : Biochemistry

Signature:   
Date: 05/12/2020

Name : Mr. Bhawar S. R.  
Designation : Lab Technician  
Department : Biochemistry

Signature:   
Date: 05/12/2020

Customer Authorizations:

Name : Dr. S. N. Jogle Sir  
Designation : Prof. & Head Biochemistry  
Signature : S. N. Jogle  
Date : 05/12/2020


**PROFESSOR & HEAD**  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413736, (M.S.) INDIA

Ortho Clinical Diagnostics

**OPERATIONAL QUALIFICATION**

**For**

**VITROS® 5600 INTEGRATED SYSTEM**



Manufactured by:  
Ortho Clinical Diagnostics, Inc., US

## Table of Contents


<b>Sr. No.</b>	<b>Contents</b>	<b>Page No.</b>
I	Approval of the OQ procedure	3
II	Instructions	4
III	Scope	5
IV	Operational Qualification	6
IV.I	Operational Procedure	7
V	Operational Qualification Report	21
VI	Comments	23
VII	System Certification	24

**I. APPROVAL OF THE OQ PROCEDURE:**

Both Central Clinical Laboratory Pravara Rural Hospital and Ortho Clinical Diagnostics are responsible for Operational check of VITROS® 5600 Integrated System bearing Sr. No. J56001041 installed in Central Clinical Laboratory Pravara Rural Hospital, Loni as per the attached protocol.

Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Kailas Ghait

Signature: 


Designation : Sr. Territory Manager

Date: 10/12/2020

Company : Ortho Clinical Diagnostics

Validation Team from Central Clinical Laboratory Pravara Rural Hospital:

Name : Mr. K. N. Hattar


Signature: 

Designation : ASSO. PROF.

Date: 10/12/2020

Department : Biochemistry

Name : Mr. Bhawar S. R.

Signature: 

Designation : Sr. Lab Tech

Date: 10/12/2020

Department : Biochemistry

Customer Authorizations:

Name : Dr. S. N. Jangle Sir

Designation: Prof. & Head Biochemistry

Signature : S. N. Jangle

Date : 10/12/2020

**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413736, (M.S.) INDIA**

**II. INSTRUCTIONS:**

1. This document is to be verified / completed at the time, the system is going for operational check of each purpose and set up for operation.
2. An authorized Ortho Clinical Diagnostics representative will check the system and enter the specific data as outlined in the appropriate Operational Qualification.
3. Employees of Central Clinical Laboratory, Pravara Rural Hospital will verify each result and sign in each page. The member of the validation team will carry out this procedure.
4. ~~ALL deviations from normal specification during operation check will be noted under~~ COMMENTS. All resolution to such problems will also be noted in the COMMENTS section. Additional space is provided at the end of this protocol for the same.

### III. SCOPE

This Operational Qualification protocol will be performed on the VITROS® 5600 Integrated System, bearing Sr. No. J56001041 located at Central Clinical Laboratory, Pravara Rural Hospital Loni.

This protocol will define the documentation that will be used to evaluate the instrument's operational check in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been operated in accordance with the intended usage.

~~Operational checks will also be performed to verify that the instrument has been operated with proper information / sequence and utilities.~~

Trained, knowledgeable personnel will perform operational Qualification studies.

Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated, and the appropriate course of action determined. All documents will be initialed and dated.

## IV. OPERATIONAL QUALIFICATION:

## A. Instrument Identification

Verified Date

1. Model Name VITROS® 5600 Integrated System

2. Serial Number J56001041 08/12/2020

## B. Following is a list of tests to be performed and verified:

Test No.	Test Name	Test purpose	Initial / Date
01	System centers Overview	To make the operator to identify the instrument subsystem.	Kailas Ghait 08/012/2020
02	Start up & Shutdown	To make the equipment ready for operation.	Kailas Ghait 08/012/2020
03	User Inter Face Overview	Different functionality of software utility available for the operator interaction.	Kailas Ghait 08/012/2020
04	Sample programming and Analysis	To process samples either by manual assigning or through LIS.	Kailas Ghait 08/012/2020
05	Performing Calibration	To calibrate the system for every new lot of assay or after calibration expiry.	Kailas Ghait 08/012/2020
06	Maintenance & System clean	To perform maintenance process to keep the system operating properly.	Kailas Ghait 08/012/2020
07	Reagent Management & supply	To update & monitor the status of reagents required for assay processing.	Kailas Ghait 08/012/2020
08	Performing Quality control	To confirm that systems, reagents and consumables are acceptable and working within specifications for each assay used	Kailas Ghait 08/012/2020
09	Result Review	To review the processed results in the system.	Kailas Ghait 08/012/2020
10	Result Intellicheck.	To check the Intellicheck function of the system.	Kailas Ghait 08/012/2020
11	Option & Configuration	To setup the system as per Laboratory requirement.	Kailas Ghait 08/012/2020

**Test: 1** : **System Hardware Overview**  
**Purpose** : To make the operator to identify the instrument subsystem.  
**Reference** : Operator Reference Guide - Pages 4-1 to 4-5

**Summary:**

For better understanding purpose, Instrument has been divided into several parts according to its operation mode, so we call this partition as centers. And those centers are named as mention below.

**Procedure:**

This will list the available system centers in the instrument and its subsystem, contend to operator understanding. The operator has to overview the Service V-Docs to get an idea about the system centers overview.

- Sampling Centers
  - Sample Supply
  - Primary Tip sealer
  - Micro sensor subsystem
- Micro Immunoassay Center
  - Micro Immunoassay Metering & Reagent Metering
  - Micro Immunoassay Versa Tip Ring
  - Micro Well Incubator
  - Micro Well Wash Assembly
  - Signal Reagent Assembly
  - Luminometer
- Command Center
  - Master Computer & Monitor
  - Keyboard & Touch system
- System Frame and Cabinetry



- Test: 2** : **Starting Up and Shutting Down**
- Purpose** : To make the instrument 'READY' for operation
- Reference** : Operator Reference Guide - Pages 3-0 to 3-25

**Summary:**

Instrument will check status of different parts of the instrument automatically after booting up to system status screen; if there is an error code posted, initialize the system and follow corrective action instructions provided for the error code.

**Procedure:**

**Starting Up the System**

- Check the room temperature and switch on the Air Conditioner.
- Check the UPS.
- Switch on the Printer & Load the paper.
- Switch on the VITROS<sup>®</sup> 5600 Integrated System by lift the main switch up and hold it for about 5 sec.
- Wait for the instrument to get ready after initialization.
- The machine is ready for next step if it displays "READY" on the status console.
- If not, initialize by pressing the initialize button on the error code screen.
- Follow the instructions provided for the error codes.

**Shutting Down the system**

- Touch **Shutdown** in the main menu.
- Press 'Y/N' to continue the shutdown process.
- Desire normal shutdown or final shutdown and then select desired menu.
- If, you selected **Final Shutdown**, press the RESET button to restart the instrument, or press the Main power switch down to make it OFF.

## Ortho Clinical Diagnostics

<b>Observation</b>	System status console shows "Ready". Instrument is ready for operation	<b>Remarks</b>	<b>Initial/Date</b>
		Pass	Kailas Ghait 08/12/2020

**Test: 3** : **User Interface Overview**


**Purpose** : To make the operator to understand the system screens.

**Reference** : Operator Reference Guide - Pages 6-0 to 6-12

**Summary:**

The System status screen will make the operator to understand of different functionality of software utility available for the operator. This will helps operator to check system status as well as to instruct any commands to the system for required operation.

Procedure:



The above (as shown in the Previous Page) picture is actual system software screen and the number in the blue circle is to identify the several functionality of the software icons designed for the operator to interact with the system. We call these software icon as mentioned below.,

1. Status Line
2. Status Console
3. Function Screen
4. Prompt Line

## Ortho Clinical Diagnostics

5. Time, Date and Version Display
6. Status Indicators
7. Process Buttons

**Test: 4** : **Sample programming and Analysis**  
**Purpose** : To program and process the samples  
**Reference** : Operator Reference Guide (pages 9-1 to 9-13)

**Summary:** The operator can process assay by assigning program manually in the Universal Sample tray in '**Sample Program**' menu or they can download sample program through '**LIS**'. Sample programming is the process of selecting assays and programming characteristics for samples. The system uses the sample program to meter appropriate sample and select the right reagent for the assay, process and then report results with the correct identification.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Sample Programming methods & Overview	Desire the sample Programming method.	Pass	Kailas Ghait 09/12/2020
02	Loading and Processing of samples	Tray cover opened and samples loaded in sample tray placed in sample supply.	Pass	Kailas Ghait 09/12/2020
03	Programming samples	Sample program assigned for selected tray.	Pass	Kailas Ghait 09/12/2020
04	Processing samples	Samples are processed automatically by the system.	Pass	Kailas Ghait 09/12/2020
05	Unloading the samples	Tray cover opened and processed sample tray unloaded.	Pass	Kailas Ghait 09/12/2020
06	Viewing samples in process	Sample under process are displayed on the 'View Sample Status' Screen.	Pass	Kailas Ghait 09/12/2020

- Test: 5** : **Performing Calibration**
- Purpose** : To calibrate the system for every new lot of assay
- Reference** : Operator Reference Guide (pages 10-1 to 10-9)

**Summary:** The system requires its own calibration for every individual assay to measure the analytic concentration as well as to accept reagent pack status ready for processing. Assay calibration is a process that relates the response of the system to analyte concentration or activities. Calibration is performed periodically to adjust for changes in the system, assay protocols, or assay reagent lots.

The system requires calibration for individual assays when:

- A new assay is uploaded to the system
- The calibration expires (up to 28 days after it is processed, depending on the assay; refer to the package insert for expiration information)
- An assay reagent lot number changes
- Government regulations specified
- An assay's protocol changes

You also may need to perform calibration when:

- Certain service procedures are performed
- Quality control performance is out of range

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Load New ADD via CD or Downloaded file (New Gen Lot, Protocol , reagent lot calibration , Diluents lot information).	Calibrator identified and updated for protocol & master calibration data.	Pass	Kailas Ghait 09/12/2020
02	Preparing calibrators.	Calibrators are ready for processing.	Pass	Kailas Ghait 09/12/2020
03	Performing Calibration with Bar code label.	System recognizes the barcode and processing the calibration automatically.	Pass	Kailas Ghait 09/12/2020
04	Performing Calibration with Sample processing screen.	In the sample programming, the calibration program is assigned and processed the calibration assigned for each assay.	Pass	Kailas Ghait 09/12/2020

## Ortho Clinical Diagnostics

05	Calibration report.	Calibration completed successfully. Report printed.	Pass	Kailas Ghait 09/12/2020

**Test: 6** : **Maintenance & System Clean**

**Purpose** : Clean appropriate modules to maintain Accuracy and precision.

**Reference** : Operator Reference Guide 16-13 to 16-17

**Summary:**

~~Maintenance procedures are tasks that are performed to keep the system operating properly.~~  
Maintenance protocols to be performed according to the recommended schedule (daily, weekly, monthly, or as required). Ensure that we need to use 70% Isopropyl alcohol to disinfect the appropriate module to keep cleanliness and maintain the accuracy & precision.

The Maintenance is classified into four-category ie.,

1. Daily Maintenance
2. Weekly Maintenance
3. Monthly Maintenance
4. As required Maintenance

**Daily Maintenance:** Pages from 16-14 to 16-15

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Perform Metering Maintenance	Metering Maintenance Performed	Pass	Kailas Ghait 09/12/2020
02	Empty Solid and Liquid waste container	Solid & Liquid waste containers are emptied.	Pass	Kailas Ghait 09/12/2020
03	Remove outdated or empty reagent packs, Signal Reagent packs and Universal Wash Buffer	Outdated empty Reagent packs, SR packs and UWR bottles are removed & discarded.	Pass	Kailas Ghait 09/12/2020
04	Inspect sample trays and adaptors	Sample Trays are cleaned.	Pass	Kailas Ghait 09/12/2020
05	Clean the SR Probe assembly	SR Probes are cleaned.	Pass	Kailas Ghait

## Ortho Clinical Diagnostics

				09/12/2020
06	Load required reagent packs, Signal Reagent pack and Universal Wash Buffer.	All the required reagent packs are loaded and updated in the system as required.	Pass	Kailas Ghait 09/12/2020
07	Run Q.C fluids	Q.C samples are processed successfully.	Pass	Kailas Ghait 09/12/2020

### Weekly Maintenance: Pages from 16-15 to 16-16

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Clean the Micro well Incubator.	Micro well incubator – Inner ring, outer ring, middle ring, shuttle weight, drop holes, Luminometer FOB, Wash reagent and signal reagent probes are cleaned.	Pass	Kailas Ghait 09/12/2020
02	Clean the Primary tip sealer.	Primary Tip Sealer cleaned.	Pass	Kailas Ghait 09/12/2020
03	Clean the Secondary Tip Sealer.	Secondary Tip Sealer cleaned.	Pass	Kailas Ghait 09/12/2020
04	Clean the Sample Supply and Cap Retainer.	Cap Retainer & Sample Supply cleaned.	Pass	Kailas Ghait 09/12/2020
05	Clean the Touch Screen Monitor & Keyboard.	Touch Screen Monitor & Key board Clean Done.	Pass	Kailas Ghait 09/12/2020
06	Run the maintenance Pack for Subsystem cleaning.	Maintenance pack is loaded; subsystem cleaning done automatically by the system.	Pass	Kailas Ghait 09/12/2020
07	Run QC Fluids	Q.C Processed successfully.	Pass	Kailas Ghait 09/12/2020

Ortho Clinical Diagnostics

Monthly Maintenance: Pages from 16-15 to 16-16

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Clean Micro sensor Cover & Ring Area.	Micro Sensor Cover & Ring Surface cleaned.	Pass	Kailas Ghait 09/12/2020
02	Inspect/Clean Micro Immuno Assay reagent Supply top Cover.	Micro Immuno Assay Reagent Supply top Cover Inspected and Cleaned.	Pass	Kailas Ghait 09/12/2020
03	Clean VITROS Versa Tip supply Registration Rail.	Versa tip Supply Registration Rail cleaned.	Pass	Kailas Ghait 09/12/2020
04	Inspect Reagent cooler filter for cleanliness.	Reagent cooler filter removed & cleaned.	Pass	Kailas Ghait 09/12/2020
05	Replace Vapors adsorption cartridge for every two months.	Every two months once, VAC replaced.	Pass	Kailas Ghait 09/12/2020
06	Make a backup of Q.C, Calibration and Configuration.	Backup of QC, Calibration and Configuration made successfully.	Pass	Kailas Ghait 09/12/2020
07	Inspect / Clean Master Computer Filter.	Inspected and Cleaned Master computer Air Filter.	Pass	Kailas Ghait 09/12/2020



**Test: 7** : **Managing reagents Inventory and Supply**

**Purpose** : To Maintain & monitor the status of reagents or supply required for assay processing.

**Reference** : Operator Reference Guide (pages 15-1 to 15-7)

**Summary:** The Reagent Management feature enables you to review current inventory information for the reagents loaded on the system. Using this function, you can load and unload reagents as necessary. To maintain the required reagents in the system for processing, the operator should review the Reagent management screen.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass/Fail	Date
01	Review the reagent inventory to plan for the day.	The reagent inventory for the day planned.	Pass	Kailas Ghait 09/12/2020
02	Loading of Reagent Pack automatically	Required Reagent Packs loaded automatically by software request.	Pass	Kailas Ghait 09/12/2020
03	Loading of Reagent Pack with help of Manual Lot Entry button.	Requested Lot Information fed and the reagent pack loaded.	Pass	Kailas Ghait 09/12/2020
04	Loading of Signal Reagent automatically.	SR Pack loaded in position 1 & 2 and accepted by barcode reading.	Pass	Kailas Ghait 09/12/2020
05	Loading of Signal Reagent with Manual Load Button.	SR packs information fed and loading done.	Pass	Kailas Ghait 09/12/2020
06	Loading of Universal Wash Buffer	UWR buffer loading done through Load supply Software icons.	Pass	Kailas Ghait 09/12/2020
07	Unloading of Reagents	The entire Empty & expired reagents packs are unloaded by Load/Unload software icons.	Pass	Kailas Ghait 09/12/2020

**Test: 8 : Performing Quality control**

**Purpose :** Quality Control (QC) is important in determining the performance and accuracy of the system. To perform Quality Control, QC materials are run with either known, or unknown values along with patient samples to determine whether the system is functioning within the established ranges for your lab.

**Reference :** Operator Reference Guide (Pages 9-6 to 9-8)

**Summary:** Performing quality control procedures is an important part of using or maintaining the system. This section explains:

- When you should perform quality control
- How to choose a control fluid

The recommended frequency for processing quality control fluids is once in every 24 hours. However, the frequency with which you perform quality control procedures may vary, depending on the requirements and regulations for processing control fluids by your national, state, provincial, and local governments. Quality control procedures within your own Laboratory may also require a different frequency. You should also perform quality control procedures when:

- Assays have been calibrated
- Certain service procedures are performed, other than routine maintenance

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Choosing the Control fluid	Required control fluid identified.	Pass	Kailas Ghait 09/12/2020
02	Preparing Liquid or Lyophilized control fluids	Control fluid prepared and ready for processing.	Pass	Kailas Ghait 09/12/2020
03	Creating QC file.	Q.C file created for assay in the system according to control fluid.	Pass	Kailas Ghait 09/12/2020
04	Process QC samples	QC samples are programmed in the sample programming window and the QC samples are loaded and processed automatically	Pass	Kailas Ghait 09/12/2020
05	Review Q.C result.	Processed Q.C results are reviewed and found satisfactory.	Pass	Kailas Ghait 09/12/2020
06	Display & printing graph.	Q.C graph reviewed and printed.	Pass	Kailas Ghait 09/12/2020

## Ortho Clinical Diagnostics

07	Managing Quality control Reports	Required reports printed and filed.	Pass	Kailas Ghait 09/12/2020
----	----------------------------------	-------------------------------------	------	----------------------------

**Test: 9** : **Result Review.**

**Purpose** : To review the processed results in the system.

**Reference** : Operator Reference Guide (Pages 11-1 to 11-6)

**Summary:** The Results Review function helps to evaluate result records for patient and quality control samples. The results will be displayed along with the Reagents Lot information, Dilution information & if there is any error codes or Flags.

Result records contain the data generated by the system when assays are processed. The system can store up to 25,000 result records. When this limit is reached, new result records overwrite the oldest records. The system permanently deletes the overwritten records from computer memory.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Update List.	Sample under process status displayed with all information.	Pass	Kailas Ghait 09/12/2020
02	Monitoring Results.	Completed Recent Assay Results displayed on the screen.	Pass	Kailas Ghait 09/12/2020
03	Filter Results.	Processed Assay Results displayed as per the selected criteria.	Pass	Kailas Ghait 09/12/2020
04	Edit Patient Data	User can Edit/Add Patient Demography information, but the Patient ID will remain same.	Pass	Kailas Ghait 09/12/2020
05	Retrieving and Reviewing Archive Results by Set Report Status.	Archived Results are updated successfully in the CD/Pen Drive. The same Retrieved from the CD/Pen Drive.	Pass	Kailas Ghait 09/12/2020
06	Managing Reports by Set Report Status.	Required Reports got printed and filed.	Pass	Kailas Ghait 09/12/2020
07	Integrated Codes and Flags	Reported Codes and Flags are referred in Flags and Code chart. Necessary corrective action taken.	Pass	Kailas Ghait 09/12/2020

**Test: 10** : **Result Intellicheck.**

**Purpose** : To check the Integrity of Performed assays.

**Reference** : Operator Reference Guide (Pages 11-8 to 11-9).

**Summary:** The Result Intellicheck screen to view Intellicheck Technology Verifications performed for each sample and assay processed. For the selected result record, you can view verification data and detected exceptions for each analyte.

Example:

~~To Analyze Intelli Report, select the sample ID listed on result review screen.~~

- On Review results screen, touch the 'Result Intellicheck' Icon.
- Result Intellicheck report comes on the screen.
- Check the parameters of Sample Metering, Sample + Reagent volume, Signal reagent volume and well wash verification for their acceptance.
- Take print of the Result Intellicheck report.

**Acceptance criteria:**

Sr. No.	Parameter	Acceptance limit	Remarks	Done By
			Pass / Fail	Date
01	<b>Sample Metering</b> <ul style="list-style-type: none"> <li>• Clot</li> <li>• Bubble</li> <li>• Short sample</li> <li>• Viscosity</li> <li>• Thin layer of fluid</li> </ul>	Against all the parameters, "No" should be displayed on screen	Pass	Kailas Ghait 09/12/2020
02	<b>Reagent Metering</b> Sample + Reagent volume	No Exception. Range: 12700 – 19000	Pass	Kailas Ghait 09/12/2020
03	Signal Reagent	No Exception Range: 17500 – 22800	Pass	Kailas Ghait 09/12/2020
04	Well wash verification	No Exception Range: 21300 – 25000	Pass	Kailas Ghait 09/12/2020
05	Luminometer – Self calibration	No Exception	Pass	Kailas Ghait 09/12/2020

**Test: 9 : Option & Configuration**

**Purpose** : To setup the system as per laboratory requirement

**Reference** : Operator Guide (Pages 11-8 to 11-14)

**Summary:** The Options & Configuration function provides many features for customizing your VITROS® 5600 Integrated System. It is organized into three main groups:

- Configure Analyte Data & Review/Edit Calibration Data
- System Setup
- System Services

Selections within these groups allow you to customize analyte parameters, review calibration data, perform user calibrations, configure and set system parameters, review usage inventory, configure e-Connectivity® and network device parameters, configure printer, laboratory computer and auxiliary ports, perform backup for quality control, calibration and configuration files and perform an archive procedure for result records.

**Procedure:**

Sr. No.	Activity	Observation	Remarks	Done By
			Pass / Fail	Date
01	Configure Assays	Analyte parameters are configured as desired by laboratory.	Pass	Kailas Ghait 09/12/2020
02	Review / Edit Calibration Data	Calibration data updated.	Pass	Kailas Ghait 09/12/2020
03	Configure System Setup	Required subsystem configuration is done successfully.	Pass	Kailas Ghait 09/12/2020
04	Configure Subsystem Setup	The required subsystem can be disabled / deactivated as per needs.	Pass	Kailas Ghait 09/12/2020
05	Configure Report Control	System report parameter has set for printer & LIS.	Pass	Kailas Ghait 09/12/2020
06	Configure Communication	System interface protocol set for Laboratory Information System (LIS), Ethernet and e-Connectivity® communications,	Pass	Kailas Ghait 09/12/2020
07	Configure Demography	Global demographic attributes to be used when configuring age, sex, and normal ranges for specific	Pass	Kailas Ghait 09/12/2020

## Ortho Clinical Diagnostics

		assay/body fluids is defined.		
08	System Services 1.Datalogger 2.Perform Backup 3.Usage Counters 4.Option Summary 5.Load System Data	Shall be performed and reviewed as & when required.	Pass	Kailas Ghait 09/12/2020

### V. Operational procedure:

#### ~~Certificate of Training~~

#### 1. Technician Training

This certifies that the technicians have received basic user training in the following categories for the system described in this Operational Qualification.

Mr. Kailas Ghait, Sr. Territory Manager from Ortho Clinical Diagnostics has conducted the training.

Sr. No.	Training program	Initials	Date
1.	Instrument Setup	Kailas Ghait	
2.	System Operation	Kailas Ghait	
3.	Basic trouble shooting and Maintenance	Kailas Ghait	

2. Operator Training

The users responsible for the operation of this equipment have been trained in the proper usage of the system. Training focused on the basic operation and maintenance of the system.

Sr. No.	Operators	Department	Initials	Date
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				

b. Customer SOP / Manuals:

Title	Number	Version	Verified by	Date
Vitros 5600 Integrated System Reference Guide				
Microslide Instruction for use Manual				
Microtip Chemistry Instruction For Use Manual				
Micro Well Instruction For Use Manual				

**VI. COMMENTS :**

Deviation:

Nil

Impact On Operation:

Nil

Corrective Action:

Nil




**VII. SYSTEM CERTIFICATION**

Study data has determined that the system described in this document either meets all criteria outlined in this Operational Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been / not been addressed. The system is ready for Performance Qualification.


Protocol Performed By: Ortho Clinical Diagnostics Representative

Name : Kailas Ghait  
Designation : Sr. Territory Manager  
Company : Ortho Clinical Diagnostics


Signature:   
Date: 10/12/2020

Validation Team From Central Clinical Laboratory Pravara Rural Hospital:

Name : Mr. K. N. Hotkar  
Designation : Asso. prof.  
Department : Biochemistry

Signature:   
Date: 10/14/2020

Name : Mr. Shrawan S. P.  
Designation : Sr. Lab Tech.  
Department : Biochemistry

Signature:   
Date: 10/14/2020

Customer Authorizations:

Name : Dr S. N. Jangle Sir  
Designation : Prof. & Head Biochemistry  
Signature : S. N. Jangle  
Date : 10/12/2020

**PROFESSOR & HEAD  
DEPARTMENT OF BIOCHEMISTRY  
RURAL MEDICAL COLLEGE  
LONI-413738, (M.S.) INDIA**

ASSAY: CHOL  
FLUID: Serum  
Status: Passed  
OP ID:

KIT LOT: 0230

CAL DATE/TIME: 7/12/2020 12:34:11  
REAGENT GEN/LOT: 08435770

SITE TEMP: 29.3C

SUPPLY HUMIDITY: 33

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 6.83330			mg/dL	
SLOPE: 93.9901	1	1	48	0.502765
CURVATURE/SLOPE2: 0.731799	2	2	183	0.917953
	3	3	421	1.30600

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:20:11

ASSAY: CHOL

CAL DATE/TIME: 24/1/2022 12:41:31

FLUID: Serum

KIT LOT: 0230

REAGENT GEN/LOT: 08459641

Status: Passed

OP ID:

SITE TEMP: 27.3C

SUPPLY HUMIDITY: 33

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 5.02189			mg/dL	
SLOPE: 93.7497	1	1	45	0.490888
CURVATURE/SLOPE2: 0.812442	2	2	184	0.904349
	3	3	410	1.29182

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:20:06

ASSAY: Mg

CAL DATE/TIME: 7/12/2020 16:31:09

FLUID: Serum

KIT LOT: 0189

REAGENT GEN/LOT: 32064887

Status: Passed

OP ID:

SITE TEMP: 29.2C

SUPPLY HUMIDITY: 32

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.345463			mg/dL	
SLOPE: 101.261	1	1	0.7	0.964788
CURVATURE/SLOPE2: 21.1311	2	2	3.6	1.37181
	3	3	9.3	2.01105

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:19:32

ASSAY: Mg

CAL DATE/TIME: 16/9/2021 11:21:29

FLUID: Serum

KIT LOT: 0110

REAGENT GEN/LOT: 32068342

Status: Passed

OP ID:

SITE TEMP: 28.8C

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE
INTERCEPT: 0.361532				
SLOPE: 101.621	1	1	0.8	0.976550
CURVATURE/SLOPE2: 103.589	2	2	3.8	1.38205
	3	3	9.6	1.99126

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:19:19

ASSAY: dHDL  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.4C

KIT LOT: 2599

CAL DATE/TIME: 7/12/2020 12:35:54  
 REAGENT GEN/LOT: 11256686

SUPPLY HUMIDITY: 33

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE
INTERCEPT: -0.901667				
SLOPE: 110.023	1	1	7	0.331437
CURVATURE/SLOPE2: -11.6830	2	2	.40	0.611254
	3	3	121	0.900070

EXPIRED CALIBRATOR

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:20:53

ASSAY: dHDL  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.2C

KIT LOT: 2540

CAL DATE/TIME: 5/1/2022 12:42:27  
 REAGENT GEN/LOT: 11322283

SUPPLY HUMIDITY: 31

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.244157			mg/dL	
SLOPE: 93.8445	1	1	7	0.342356
CURVATURE/SLOPE2: 9.57231	2	2	34	0.598629
	3	3	108	0.836760

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:20:43

ASSAY: URIC  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.2C

KIT LOT: 0189

CAL DATE/TIME: 7/12/2020 16:30:12  
 REAGENT GEN/LOT: 05386491

SUPPLY HUMIDITY: 32

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE
INTERCEPT: 0.0487304	1	1	0.7	0.193693
SLOPE: 103.169	2	2	5.9	0.692233
CURVATURE/SLOPE2: 11.5468	3	3	16.5	1.34432

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:18:11



ASSAY: URIC  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 27.3C

KIT LOT: 0110

CAL DATE/TIME: 27/11/2021 16:02:07  
 REAGENT GEN/LOT: 05489467

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.00268656			mg/dL	
SLOPE: 102.623	1	1	0.9	0.236541
CURVATURE/SLOPE2: 8.31990	2	2	5.8	0.707108
	3	3	16.6	1.38348

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:18:04

ASSAY: PHOS  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.2C

KIT LOT: 0189

CAL DATE/TIME: 7/12/2020 16:30:31  
 REAGENT GEN/LOT: 12494857

SUPPLY HUMIDITY: 32

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.0524050			mg/dL	
SLOPE: 97.3316	1	1	1.1	0.234234
CURVATURE/SLOPE2: 89.6172	2	2	5.3	0.468303
	3	3	12.5	0.690562

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:19:48

ASSAY: PHOS  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.0C

KIT LOT: 0110

CAL DATE/TIME: 9/12/2021 10:53:30  
 REAGENT GEN/LOT: 12483461

SUPPLY HUMIDITY: 35

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE
INTERCEPT: 0.546283				
SLOPE: 86.2334	1	1	1.2	0.217389
CURVATURE/SLOPE2: 213.892	2	2	5.6	0.473451
	3	3	12.9	0.680923

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:19:40

ASSAY: UREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.2C

KIT LOT: 0189

CAL DATE/TIME: 7/12/2020 16:29:34  
REAGENT GEN/LOT: 01315247

SUPPLY HUMIDITY: 32

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.341656			mg/dL(A)	
SLOPE: 102.395	1	1	7	0.136546
CURVATURE/SLOPE2: -1.60288	2	2	107	0.974153
	3	3	233	1.80207

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:13:20

ASSAY: UREA  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 27.6C

KIT LOT: 0110

CAL DATE/TIME: 30/11/2021 14:31:37  
 REAGENT GEN/LOT: 01318070

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL(A)	RESPONSE
INTERCEPT: 0.209745				
SLOPE: 99.6223	1	1	8	0.149922
CURVATURE/SLOPE2: 7.26464	2	2	105	0.949782
	3	3	236	1.73500

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:13:12

ASSAY: AMYL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.4C

KIT LOT: 0310

CAL DATE/TIME: 7/12/2020 16:52:50  
REAGENT GEN/LOT: 60052712

SUPPLY HUMIDITY: 31

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 6.31413			U/L	
SLOPE: 97.5707	1	1	25	0.006800
CURVATURE/SLOPE2: -0.764502	2	2	359	0.046370
SUB DEP DENS: 1.23509	3	3	978	0.106931

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:17:18

ASSAY: AMYL  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.5C

KIT LOT: 0320

CAL DATE/TIME: 7/10/2021 10:58:30  
 REAGENT GEN/LOT: 60169168

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 11.1882			U/L	
SLOPE: 112.761	1	1	51	0.009416
CURVATURE/SLOPE2: 2.95956	2	2	392	0.045173
SUB DEP DENS: 1.18001	3	3	997	0.091037

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:17:14

ASSAY: Ca  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.2C

KIT LOT: 0189

CAL DATE/TIME: 7/12/2020 16:30:50  
REAGENT GEN/LOT: 03353622

SUPPLY HUMIDITY: 32

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.0592083			mg/dL	
SLOPE: 95.9383	1	3	1.9	0.672478
CURVATURE/SLOPE2: 0.000000000000	2	2	9.0	1.09009
	3	1	14.2	1.30235

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:18:43



ASSAY: Ca

CAL DATE/TIME: 30/11/2021 14:32:53

FLUID: Serum

KIT LOT: 0110

REAGENT GEN/LOT: 03039170

Status: Passed

OP ID:

SITE TEMP: 27.5C

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE
INTERCEPT: 0.0384845	1	3	1.7	0.683448
SLOPE: 95.6149	2	2	9.0	1.12001
CURVATURE/SLOPE2: 0.000000000000	3	1	13.9	1.33930

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:18:35

ASSAY: GLU  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.2C

KIT LOT: 0189

CAL DATE/TIME: 7/12/2020 16:29:53  
 REAGENT GEN/LOT: 00493273

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.382906			mg/dL	
SLOPE: 100.090	1	1	29	0.251579
CURVATURE/SLOPE2: 0.0730982	2	2	264	1.04565
	3	3	574	1.58001

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:10:46

ASSAY: GLU  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.7C

KIT LOT: 0110

CAL DATE/TIME: 17/12/2021 10:33:26  
 REAGENT GEN/LOT: 00098952

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE
INTERCEPT: 1.63615				
SLOPE: 96.2517	1	1	33	0.266175
CURVATURE/SLOPE2: 0.584552	2	2	293	1.12835
	3	3	565	1.57345

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:10:25

ASSAY: CREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.2C

KIT LOT: 0189

CAL DATE/TIME: 7/12/2020 16:29:15  
REAGENT GEN/LOT: 15316476

SUPPLY HUMIDITY: 13

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.00680375			mg/dL	
SLOPE: 105.455	1	1	0.4	0.002851
CURVATURE/SLOPE2: 0.000000000000	2	2	1.5	0.009460
SUB DEP DENS: 0.454137	3	3	13.1	0.069085

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:14:05

ASSAY: CREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 27.6C

KIT LOT: 0110

CAL DATE/TIME: 30/11/2021 14:31:11  
REAGENT GEN/LOT: 15319728

SUPPLY HUMIDITY: 13

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.0541661			mg/dL	
SLOPE: 111.969	1	1	0.6	0.003826
CURVATURE/SLOPE2: 0.000000000000	2	2	1.5	0.009536
SUB DEP DENS: 0.430258	3	3	13.2	0.065823

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:14:01

ASSAY: ALTV  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.4C

KIT LOT: 0310

CAL DATE/TIME: 7/12/2020 12:33:03  
 REAGENT GEN/LOT: 56024772

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.529878			U/L	
SLOPE: 104.161	1	1	6	0.008772
CURVATURE/SLOPE2: -0.0741566	2	2	219	0.137006
SUBS DEPL DENS: 0.517170	3	3	792	0.402204
DELTA DENSITY: NA				
1ST POINT REF: 3.22908				

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:15:55

ASSAY: ALTV  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 26.8C

KIT LOT: 0320

CAL DATE/TIME: 27/11/2021 17:35:12  
 REAGENT GEN/LOT: 56060225

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.180227			U/L	
SLOPE: 100.658	1	1	9	0.011726
CURVATURE/SLOPE2: 0.252039	2	2	216	0.137514
SUBS DEPL DENS: 0.505881	3	3	761	0.404376
DELTA DENSITY: NA				
1ST POINT REF: 3.15859				

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:15:50

ASSAY: AST  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.4C

KIT LOT: 0310

CAL DATE/TIME: 7/12/2020 12:32:06  
 REAGENT GEN/LOT: 73265542

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.177577			U/L	
SLOPE: 101.982	1	1	16	0.005903
CURVATURE/SLOPE2: -0.355190	2	2	223	0.078792
SUBS DEPL DENS: 0.525980	3	3	750	0.263029
DELTA DENSITY: NA				
1ST POINT REF: 7.47784				

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:14:49



ASSAY: AST  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.1C

KIT LOT: 0320

CAL DATE/TIME: 9/3/2022 11:47:24  
REAGENT GEN/LOT: 73285153

SUPPLY HUMIDITY: 13

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 1.44557			U/L	
SLOPE: 98.6854	1	1	15	0.005001
CURVATURE/SLOPE2: 0.983398	2	2	208	0.072998
SUBS DEPL DENS: 0.489488	3	3	690	0.226543
DELTA DENSITY: NA				
1ST POINT REF: 7.03013				

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:14:34

ASSAY: TBIL  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 30.3C

KIT LOT: 0460

CAL DATE/TIME: 10/1/2022 12:54:51  
 REAGENT GEN/LOT: 14483001

SUPPLY HUMIDITY: 35

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -3.55472			mg/dL		
SLOPE: 31.8012	1	1	1.1	0.189014	0.197460
CURVATURE/SLOPE2: 3.80023	2	3	9.8	0.408233	0.329984
	3	2	17.6	0.520751	0.424288

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:14:17

ASSAY: TBIL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.4C

KIT LOT: 0459

CAL DATE/TIME: 7/12/2020 12:42:53  
REAGENT GEN/LOT: 14406459

SUPPLY HUMIDITY: 33

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -4.12256			mg/dL		
SLOPE: 34.8456	1	1	0.9	0.180457	0.199903
CURVATURE/SLOPE2: 1.43135	2	3	10.4	0.406751	0.325850
	3	2	17.3	0.504729	0.407470

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:14:23

ASSAY: Bu  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.0C

KIT LOT: 0459

CAL DATE/TIME: 5/12/2020 17:01:58  
REAGENT GEN/LOT: 02465743

SUPPLY HUMIDITY: 31

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: 4.19194			mg/dL		
SLOPE: 295.712	1	1	0.6	0.128866	0.161449
CURVATURE/SLOPE2: -206.826	2	2	7.2	0.643274	0.550514
	3	3	10.2	0.485024	0.323911
	4	4	22.2	0.811303	0.532336

---

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:41:56

ASSAY: Bu  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.5C

KIT LOT: 0460

CAL DATE/TIME: 20/10/2021 11:56:17  
REAGENT GEN/LOT: 02079935

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: 2.86560			mg/dL		
SLOPE: 249.097	1	1	0.6	0.130505	0.164592
CURVATURE/SLOPE2: -207.053	2	2	6.8	0.641979	0.556780
	3	3	9.7	0.476381	0.314441
	4	4	21.7	0.789259	0.520370

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:21:30

ASSAY: Bc  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.5C

KIT LOT: 0460

CAL DATE/TIME: 20/10/2021 11:56:17  
 REAGENT GEN/LOT: 02079935

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -4.57232			mg/dL		
SLOPE: -135.721	1	1	0.1	0.130505	0.164592
CURVATURE/SLOPE2: 275.017	2	3	1.0	0.476381	0.314441
	3	4	8.0	0.789259	0.520370
	4	2	20.1	0.641979	0.556780

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:21:57

ASSAY: Bc  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 29.0C

KIT LOT: 0459

CAL DATE/TIME: 5/12/2020 17:01:58  
 REAGENT GEN/LOT: 02465743

SUPPLY HUMIDITY: 31

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE 1	RESPONSE 2
INTERCEPT: -6.60331	1	1	0.1	0.128866	0.161449
SLOPE: -128.008	2	3	1.9	0.485024	0.323911
CURVATURE/SLOPE2: 378.395	3	4	8.8	0.811303	0.532336
	4	2	18.9	0.643274	0.550514

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:22:40

ASSAY: TRIG  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.3C

KIT LOT: 0230

CAL DATE/TIME: 7/12/2020 12:34:38  
REAGENT GEN/LOT: 07456467

SUPPLY HUMIDITY: 33

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 9.30068			mg/dL	
SLOPE: 98.6622	1	1	35	0.267047
CURVATURE/SLOPE2: 1.92280	2	3	201	0.857672
	3	2	488	1.55527

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:20:28



ASSAY: TRIG  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.3C

KIT LOT: 0230

CAL DATE/TIME: 15/12/2021 09:53:55  
REAGENT GEN/LOT: 07049833

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -23.8010			mg/dL	
SLOPE: 109.158	1	1	27	0.361645
CURVATURE/SLOPE2: -1.55834	2	3	200	0.886041
	3	2	463	1.62546

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:20:19

ASSAY: Bc  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.4C

KIT LOT: 0460

CAL DATE/TIME: 20/10/2021 11:47:15  
REAGENT GEN/LOT: 02079935

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -3.78743			mg/dL		
SLOPE: -133.900	1	1	0.1	0.136440	0.162043
CURVATURE/SLOPE2: 254.194	2	3	1.0	0.482628	0.322557
	3	4	8.0	0.808774	0.545860
	4	2	20.1	0.654255	0.581417

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:34:40

ASSAY: TRIG  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.8C

KIT LOT: 0257

CAL DATE/TIME: 4/7/2018 15:19:28  
REAGENT GEN/LOT: 07403027

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -1.32195			mg/dL	
SLOPE: 97.1820	1	1	31	0.297328
CURVATURE/SLOPE2: 0.352217	2	3	197	0.896307
	3	2	472	1.61410

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:32:27

ASSAY: TRIG  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 27.2C

KIT LOT: 0230

CAL DATE/TIME: 14/12/2021 10:24:10  
REAGENT GEN/LOT: 07049833

SUPPLY HUMIDITY: 35

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -2.56962			mg/dL	
SLOPE: 90.1671	1	1	27	0.292350
CURVATURE/SLOPE2: 2.14201	2	3	200	0.892900
	3	2	463	1.58789

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:32:16

ASSAY: CHOL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.9C

KIT LOT: 0257

CAL DATE/TIME: 4/7/2018 15:19:09  
REAGENT GEN/LOT: 08334811

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -5.95069			mg/dL	
SLOPE: 112.935	1	1	27	0.503357
CURVATURE/SLOPE2: -3.91375	2	2	186	0.906140
	3	3	445	1.31943

---

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:09:25

ASSAY: CHOL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 26.1C

KIT LOT: 0230

CAL DATE/TIME: 24/1/2022 12:31:37  
REAGENT GEN/LOT: 08459641

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 3.07877			mg/dL	
SLOPE: 96.9900	1	1	45	0.494310
CURVATURE/SLOPE2: 0.129873	2	2	184	0.900709
	3	3	410	1.29231

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:31:16

ASSAY: Mg  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.6C

KIT LOT: 0147

CAL DATE/TIME: 3/7/2018 16:51:06  
REAGENT GEN/LOT: 32012101

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.115353			mg/dL	
SLOPE: 99.5047	1	1	0.6	0.990735
CURVATURE/SLOPE2: -10.1133	2	2	3.8	1.45299
	3	3	9.4	2.08643

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:30:17

ASSAY: Mg  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.1C

KIT LOT: 0110

CAL DATE/TIME: 1/4/2022 15:09:50  
REAGENT GEN/LOT: 32068342

SUPPLY HUMIDITY: 33

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.382835			mg/dL	
SLOPE: 108.008	1	1	0.8	0.969862
CURVATURE/SLOPE2: 99.0107	2	2	3.8	1.35466
	3	3	9.6	1.95383

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:29:11



ASSAY: dHDL  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.6C

KIT LOT: 2517

CAL DATE/TIME: 3/7/2018 16:44:18  
 REAGENT GEN/LOT: 11132158

SUPPLY HUMIDITY: 35

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.568620			mg/dL	
SLOPE: 114.639	1	1	7	0.309458
CURVATURE/SLOPE2: -7.63799	2	2	38	0.590698
	3	3	116	0.848789

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:34:22

ASSAY: dHDL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 27.1C

KIT LOT: 2540

CAL DATE/TIME: 5/1/2022 12:34:55  
REAGENT GEN/LOT: 11322283

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.511405			mg/dL	
SLOPE: 100.498	1	1	7	0.337742
CURVATURE/SLOPE2: 3.33635	2	2	34	0.591347
	3	3	108	0.836992

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:33:00

ASSAY: URIC  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.6C

KIT LOT: 0147

CAL DATE/TIME: 3/7/2018 16:50:00  
REAGENT GEN/LOT: 05264220

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.0599467			mg/dL	
SLOPE: 113.225	1	1	0.6	0.197550
CURVATURE/SLOPE2: -32.8309	2	2	5.9	0.669581
	3	3	16.8	1.33787

---

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:06:39

ASSAY: URIC  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 26.6C

KIT LOT: 0110

CAL DATE/TIME: 30/11/2021 14:20:25  
 REAGENT GEN/LOT: 05489467

SUPPLY HUMIDITY: 34

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.0466648			mg/dL	
SLOPE: 103.999	1	1	0.9	0.227342
CURVATURE/SLOPE2: 10.2910	2	2	5.8	0.697498
	3	3	16.6	1.36498

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:27:59

ASSAY: PHOS  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.5C

KIT LOT: 0147

CAL DATE/TIME: 3/7/2018 16:50:19  
REAGENT GEN/LOT: 12454254

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.651375			mg/dL	
SLOPE: 106.397	1	1	1.1	0.257896
CURVATURE/SLOPE2: -28.9960	2	2	5.3	0.484581
	3	3	13.0	0.730285

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:30:53

ASSAY: PHOS  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.3C

KIT LOT: 0110

CAL DATE/TIME: 9/12/2021 10:43:57  
REAGENT GEN/LOT: 12483461

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.597449			mg/dL	
SLOPE: 88.2428	1	1	1.2	0.212526
CURVATURE/SLOPE2: 226.099	2	2	5.6	0.466079
	3	3	12.9	0.672439

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:30:40

ASSAY: UREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.6C

KIT LOT: 0147

CAL DATE/TIME: 3/7/2018 16:49:22  
REAGENT GEN/LOT: 01272909

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.506021			mg/dL (A)	
SLOPE: 101.670	1	1	8	0.158504
CURVATURE/SLOPE2: 3.10863	2	2	106	0.972468
	3	3	245	1.80051

---

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:02:23

ASSAY: UREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 26.7C

KIT LOT: 0110

CAL DATE/TIME: 30/11/2021 14:19:47  
REAGENT GEN/LOT: 01318070

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.113747			mg/dL(A)	
SLOPE: 100.413	1	1	8	0.151459
CURVATURE/SLOPE2: 5.92787	2	2	105	0.950080
	3	3	236	1.74326

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:22:14



ASSAY: AMYL  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.5C

KIT LOT: 0357

CAL DATE/TIME: 3/7/2018 16:55:32  
 REAGENT GEN/LOT: 60419833

SUPPLY HUMIDITY: 35

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -19.9078			U/L	
SLOPE: 92.0541	1	1	36	0.009776
CURVATURE/SLOPE2: -0.112352	2	2	363	0.051177
SUB DEP DENS: 1.17700	3	3	1098	0.114661

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:05:46

ASSAY: AMYL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.5C

KIT LOT: 0320

CAL DATE/TIME: 7/10/2021 10:51:00  
REAGENT GEN/LOT: 60169168

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 9.29706			U/L	
SLOPE: 106.919	1	1	51	0.009868
CURVATURE/SLOPE2: 2.79488	2	2	392	0.047469
SUB DEP DENS: 1.18899	3	3	997	0.094086

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:25:44

ASSAY: Ca  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.1C

KIT LOT: 0147

CAL DATE/TIME: 5/7/2018 16:35:33  
REAGENT GEN/LOT: 03013731

SUPPLY HUMIDITY: 33

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.0211596			mg/dL	
SLOPE: 87.3436	1	3	2.1	0.692422
CURVATURE/SLOPE2: 0.000000000000	2	2	9.2	1.10824
	3	1	13.9	1.32929

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:29:00

ASSAY: Ca  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 26.5C

KIT LOT: 0110

CAL DATE/TIME: 30/11/2021 14:21:03  
REAGENT GEN/LOT: 03039170

SUPPLY HUMIDITY: 34

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.00562739			mg/dL	
SLOPE: 89.1140	1	3	1.7	0.698419
CURVATURE/SLOPE2: 0.000000000000	2	2	9.0	1.15114
	3	1	13.9	1.38325

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:28:35

CALIBRATION REPORT

SYSTEM NAME: 5600703

ASSAY: GLU  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.6C

KIT LOT: 0147

CAL DATE/TIME: 3/7/2018 16:49:41  
REAGENT GEN/LOT: 00254056

SUPPLY HUMIDITY: 13

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -3.48957			mg/dL	
SLOPE: 114.686	1	1	34	0.275070
CURVATURE/SLOPE2: -3.22903	2	2	302	1.12430
	3	3	595	1.63088

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:22:06

ASSAY: GLU  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.4C

KIT LOT: 0110

CAL DATE/TIME: 17/12/2021 10:24:20  
 REAGENT GEN/LOT: 00098952

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.267241			mg/dL	
SLOPE: 97.6219	1	1	33	0.270005
CURVATURE/SLOPE2: 0.424132	2	2	293	1.12527
	3	3	565	1.57205

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:21:39

ASSAY: CREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.6C

KIT LOT: 0147

CAL DATE/TIME: 3/7/2018 16:49:03  
REAGENT GEN/LOT: 15134267

SUPPLY HUMIDITY: 13

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.00127209			mg/dL	
SLOPE: 101.844	1	1	0.5	0.003100
CURVATURE/SLOPE2: 0.000000000000	2	2	1.6	0.009639
SUB DEP DENS: 0.421617	3	3	13.4	0.069430

---

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:18:50

ASSAY: CREA  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 26.8C

KIT LOT: 0110

CAL DATE/TIME: 30/11/2021 14:19:18  
REAGENT GEN/LOT: 15319728

SUPPLY HUMIDITY: 16

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.0343734			mg/dL	
SLOPE: 112.430	1	1	0.6	0.003698
CURVATURE/SLOPE2: 0.000000000000	2	2	1.5	0.009392
SUB DEP DENS: 0.425994	3	3	13.2	0.065449

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:23:20



ASSAY: ALTV  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 30.1C

KIT LOT: 0378

CAL DATE/TIME: 7/6/2019 12:18:33  
 REAGENT GEN/LOT: 56027506

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: -0.388706			U/L	
SLOPE: 101.641	1	1	4	0.006864
CURVATURE/SLOPE2: -0.463149	2	2	222	0.142138
SUBS DEPL DENS: 0.493664	3	3	762	0.408025
DELTA DENSITY: NA				
1ST POINT REF: 3.08232				

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:25:25

ASSAY: ALTV  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 27.7C

KIT LOT: 0320

CAL DATE/TIME: 30/3/2022 17:07:23  
REAGENT GEN/LOT: 56064173

SUPPLY HUMIDITY: 9

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 0.324064			U/L	
SLOPE: 99.9461	1	1	9	0.011655
CURVATURE/SLOPE2: -0.0798065	2	2	216	0.139018
SUBS DEPL DENS: 0.508329	3	3	761	0.416151
DELTA DENSITY: NA				
1ST POINT REF: 3.17388				

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:24:37

ASSAY: AST  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.5C

KIT LOT: 0357

CAL DATE/TIME: 3/7/2018 16:56:01  
 REAGENT GEN/LOT: 73234490

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 1.43616			U/L	
SLOPE: 102.976	1	1	10	0.004214
CURVATURE/SLOPE2: -0.235159	2	2	237	0.083587
SUBS DEPL DENS: 0.555034	3	3	764	0.270352
DELTA DENSITY: NA				
1ST POINT REF: 7.82830				

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:12:46

ASSAY: AST  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.7C

KIT LOT: 0320

CAL DATE/TIME: 9/3/2022 11:11:58  
REAGENT GEN/LOT: 73285153

SUPPLY HUMIDITY: 13

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE
INTERCEPT: 1.54187			U/L	
SLOPE: 96.7002	1	1	15	0.005061
CURVATURE/SLOPE2: 2.09016	2	2	208	0.072773
SUBS DEPL DENS: 0.489897	3	3	690	0.217001
DELTA DENSITY: NA				
1ST POINT REF: 7.03519				

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:12:27

ASSAY: TBIL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 28.9C

KIT LOT: 0497

CAL DATE/TIME: 4/7/2018 15:14:54  
REAGENT GEN/LOT: 14264583

SUPPLY HUMIDITY: 33

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -3.83972			mg/dL		
SLOPE: 39.3931	1	1	1.2	0.184129	0.201493
CURVATURE/SLOPE2: -2.04459	2	3	10.4	0.395317	0.317044
	3	2	18.4	0.505297	0.397805

---

End of Report

PAGE 1

PRINT DATE/TIME: 5/4/2022 11:03:27

: ASSAY: TBIL  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 29.9C

KIT LOT: 0460

CAL DATE/TIME: 10/1/2022 12:34:57  
REAGENT GEN/LOT: 14483001

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -3.43022			mg/dL		
SLOPE: 30.3032	1	1	1.1	0.190473	0.199040
CURVATURE/SLOPE2: 6.71292	2	3	9.8	0.408122	0.328641
	3	2	17.6	0.516902	0.418307

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:23:36

ASSAY: Bu  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 25.0C

KIT LOT: 0487

CAL DATE/TIME: 15/5/2018 14:45:10  
REAGENT GEN/LOT: 02221932

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: 1.53569			mg/dL		
SLOPE: 289.314	1	1	0.5	0.123278	0.152941
CURVATURE/SLOPE2: -217.394	2	2	6.6	0.643949	0.571400
	3	3	9.8	0.460618	0.304305
	4	4	21.1	0.766061	0.522103

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:35:13

ASSAY: Bu  
 FLUID: Serum  
 Status: Passed  
 OP ID:  
 SITE TEMP: 28.4C

KIT LOT: 0460

CAL DATE/TIME: 20/10/2021 11:47:15  
 REAGENT GEN/LOT: 02079935

SUPPLY HUMIDITY: 35

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE mg/dL	RESPONSE 1	RESPONSE 2
INTERCEPT: 2.41299					
SLOPE: 237.587	1	1	0.6	0.136440	0.162043
CURVATURE/SLOPE2: -186.980	2	2	6.8	0.654255	0.581417
	3	3	9.7	0.482628	0.322557
	4	4	21.7	0.808774	0.545860

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:35:01



ASSAY: Bc  
FLUID: Serum  
Status: Passed  
OP ID:  
SITE TEMP: 25.0C

KIT LOT: 0487

CAL DATE/TIME: 15/5/2018 14:45:10  
REAGENT GEN/LOT: 02221932

SUPPLY HUMIDITY: 35

---

PARAMETERS	LEVEL	BOTTLE	CALIB VALUE	RESPONSE 1	RESPONSE 2
INTERCEPT: -6.12241			mg/dL		
SLOPE: -131.589	1	1	0.2	0.123278	0.152941
CURVATURE/SLOPE2: 372.428	2	3	0.8	0.460618	0.304305
	3	4	9.1	0.766061	0.522103
	4	2	20.6	0.643949	0.571400

---

End of Report

PAGE 1

PRINT DATE/TIME: 4/4/2022 11:34:53