

ASIAN TECHNOLOGY



(A HOUSE OF CALIBRATION)

CALIBRATION CERTIFICATE

| Customer name | M/S: KVR Hospital Kashipur And Near Reliance Petrol Pump Moradabad Road, Kashipur-244713 Distt- US Nagar (Uttrakhand) | ULR No. | CC223922000001234 F |
|---------------|--|---------------------------|---------------------|
| | | Service request No & date | P/05/12/03/2022 |
| Address | | Date of calibration | 14/03/2022 |
| | | Valid up to | 13/03/2023 |
| | | Date of issue | 15/03/2022 |

| | Instrument Deta | nils | |
|-----------------|------------------------------------|-------------------|--------|
| Instrument name | Digital Thermometer (Refrigerator) | S. No. | TL8009 |
| Make | Aptechdeals | Location | Lab |
| Range / Size | -50 to 110 °C | I.D. No. | DTM-01 |
| Least Count | 0.1 °C | Visual Inspection | OK |

| | Detail of reference standards & Major ed | quipments used |
|----------------------|--|----------------|
| Equipment Name | RTD With Digital Indicator | |
| Make | YCT / SBI | |
| Model / SR No. | 11080368 | |
| Certificate No. | CC325722000000135F | |
| Calibration Validity | 10/02/2023 | |
| Calibration by | KL Technology | |

| Environmental | Temperature | 25±4°C | Calibration Reference | IS:6274 |
|---------------|-------------------|---------|-----------------------|---------|
| Condition | Relative Humidity | 50±20 % | Calibration Procedure | CP-02T |

Calibration Results (Up Ward)

| Serial No. | Std. Reading in (°C) | Measured Reading in | Uncertainty At 95% C.L. (coverage factor K=2) |
|------------|----------------------|---------------------|--|
| 01. | -18.31 | -18.5 | (straige metor it 2) |
| 02. | -14.21 | -14.1 | |
| 03. | -10.15 | -9.9 | ±0.3 °C |
| 04. | -8.16 | -8.0 | -0.5 |
| 05. | 0.04 | 0.2 | |

| Serial No. | Std. Reading in | Measured Reading in | Uncertainty At 95% C.L. (coverage factor K=2) |
|------------|-----------------|---------------------|--|
| 01. | 2.15 | 2.1 | (coverage metor ix 2) |
| 02. | 4.23 | 4.0 | |
| 03. | 6.38 | 6.2 | ±0.3 °C |
| 04. | 8.21 | 8.3 | Trove by |

Remarks

(1) Standard equipment use for calibration are traceable to national/ international standards.

Calibrated By (Calibration Engg./TM) (D.B SAXSENA) Form No.QF-47



Approved By
(QM/TM)
(NEERA) TYAGI)
Page No. 1 of 1

⁽²⁾The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.