



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: KVR Hospital Kashipur Near Reliance Petrol Pump Moradabad Road, Kashipur-244713 Distt- US Nagar (Uttarakhand)	ULR No.	CC223922000001232 F
		Service request No & date	P/05/12/03/2022
		Date of calibration	14/03/2022
		Valid up to	13/03/2023
		Date of issue	15/03/2022

### Instrument Details

Instrument name	Centrifuge	S. No.	ZFIN-31004
Make / Model No.	Remi / C-854/6	Location	Lab
Range / Size	3500 rpm	I.D. No.	CF-01
Least Count	N/A	Visual Inspection	OK

### Detail of reference standards & Major equipments used

Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/AC92792		
Certificate No.	QCTS/000229/10		
Calibration Validity	21/10/2022		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2 °C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37


### Calibration Results

Serial No.	Position	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	Position: 1	1225	±0.2 %rdg.
02.	Position: 2	1554	
03.	Position: 3	1856	
04.	Position: 4	2525	
05.	Position: 5	3458	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By  
(Calibration Engg./TM)  
(MAHIPAL)  
Form No.-QF-47



Approved By  
(QM/TM)  
(NEERAJ TYAGI)  
Page No. 1 of 1

---End of Report---