



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: KVR Hospital Kashipur Near Reliance Petrol Pump Moradabad Road, Kashipur-244713 Distt- US Nagar (Uttarakhand)	ULR No.	CC223922000001233 F
		Service request No & date	P/05/12/03/2022
		Date of calibration	14/03/2022
		Valid up to	13/03/2023
		Date of issue	15/03/2022

Instrument Details			
Instrument name	Centrifuge	S. No.	ZGGN-13407(
Make / Model No.	Remi / C-854/6(Location	Lab
Range / Size	3500 rpm	I.D. No.	CF-02
Least Count	N/A	Visual Inspection	OK

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/AC92792		
Certificate No.	QCTS/000229/10		
Calibration Validity	21/10/2022		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %		Calibration Procedure

Serial No.	Position	Calibration Results	
		Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	Position: 1	1210	±0.2 %rdg.
02.	Position: 2	1553	
03.	Position: 3	1855	
04.	Position: 4	2523	
05.	Position: 5	3459	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(MANIPAL)
Form No.-QF-47



Approved By
(QM/TM)
(NEERAJTYAGI)
Page No. 1 of 1

---End of Report---