



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100005048	Issued On	22/03/2022
ULR No.	CC214622100005048F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5048	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	Instrumentics / HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	THM-01
Range	Up to 60°C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

#### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS	Environmental Chamber
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration (On Working Range)

U.U.C. Reading		Corrected Value of Standard (Average)		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
15.0 °C	20 % RH	14.94 °C	19.64 % RH	± 0.37 °C	± 1.6 % RH
20.0 °C	40 % RH	19.91 °C	39.33 % RH	± 0.37 °C	± 1.6 % RH
30.0 °C	75 % RH	29.89 °C	73.94 % RH	± 0.43 °C	± 1.6 % RH
40.0 °C	90 % RH	39.85 °C	88.42 % RH	± 0.43 °C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note :
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

Authorised Signatory  
Quality Manager





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100005050	Issued On	22/03/2022
ULR No.	CC214622100005050F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5050	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	Instrumentics / HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	THM-03
Range	Up to 60°C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

#### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS	Environmental Chamber
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

#### RESULTS: Thermal Calibration (On Working Range)

U.U.C. Reading		Corrected Value of Standard (Average)		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
15.0 °C	20 % RH	15.04 °C	20.43 %RH	± 0.37 °C	± 1.6 % RH
20.0 °C	40 % RH	20.07 °C	40.67 %RH	± 0.37 °C	± 1.6 % RH
30.0 °C	75 % RH	30.09 °C	76.08 %RH	± 0.43 °C	± 1.6 % RH
40.0 °C	90 % RH	40.14 °C	91.59 %RH	± 0.43 °C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARTIK KAPOOR  
Authorised Signatory





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100005049	Issued On	22/03/2022
ULR No.	CC214622100005049F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5049	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd.
	Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)

### EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	Instrumentics / HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	THM-02
Range	Up to 60°C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/CIF/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

### USED EQUIPMENT DETAILS

Environmental Chamber

### PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

### RESULTS: Thermal Calibration (On Working Range)

U.U.C. Reading		Corrected Value of Standard (Average)		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
15.0 °C	20 % RH	15.06 °C	19.57 % RH	± 0.37 °C	± 1.6 % RH
20.0 °C	40 % RH	20.08 °C	39.21 % RH	± 0.37 °C	± 1.6 % RH
30.0 °C	75 % RH	30.11 °C	73.82 % RH	± 0.43 °C	± 1.6 % RH
40.0 °C	90 % RH	40.16 °C	88.37 % RH	± 0.43 °C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARSHI KAPOOR  
Authorised Signatory





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Calibration Certificate No.	22100005046	Issued On	22/03/2022
ULR No.	CC214622100005046F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5046	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

### EQUIPMENT DESCRIPTION

Name	Glass Thermometer for Refrigerator		
Make/Model	Zeal	Visual Inspection	Ok
Serial No.	---	I.D. No.	GEN/005
Range	-10°C to 110°C	Least Count	1°C
Location	---	Calibration Site	In-Lab

### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Methanol Bath
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90 (Comparison Method)

U.U.C. READING (in °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
2	1.97	± 0.61
4	3.95	± 0.61
5	4.92	± 0.61
7	6.89	± 0.61
8	7.86	± 0.61

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\* END OF REPORT \*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked By

Authorised Signatory





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Calibration Certificate No.	22100005047	Issued On	22/03/2022
ULR No.	CC214622100005047F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5047	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

#### CALIBRATED FOR:

M/s Path Kind Diagnostics Pvt. Ltd.  
Ground Floor, Dev Sagar Lodge, Kalaam Chouraha,  
DM Road, Bulandsahar-203001, (U.P.)

#### EQUIPMENT DESCRIPTION

Name	Glass Thermometer for Refrigerator		
Make/Model	Zeal	Visual Inspection	Ok
Serial No.	---	I.D. No.	GEN/006
Range	-10°C to 110°C	Least Count	1°C
Location	---	Calibration Site	In-Lab

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

#### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

#### USED EQUIPMENT DETAILS

Methanol Bath

#### PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90  
(Comparison Method)

#### RESULTS: Thermal Calibration

U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
2	2.04	± 0.61
4	4.07	± 0.61
5	5.08	± 0.61
7	7.11	± 0.61
8	8.13	± 0.61

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\* END OF REPORT \*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

Authorised Signatory





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF MICROPIPETTE

<b>Certificate No:</b>	22700004690	<b>Issued On:</b>	25/03/2022
<b>ULR No:</b>	CC214622700004690F	<b>Calibration Date:</b>	22/03/2022
<b>Job Identification No:</b>	ICS/C/MVD/03/4690	<b>Next Calibration Date:</b>	21/03/2023
<b>Ref. No:</b>	SRF- Dated: 19/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

<b>Name</b>	Variable Micropipette		
<b>Make/Model</b>	Eppendorf	<b>Visual Inspection</b>	Ok
<b>Serial No.</b>	I23274G	<b>I.D. No.</b>	---
<b>Range</b>	20 µl - 200 µl	<b>Least Count</b>	0.2 µl
<b>Location</b>	---	<b>Calibration Site</b>	In-Lab

#### ENVIRONMENTAL CONDITIONS

<b>Temperature</b>	25.0 °C ± 3.0° C	<b>Humidity</b>	50 ± 10 % RH
--------------------	------------------	-----------------	--------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Calibrated Balance, Distilled Water & Glass wares
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655=6

**RESULTS: Mechanical Calibration**  
Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.0200 (020.0 µl)	0.01992 (019.92 µl)	± 0.1 µl
0.1000 (100.0 µl)	0.09957 (099.57 µl)	± 0.1 µl
0.2000 (200.0 µl)	0.19934 (199.34 µl)	± 0.1 µl

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.  
\*\*\*END OF REPORT\*\*\*

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARTIK KAPOOR  
Authorised Signatory





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF MICROPIPETTE

Certificate No:	22700004691	Issued On:	25/03/2022
ULR No:	CC214622700004691F	Calibration Date:	22/03/2022
Job Identification No:	ICS/C/MVD/03/4691	Next Calibration Date:	21/03/2023
Ref. No:	SRF- Dated: 19/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

### EQUIPMENT DESCRIPTION

Name	Variable Micropipette		
Make/Model	Eppendorf	Visual Inspection	Ok
Serial No.	H47294G	I.D. No.	---
Range	100 µl - 1000 µl	Least Count	1 µl
Location	---	Calibration Site	In-Lab

### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Calibrated Balance, Distilled Water & Glass wares
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6


**RESULTS: Mechanical Calibration**  
Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.100 (0100 µl)	0.09972 (0099.72 µl)	± 0.6 µl
0.500 (0500 µl)	0.49663 (0496.63 µl)	± 0.6 µl
1.000 (1000 µl)	0.99814 (0998.14 µl)	± 0.6 µl

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.  
\*\*\*END OF REPORT\*\*\*

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by   
Authorised Signatory 



# Indian Calibration Services



An Exclusive House for Validation & Calibration of Analytical Instruments  
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900001580	Issued On:	22/03/2022
ULR No:	CC214622900001580F	Calibration Date:	19/03/2022
Job Identification No:	ICS/C/AS/03/1580	Next Calibration Date:	18/03/2023
Ref. No:	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

Name	Laboratory Centrifuge, Fitted with Digital RPM Indicator		
Make/Model	Remi / R-8C Plus	Visual Inspection	Ok
Serial No.	ZDGN-18557	I.D. No.	ACC/001
Range	Up to 6000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS: Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

#### RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

CENTRIFUGE READING	CORRECTED VALUE OF TACHOMETER READING (AVERAGE)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1560	1562.5	± 6
2230	2233.7	± 6
3140	3145.2	± 6
4350	4357.1	± 7
5710	5720.8	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
3. Total liability of our Organisation is limited to the invoiced amount.  
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARISHMA KAPUR  
Quality Manager





# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

<b>Certificate No:</b>	22900001581	<b>Issued On:</b>	22/03/2022
<b>ULR No:</b>	CC214622900001581F	<b>Calibration Date:</b>	19/03/2022
<b>Job Identification No:</b>	ICS/C/AS/03/1581	<b>Next Calibration Date:</b>	18/03/2023
<b>Ref. No:</b>	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

<b>Name</b>	Laboratory Centrifuge, Fitted with Digital RPM Indicator		
<b>Make/Model</b>	Remi / R-8C Plus	<b>Visual Inspection</b>	Ok
<b>Serial No.</b>	ZDGN-18563	<b>I.D. No.</b>	GEN/001
<b>Range</b>	Up to 6000 RPM	<b>Least Count</b>	10 RPM
<b>Location</b>	---	<b>Calibration Site</b>	On-Site

#### ENVIRONMENTAL CONDITIONS

<b>Temperature</b>	25.0 °C ± 2.0° C	<b>Humidity</b>	45 to 75 % RH
--------------------	------------------	-----------------	---------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

**USED EQUIPMENT DETAILS** Not Applicable

**PRINCIPLE/METHODOLOGY OF CALIBRATION:** As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

#### RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

CENTRIFUGE READING	CORRECTED VALUE OF TACHOMETER READING (AVERAGE)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1290	1287.5	± 6
2330	2325.4	± 6
3510	3503.5	± 6
4670	4661.7	± 7
5820	5808.3	± 8

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*


- Note:**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the WUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARTIK KANDOR  
Authorised Signatory  
Quality Manager





**Twin Cooling Plus**

-23°C	Freezer Convert	1°C
-21°C	Fridge	2°C
-19°C	off	3°C
-17°C	❄️	5°C
-15°C	💧	7°C

Freezer Power Freeze (3 sec) | Freezer Convert (3 sec) | Fridge Power Cool (3 sec)

INV  
TECH  
**10 YEAR - ON THE  
WARRANTY**

**MORE STARS  
MORE SAVINGS**

**POWER SAVING  
GUIDE**


**ELECTRICITY CONSUMPTION**  
**224**  
**UNITS PER YEAR**  
Label Period: 1st Jan 2018 - 31st Dec 2019

Appliance	Refrigerator
Brand	SAMSUNG
Model / Year	RT6M551E3L7L / 2018
Type	Frost Free
Gross Volume	415 Liters
Storage Volume	384 Liters

**ENERGY STAR**

\*Under test conditions, under normal circumstances and normal operation. Actual electricity consumption will depend on how the appliance is used.



**CONTROLLED COPY**  
 AUTH:   
**QA FUNCTION**

	PATHKIND DIAGNOSTICS PVT LTD	ISSUE DATE: 01-08-2017	AMENDMENT DATE:	KOTHIYAT, GROUND FLOOR, KALA AAM CHORAHA, DM ROAD, BULANDSHEHAR, UTTAR PRADESH-201001
DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER		ISSUE NO: 01	AMENDMENT NO:	

YEAR: 2022

SR NO:

DATE		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
JAN	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER	ⓐ																														
	SIGNATURE	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	
FEB	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER																															
	SIGNATURE	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	
MAR	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER																															
	SIGNATURE	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	
APR	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER																															
	SIGNATURE	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	ⓐ	
MAY	REFRIGERATOR																															
	FREEZER																															
	SIGNATURE																															
JUN	REFRIGERATOR																															
	FREEZER																															
	SIGNATURE																															

ACCEPTABLE LIMITS  
 REFRIGERATOR: 2-8° C      FREEZER: -10 to -30° C


SECTION HEAD SIGN:

DOC NO.: P005/GEN/007.0	DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER	Page 01 of 02
ISSUED BY: DR ASHOK RATTAN		

Scanned with CamScanner



**CONTROLLED COPY**

AUTH: 

**QA FUNCTION**



PATHKIND DIAGNOSTICS PVT LTD

ISSUE DATE:

01-08-2017

AMENDMENT DATE:

AMENDMENT NO:

KOTHYAT, GROUND FLOOR,  
KALA AAM CHORANA, DM  
ROAD, BULANOSHEHAR,  
UTTAR PRADESH-203001

DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER

ISSUE NO:  
01

YEAR: 2022

SR NO:

DATE		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
JAN	REFRIGERATOR	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	
	FREEZER	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>		
	SIGNATURE																															
FEB	REFRIGERATOR	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>		
	FREEZER	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>		
	SIGNATURE																															
MAR	REFRIGERATOR	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>		
	FREEZER	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>		
	SIGNATURE																															
APR	REFRIGERATOR	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>	19 <sup>o</sup>		
	FREEZER	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>	-5 <sup>o</sup>		
	SIGNATURE																															
MAY	REFRIGERATOR																															
	FREEZER																															
	SIGNATURE																															
JUN	REFRIGERATOR																															
	FREEZER																															
	SIGNATURE																															

ACCEPTABLE LIMITS

REFRIGERATOR: 2-8° C


FREEZER: -10 to -30° C


SECTION HEAD SIGN:

DOC NO.: P005/GEN/007.D	DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER	Page 01 of 02
ISSUED BY: DR ASHOK RATTAN		

Scanned with CamScanner



AUTH:   
QA FUNCTION

	PATHKIND DIAGNOSTICS PVT LTD	ISSUE DATE: 01-08-2017	AMENDEMENT DATE:	KOTHIYAT, GROUND FLOOR KALA AAM CHORAHIA, D1 ROAD, BULANDSHIEHAR UTTAR PRADESH-203001
	DAILY ROOM TEMPERATURE & HUMIDITY MONITORING	ISSUE NO: 01	AMENDEMENT NO:	

YEAR: 2022

SR NO:

DATE		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
JAN	TEMPERATURE	20.5	21.5	21.5	21.0	21.8	22.2	22.0	22.5	21.0	22.0	21.0	23.0	21.0	22.0	20.7	20.7	20.0	21.0	22.5	22.3	22.6	21.9	22.1	20.9	21.5	21.1	21.5	21.0	21.7	22.0	20.9	
	HUMIDITY %	49%	52%	52%	50%	67%	58%	65%	67%	57%	55%	54%	51%	51%	56%	53%	53%	52%	59%	62%	60%	62%	57%	59%	57%	61%	54%	58%	60%	55%	55%	55%	
	SIGNATURE	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	
FEB	TEMPERATURE	21.5	20.6	21.1	20.4	21.8	21.0	21.0	20.7	21.2	21.4	21.4	21.1	20.9	22.0	21.9	22.1	22.1	22.1	22.1	23.1	22.0	21.0	22.0	21.0	22.0	22.0	21.0	21.0	21.0	21.0		
	HUMIDITY %	55%	52%	55%	53%	57%	55%	57%	62%	59%	56%	57%	56%	55%	54%	55%	54%	55%	55%	56%	55%	54%	52%	51%	52%	56%	57%	57%	58%	58%	58%		
	SIGNATURE	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	
MAR	TEMPERATURE	23.0	22.3	23.1	24.3	21.9	23.1	24.0	24.5	23.0	24.0	23.0	24.0	25.0	25.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0	26.0		
	HUMIDITY %	58%	57%	58%	56%	51%	49%	56%	57%	51%	49%	51%	55%	50%	48%	49%	51%	50%	47%	56%	51%	51%	50%	51%	50%	51%	46%	56%	47%	55%	49%	50%	
	SIGNATURE	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	
APR	TEMPERATURE	22.0	22.1	22.3	22.7	22.4	22.0	22.7	22.0	22.1	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0	22.0		
	HUMIDITY %	36%	38%	38%	38%	35%	34%	32%	35%	31%	35%	34%	45%	44%	51%																		
	SIGNATURE	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	AS	
MAY	TEMPERATURE																																
	HUMIDITY %																																
	SIGNATURE																																
JUN	TEMPERATURE																																
	HUMIDITY %																																
	SIGNATURE																																

ACCEPTABLE LIMITS  
ROOM TEMPERATURE: 21-30° C      HUMIDITY: 30 % to 80 %

SECTION HEAD SIGN:

DOC NO.: P005/GEN/006.0	DAILY ROOM TEMPERATURE & HUMIDITY MONITORING	Page 01 of 02
ISSUED BY: DR ASHOK RATTAN		







