

CALIBRATION REPORT

Details of Device Under Calibration :

Date:05-03-2022

Customer Name & Address: Kasturba Hospital,Sewa Rural,Jhaghadia	Certificate ID : AZ/21-22/ 1431
Name of Equipment: Refrigerator	Calibration Date : :22-02-2022
Manufacturer : WHIRLPOOL	Calibration Due Date : :21-02-2023
Model : FF2D35545/2021	Location of equipment : :LAB
Serial No : NS	Ambient Temperature : 22 °C

Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali. date	Cali. due date
1	Temp. Detector with sensor	Omron	DS708	19090715	01-02-2022	31-01-2023

Condition Of Equipment at the time of Calibration : GOOD

Test Report:

Sr No	Test	Unit	Range	Reading at	Observed Value by Test equipment	Result
1	Temperature	Deg.° C	2°C	Rack 1	4.6	Pass
2			to	Rack 2	4.6	Pass
3			8 °C	Container	4.7	Pass
4			N.S.	Freezer	-15.1	Pass

Note:

- (1) Results reported are valid at the time of and under the stated conditions of measurement.
- (2) Calibration has been done as per customer's requirement.
- (3) Measurements are traceable to the National Standard
- (4) NS means Not Specified and DUC means Device Under Calibration
- (5) The reported measurement uncertainty is estimated at a level of confidence of approx 95 % with a coverage factor $k = 2$.
- (6) Cali means calibration .

Calibrated By: 
Gaurang Jansari

Approved By:


Gaurav Desai

THIS DOCUMENT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL OF "A TO Z BIOMEDICAL SERVICES"

CALIBRATION REPORT

Details of Device Under Calibration :

Date:05-03-2022

Customer Name & Address: Kasturba Hospital,Sewa Rural,Jhaghadia		Certificate ID : AZ/21-22/ 1432
Name of Equipment: Refrigerator		Calibration Date : :22-02-2022
Manufacturer : SAMSUNG		Calibration Due Date : :21-02-2023
Model : RT-39MSS3858/TL/2019		Location of equipment : :LAB
Serial No : NS		Ambient Temperature : 22 °C

Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali. date	Cali. due date
1	Temp. Detector with sensor	Omron	DS708	19090715	01-02-2022	31-01-2023



Condition Of Equipment at the time of Calibration : GOOD

Test Report:

Sr No	Test	Unit	Range	Reading at	Observed Value by Test equipment	Result
1	Temperature	Deg.° C	2°C	Rack 1	4.5	Pass
2			to	Rack 2	4.4	Pass
3			8 °C	Container	4.6	Pass
4			N.S.	Freezer	-15.1	Pass

Note:

- Results reported are valid at the time of and under the stated conditions of measurement.
- Calibration has been done as per customer's requirement.
- Measurements are traceable to the National Standard
- NS means Not Specified and DUC means Device Under Calibration
- The reported measurement uncertainty is estimated at a level of confidence of approx 95 % with a coverage factor $k = 2$.
- Cali means calibration .

Calibrated By:  Gaurang Jansari	Approved By:  Gaurav Desai
---	---