



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2021



PC-1024

Lab Name **GREEN CROSS DIAGNOSTIC CENTER** Lab No **5861**
 Constituent Group **Chemistry I** Date of Result Entered : **15/12/2021**
 PT item **Lyophilized Serum** Date of Report Published : **03/01/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Mindray	196	191.38	4.85	9.28	163.8 mg/dL	-2.97	1.33
2	UREA	Urease UV / GLDH	Mindray	191	87.91	6.88	6.05	77.8 mg/dL	-1.67	0.88
3	CREATININE	Enzymatic Colorimetric	Mindray	57	7.31	8.14	0.60	6.53 mg/dL	-1.31	0.16
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Mindray	165	2.36	10.94	0.26	2.5 mg/dL	0.54	0.04
5	T-PROTEIN	Biuret - Colorimetric	Mindray	197	5.21	6.26	0.33	5.26 g/dL	0.15	0.05
6	ALBUMIN	BCG - colorimetric	Mindray	193	3.29	8.22	0.27	3.1 g/dL	-0.70	0.04
7	CALCIUM	Arsenazo III	Mindray	141	10.60	6.60	0.70	11.4 mg/dL	1.14	0.12
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Mindray	103	4.67	10.28	0.48	4.4 mg/dL	-0.56	0.09
9	URIC ACID	Enzymatic / Uricase Colorimetric	Mindray	194	8.69	8.08	0.70	6.7 mg/dL	-2.83	0.10
10	CHOLESTEROL	CHOD-PAP	Mindray	203	108.30	5.93	6.42	103.8 mg/dL	-0.70	0.90
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	181	209.18	5.82	12.17	193.6 mg/dL	-1.28	1.81
12	HDL CHO	Direct method / Enzymatic colorimetric	Mindray	144	24.96	10.71	2.67	27.1 mg/dL	0.80	0.45
13	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	576	141.61	2.90	4.10	249.1 mmol/L	26.22	0.34
14	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	579	3.32	5.21	0.17	3.27 mmol/L	-0.29	0.01
15	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	495	107.44	4.39	4.71	199 mmol/L	19.42	0.42
16	AST	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	629	89.46	12.82	11.47	98.9 U/L	0.82	0.91
17	ALT	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	621	54.20	19.44	10.53	56.8 U/L	0.25	0.85
18	ALP	PNP AMP kinetic	Mindray	123	236.91	14.88	35.26	118.6 U/L	-3.36	6.36
19	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	411	55.93	18.85	10.54	91.37 U/L	3.36	1.04
20	IRON	Ferrene (No Protein Removal)	Any Analyser (Automation / Semi Automation)	58	85.26	10.78	9.19	72.9 ug/dL	-1.35	2.41
21	MAGNESIUM	Xylidyl Blue	Mindray	22	2.60	12.07	0.31	1.49 mg/dL	-3.58	0.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.

Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

GREEN CROSS DIAGNOSTIC CENTER

GREEN CROSS DIAGNOSTIC CENTER SOLA ROAD 201, Pushpraj Arcade, Above Axis bank, Nr. Bhuyangdev Cross Road, Sola Road, Ghatlodiya

AHMEDABAD

GUJARAT380061

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****