

Proficiency Testing Scheme for HIV Serology

Ref No.:

Date: 11/03/22

PROFICIENCY TESTING REPORT (SRL to ICTCs/Blood banks)

Round: I/II 2021-22

Name of Proficiency Testing Provider: NRL, NARI-Pune

Date Panel Distributed: 01/03/22

Name of the participating laboratory: ICTC-1, SDH, Ponda Date results received: 04/03/22
EA-14

Sample Code	Standard Result	Results obtained by participating laboratory	% Matching
A	Non Reactive	Non Reactive	100%
B	Non Reactive	Non Reactive	100%
C	Reactive	Reactive	100%
D	Non Reactive	Non Reactive	100%

Keyword: R: Reactive, NR: Non reactive

Remarks: Performance is Satisfactory / Unsatisfactory

Name of Laboratory Incharge: Dr. M J W Pinto

Signature of Laboratory Incharge

Date: 11/03/22

Dr. M J W Pinto

Associate Professor

Department of Microbiology

Goa Medical College

Bambolim-Goa

***End of Report ***

CONFIDENTIAL

Page 1 of 1