



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2022



PC-1024

Lab Name **SHRADHHADEEP GREENCROSS PATHOLOGY** Lab No **5524**  
 Constituent Group **HbA1c** Date of Result Entered : **09/03/2022**  
 PT item **Lyophilized Whole blood** Date of Report Published : **06/04/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	121	7.19	4.55	0.33	7.1 %	-0.28	0.06

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

SHRADHHADEEP GREENCROSS PATHOLOGY  
LABORATORY ,PLOT No.910/1,FIRST FLOOR, NEAR GH-3 CIRCLE,SECTOR-7,GANDHINAGAR  
GANDHINAGAR  
GUJARAT382002

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*