

**State Reference Laboratory
I. C. T. C.
Department of Microbiology
T N M C & B Y L Nair Hospital Mumbai-08**

Certificate of calibration

Name of the DRI: **BOBA Hospital**

Date of calibration: **6/3/2022**

Name of the Equipment: **Thermometer**

Next Calibration due on: **5/3/2023**

Serial No. / Model No.: **—**

Identification No.: **BDBA/ICTC/Thermo/05**

Acceptable range: **2 to 8 degree Celsius (°C)**

Point of Calibration: **2°C, 4°C, 8°C**

Tolerance (If Applicable): **± 2°C**

Location of Calibration: **At Site**

Condition of Receipt Item: **Good**

Method of calibration: **In- house calibration done using Calibrated Thermometer 2018-19**

Traceability Certificate no: **IDEMI, NABL C-0085/CC/ECL/0193/17-18**

Calibration Result:

Calibration Points (°C)	UUC Reading (°C)			Standard Reading (°C)	Mean
	Reading 1	Reading 2	Reading 3		
2	2°C	2°C	2°C	2°C	2°C
4	4°C	4°C	4°C	4°C	4°C
8	8°C	8°C	8°C	8°C	8°C

Acceptable Variation **± 2°C** of the Mean Value

UUC: Unit under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

RNIBNB
6/3/22
Usha Rai
Technical Officer, MDACS