

**State Reference Laboratory  
I. C. T. C.  
Department of Microbiology  
T N M C & B Y L Nair Hospital Mumbai-08**

**Certificate of calibration**

Name of the DRL.: **BDBA Hospital**

Date of calibration: **6/3/2022**

Name of the Equipment: **Timer**

Next Calibration due on: **5/3/2023**

Serial No. / Model No.: **-**

Identification No.: **BDBA / ICTC / TM / 013**

Acceptable range: **0 - 60 Mins**

Point of Calibration: **5 mins, 10 mins, 20 mins**

Tolerance (If Applicable): **± 2%**

Location of Calibration: **At Site**

Condition of Receipt Item: **Good**

Method of calibration: **In- house calibration done using Calibrated Digital stop watch 2018-19**

Traceability Certificate no: **IDEMI, NABL C-0085/CC/ECL/0193/17-18**

**Calibration Result:**

| Calibration Points (Minutes) | UUC Reading (Minutes) |           |           | Standard Reading (Minutes) | Mean  |
|------------------------------|-----------------------|-----------|-----------|----------------------------|-------|
|                              | Reading 1             | Reading 2 | Reading 3 |                            |       |
| 5                            | 5.00                  | 5.00      | 5.00      | 5.00                       | 5.00  |
| 10                           | 10.00                 | 10.00     | 10.00     | 10.00                      | 10.00 |
| 20                           | 20.00                 | 20.00     | 20.00     | 20.00                      | 20.00 |

Acceptable Variation **± 2% of the Mean Value**

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

**RUBHA**  
**6/3/22**  
**Vubha Rai**  
Technical Officer, MDACS