



UNIQUE CALIBRATION SOLUTIONS LLP


An ISO/IEC-17025:2017 / ACCREDITED BY NABL

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in



CC-3102

CALIBRATION CERTIFICATE					
UNIQUE LAB REPORT NO(ULR NO)		: CC310222000000861F			
SRF NO	: 075	CERTIFICATE No.	UCSL/03-22/075_04		
CUSTOMER DETAILS:		M/s. ICTC THALUK HEADQUARTERS HOSPITAL THRIPUNITHARA			
Device under calibration	TEMPERATURE INDICATOR OF REFRIGERATOR		Equipment & Master Used For calibration		
Make / Model	GODEREJ/GDA19B2007		01. 4 WIRE RTD SENSOR WITH INDICATOR		
Range	2 to 8 °C		Make : KELVIN / YUDIAN		
Least Count	0.1 °C		Sr. No. : 20021703 / 20L737937		
Instrument Id. No.	ICTC/THQHTPRA/04		Certificate No. : C-210713-1-5		
Instrument Sr. No.	-		Next Due Date : 13.07.2022		
Visual Inspection	OK		Calibrated By : GODREJ & BOYCE		
Location	ICTC KASP COUNTER		Accreditation No : CC-2559		
Material Receipt Date	-----				
Date of Calibration	22.03.2022				
Next Calibration Due (as per customer request)	22.03.2023				
Certificate Issue Date	29.03.2022				
Calibration performed At	ON SITE				
Discipline	THERMAL				
Method used	COMPARISON METHOD				
ENVIRONMENTAL CONDITION					
Temperature	30.7 °C				
Relative Humidity	61 %RH				
OBSERVATION RESULTS					
SLNO	SET VALUE	DUC READING	MASTER READING	ERROR	Measurement Uncertainty In ±
	(°C)		(°C)	(°C)	
01	2	2.04	2.08	-0.04	0.60
02	4	3.98	4.06	-0.08	0.60
03	5	4.90	4.97	-0.07	0.60
04	6	6.02	6.12	-0.10	0.60
05	8	7.94	8.07	-0.13	0.60
Calibration Procedure Based On : UCSL/SOP/01-T (ITS-90)					
Decision Rule Is Applied or Not				YES	NO
					✓
NOTE :-					
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.					
2. This report should not be reproduced except in full without our prior permission in writing.					
3. Calibration certificate without signature are not valid.					
4. All our measurements are Traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.					
5. DUC : Device under calibration.					
6. Reported readings of DUC are average of 5 measurements.					
7. Measurement Uncertainty reported is at approximately 95% confidence level with K=2.					
8. This Calibration Certificate relates only to the above DUC.					
Calibrated By		 Issued/ Approved By (APPU K MANI) Technical Director			

Laboratory : 5/D3, EK Apartments, Model engineering college road Thrikkakara, Edappally, Cochin, Kerala-682021
 Registered Office : Plot No. 611-P, Sector - 38, Gurgaon Haryana - 122001

SALES ★ SPARES ★ SERVICE ★ TESTING ★ CALIBRATION ★ MODERNIZATION