



UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL




UNIQUE
CALIBRATION & MAINTENANCE

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in



CC-3102

CALIBRATION CERTIFICATE				
UNIQUE LAB REPORT NO (ULR NO)		CC310222000000859F		
SRF NO.	075	CERTIFICATE NO	UCSL/03-22/075_02	
CUSTOMER DETAILS: : M/s. ICTC THALUK HEADQUARTERS HOSPITAL THRIPUNITHARA				
Device Under Calibration	VDRL ROTATOR (RPM)		Equipment & Master Used For calibration	
Make / Model No.	ROTEK		01. DIGITAL TACHO METER	
Range	180 rpm		Make : LUTRON	
Least Count	-		Model No. : DT -2236 C	
Instrument Id. No.	ICTC/THQHTPRA/28		Certificate No. : C-210713-1-8	
Instrument Sr. No.	1385		Next Due Date : 13.07.2022	
Location	ICTC LABORATORY		Calibrated by : godrej&boyce	
Visual Inspection	OK		Accreditation certificate No : CC-2559	
Material Receipt Date	---			
Calibration Performed at	ON SITE			
Date of Calibration	22.03.2022			
Suggested Due Date of Calibration	22.03.2023			
Next Calibration Due (as per customer request)	29.03.2022			
Discipline	MECHANICAL			
Method used	COMPARISON METHOD			
ENVIRONMENTAL CONDITION				
Temperature	30.5 °C			
Relative Humidity	62 %RH			
Condition Of Equipment :	Un-Loaded			
OBSERVATION RESULTS				
SL. NO.	Nominal Value on UUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At 95% Confidence Level (k=2)
1.	180.0	180.1	0.08	0.6 % rdg
Decision Rule is Applied or Not			YES	NO
				✓
Calibration Procedure Based On		: UCSSL/SOP/RPM		
NOTE : -				
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.				
2. This report should not be reproduced except in full without our prior permission in writing.				
3. Calibration certificate without signature are not valid.				
4. All our measurements are Traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.				
5. UUC : Unit under calibration , GUT : Gauge under testing .				
Calibrated By				
			Issued / Approved By (APPU K MANI) Technical Director	

Laboratory : 5/D3, EK Apartments, Model engineering college road Thrikkakara, Edappally, Cochin, Kerala-682021
Registered Office : Plot No. 611-P, Sector - 38, Gurgaon Haryana - 122001

SALES ★ SPARES ★ SERVICE ★ TESTING ★ CALIBRATION ★ MODERNIZATION