



Add : F-26, 2nd Floor, New Ashoka Vihar, XLO Point,
Ambad MIDC, Nashik - 422010.
Mobile : +91 9028646172, +91 9028777244
E-mail : rktechnologies99@gmail.com
Website : www.rktechcalibration.com

Calibration of Electro technical,
Thermal, Pressure, Mechanical,
Force & Hardness parameters.

Validation of All types of
Temperature & Humidity parameters.

CALIBRATION CERTIFICATE

Calibration Item	Micro Pipette	Certificate No	2021-08-11-503
------------------	---------------	----------------	----------------

Date of Receipt	Date of Calibration	Next Recommended Due Date	Certificate Issue Date	Page No
10 January 2022	06 August 2021	05 August 2022	08 August 2021	01 of 01

I. Customer Name	M/S Kushal Path Lab C/O Chirayoo Hospital, F-30 ,near Bajrang Chowk Besides Dashmesh Restaunt ,N5 Cidco Aurangabad
------------------	--

II. Description of Item Under Calibration :			
Instrument ID No	KPL/MP-01	Range	100 to 1000µl
Make / Model		Sr. No	--
Type	Analog	Location	Lab

III Environment Condition :	
Temperature	20 to 35 °C
Relative Humidity	45 % to 75 % RH
Condition of item	Good
Location Of Calibration	In LAB

IV. Detail of Refrence Stanadard used for calibration (Traceable To National / International Standard)					
Instrument Name	ID No	Traceability (Cert No)	Date of Calibration	Valid upto	Traceability
Weight Set	RK/STD-62	C2121452/N/N&V/01	22 February 2021	21 February 2022	AUTOCAL

V: Calibration Result :			
CalibrationPoint	Set UUC Reading	Measure Stadnard Value	Error
µl	µl	µl	µl
50	50.00	50.00	0.00
100	100.00	100.01	-0.01
500	500.00	500.03	-0.03
800	800.00	800.04	-0.04
1000	1000.00	1000.06	-0.06

VI : Note:

- 1) UUC stands for Unit Under Calibration.
- 2) Next calibration date (1 Year) mentioned in the certificate is given as per customer request
- 3) This certificate refers only to the particular item submitted for calibration
- 4) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "RK Technologies, Nashik".
- 5) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By
Mr Sharad Jagdhane
Calibration Engineer



Review & Approved By
Mr Rahul Kasture
Technical Manager

***** End of Report *****