



SHREE SAI DIAGNOSTIC

(Complete Clinical Laboratory Distributor & Biomedical Service)

73970 00464

73970 00462



No. 493, Annam Flats, First Floor,
Anna Nagar, Angammal Colony,
Salem - 636 009.



sams_002@rediffmail.com



CALIBRATION CERTIFICATE

REG: IVD/CAL/2022/8

DATE: 15-03-2022

NAME OF THE CUSTOMER AAROGYA DIAGNOSTICS CENTER 103,1 ST FLOOR,MM COMPLEX DHARMAPURI-636701	MANUFACTURER REMI MODEL -R-4C
PRODUCT NAME CENTRIFUGE	SERIAL NO EHLC-7549

This is to certify that the above mentioned equipment has been inspected and calibrated as per the protocol on 10 MARCH 2022. The Equipment is working satisfactorily. This Calibration is valid through 09 MARCH 2023.

S.NO	TEST PARMETERS	MEASURED/CALIBRATED	REMARKS
01	INPUT VOLTAGE	230V	PASSED
02	POWER CONUMPTION	300VA	PASSED
03	INPUT FREQUENCY	50HZ	PASSED
04	LAMP	6V,20 W	PASSED



SHREE SAI DIAGNOSTIC

(Complete Clinical Laboratory Distributor & Biomedical Service)

73970 00464
73970 00462



No. 493, Annam Flats, First Floor,
Anna Nagar, Angammal Colony,
Salem - 636 009.



sams_002@rediffmail.com



S.NO	DUC READING(RPM)	STANDARD READING(RPM)	CORRECTION(RPM)
01	1400	1399.58	0.42
02	2100	2099.41	0.59
03	3500	3499.12	0.88

NOTE:

1. Calibration Points were selected as per customer request
2. Each reading were taken by repeatability test (i.e.) Each reading in arrow is a mean of 10 readings
3. Result reported are valid at the time of and under the stated conditions of measurements.

ANY REMARKS:

NIL



SHREE SAI DIAGNOSTIC
SALEM