

## Medical Engineering & Services Things Beyond Thoughts

Abhcon Crown,X1/411/21,Ambadi Lane Kokkalai,Thrissur-21. Ph: 0487 2420367 Email: mescalibration.tsr@gmail.com Visit us @www.mescalibration.com

CC254322000000051P

CALIBRATION CERTIFICATE

**ULR Number** 

MES/180122/C4 Certificate Number

SEVENA HITECH LAB, PARAPPUR, THRISSUR. **Customer Details** 

18/01/2022 Date of SRF 18/01/2022 18/01/2022 Date of Calibration Date of Issue

17/01/2023 **Next Calibration Due** 

**DETAILS OF DEVICE UNDER CALIBRATION(DUC)** 

YE5A425926 Micropipette Serial No. Nomenclature

**DRAGON LAB** ID No. Make / Model

 $0.5 \mu l$ Good Resolution **Equipment Condition** 

**BIOCHEMISTRY** 5-50µL Location Range

**DETAILS OF MASTER EQUIPMENT USED** 

 $0 - 210 \, \text{gm}$ Range **Electronic Precision Balance** Nomenclature 0.01/0.1 mg Readability Boeco 27104178 Make Serial No. BBC-22

Model 05.10.2022 Calibration Due Date TVSPL21/10/227-01 Certificate No.

**ENVIRONMENTAL CONDITIONS** 

Water Temperature(°C) **Atmospheric Pressure (mbar)** Relative Humidity (%) Temperature(°C) 26.6 1004 56 26.5

## **CALIBRATION RESULTS**

CALIBRATION RESULTS							
Discipline :		Mechanical		Parameter		: Volume	
SI.	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
No.			μΙ	±μl	μl	±μl	±μl
	μl 5*	μι 5.02*	-0.02*	0.13*	0.05*	0.08*	0.420*
1	25	25.02	-0.02	0.50	0.07	0.20	0.434
2	50	50.32	-0.32	0.50	0.07	0.20	0.432
	50						

## Notes: -

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines; MES work procedure: MES/WP/VOLUME.

The tests and /or Calibration marked with an \* are not accredited by NABLS

Calibrated by

(Calibration Engineer)

THRISSUR.

Authorized S ignatory

**Binoi Thomas** (Technical Manager)