

RML – Quality Assurance Program (RML – QAP)



Urine Routine Analysis

ALL METHOD REPORT

Cycle - 11/2022 Round No -4 Date: 30/04/2022

Lab Code: 2449

Parameters	Units	No.of. Participants	Consensus		Consensus	Your Value	Remark
			Number	%	Result	Tour value	Remark
Specific Gravity		217	205	94%	1.010-1.020	1.020	Within Consensus
pH		218 '	196	90%	6.0	6.0	Within Consensus
Leukocytes	/µl	212	210	99%	Negative	Negative	Within Consensus
Nitrite		209	204	98%	Negative	Negative	Within
Protein	mg/dl	218	215	99%	Negative	Negative	Within Consensus
Glucose	mg/dl	218	218	100%	Negative	Negative	Within
Ketones		218	218	100%	Negative	Negative	Within
Urobilinogen	mg/dl	216	215	100%	Negative	Negative	Within Consensus
Bilirubin	mg/dl	216	212	98%	Negative	Negative	Within
Blood (Hemoglobin)	/µl	215	214	99%	Negative	Negative	Within Consensus

Legends Consensus Consensus Consensus Consensus	Legends	≥ 60%	(*) Outside	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---	---------	-------	-------------	---------------------	------------------	---------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Checked By:

Prepared By: SK

Programme Director

Dr.Bandana Mehrotra

Page 1 of 4

