


CONTROLLED COPY

AUTH: *[Signature]*

QA FUNCTION

	PATHKIND DIAGNOSTICS PVT LTD	ISSUE DATE: 01-08-2017	AMENDEMENT DATE:	KOTHYAT, GROUND FLOOR, KALA AAM CHORAHA, DM ROAD, BULANDSHEHAR, UTTAR PRADESH-203061
	DAILY MAINTENANCE LOG FOR CENTRIFUGE	ISSUE NO: 01	AMENDEMENT NO:	

MONTH/YEAR: *April 2022*

SR NO:

Date	CLEANING FROM OUTSIDE	CLEANING FROM INSIDE*	CLEANING OF TUBE HOLDERS	CARBON BRUSH CHANGED	SIGN OF PHLEBO	SIGN OF CC OWNER
1	✓				<i>[Signature]</i>	<i>[Signature]</i>
2	✓				<i>[Signature]</i>	<i>[Signature]</i>
3	✓				<i>[Signature]</i>	<i>[Signature]</i>
4	✓				<i>[Signature]</i>	<i>[Signature]</i>
5	✓				<i>[Signature]</i>	<i>[Signature]</i>
6	✓				<i>[Signature]</i>	<i>[Signature]</i>
7	✓				<i>[Signature]</i>	<i>[Signature]</i>
8	✓				<i>[Signature]</i>	<i>[Signature]</i>
9	✓				<i>[Signature]</i>	<i>[Signature]</i>
10	✓				<i>[Signature]</i>	<i>[Signature]</i>
11	✓				<i>[Signature]</i>	<i>[Signature]</i>
12	✓				<i>[Signature]</i>	<i>[Signature]</i>
13	✓				<i>[Signature]</i>	<i>[Signature]</i>
14	✓				<i>[Signature]</i>	<i>[Signature]</i>
15	✓				<i>[Signature]</i>	<i>[Signature]</i>
16	✓				<i>[Signature]</i>	<i>[Signature]</i>
17	✓				<i>[Signature]</i>	<i>[Signature]</i>
18	✓				<i>[Signature]</i>	<i>[Signature]</i>
19	✓				<i>[Signature]</i>	<i>[Signature]</i>
20	✓				<i>[Signature]</i>	<i>[Signature]</i>
21	✓				<i>[Signature]</i>	<i>[Signature]</i>
22	✓				<i>[Signature]</i>	<i>[Signature]</i>
23	✓				<i>[Signature]</i>	<i>[Signature]</i>
24	✓				<i>[Signature]</i>	<i>[Signature]</i>
25	✓				<i>[Signature]</i>	<i>[Signature]</i>
26	✓				<i>[Signature]</i>	<i>[Signature]</i>
27	✓				<i>[Signature]</i>	<i>[Signature]</i>
28	✓				<i>[Signature]</i>	<i>[Signature]</i>
29	✓				<i>[Signature]</i>	<i>[Signature]</i>
30	✓				<i>[Signature]</i>	<i>[Signature]</i>
31						

\* In case of spillage/tube breakage



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
 Phone / Fax : (011) 22453259, 22434362, 9599482307  
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900001580	Issued On:	22/03/2022
ULR No:	CC214622900001580F	Calibration Date:	19/03/2022
Job Identification No:	ICS/C/AS/03/1580	Next Calibration Date:	18/03/2023
Ref. No:	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

Name	Laboratory Centrifuge, Fitted with Digital RPM Indicator		
Make/Model	Remi / R-8C Plus	Visual Inspection	Ok
Serial No.	ZDGN-18557	I.D. No.	ACC/001
Range	Up to 6000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

<b>USED EQUIPMENT DETAILS</b>	Not Applicable
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

#### RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

CENTRIFUGE READING	CORRECTED VALUE OF TACHOMETER READING (AVERAGE)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1560	1562.5	± 6
2230	2233.7	± 6
3140	3145.2	± 6
4350	4357.1	± 7
5710	5720.8	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARTIK KADHAR  
Quality Manager



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
 Phone / Fax : (011) 22453259, 22434362, 9599482307  
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900001581	Issued On:	22/03/2022
ULR No:	CC214622900001581F	Calibration Date:	19/03/2022
Job Identification No:	ICS/C/AS/03/1581	Next Calibration Date:	18/03/2023
Ref. No:	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

#### EQUIPMENT DESCRIPTION

Name	Laboratory Centrifuge, Fitted with Digital RPM Indicator		
Make/Model	Romi / R-8C Plus	Visual Inspection	Ok
Serial No.	ZDGN-18563	I.D. No.	GEN/001
Range	Up to 6000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

#### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS	Not Applicable
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

RESULTS: Mechanical Calibration  
MEASUREMENT: (In RPM)

CENTRIFUGE READING	CORRECTED VALUE OF TACHOMETER READING (AVERAGE)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1290	1287.5	± 6
2330	2325.4	± 6
3510	3503.5	± 6
4670	4661.7	± 7
5820	5808.3	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

KARTIK KATHOR  
Authorised Signatory  
Quality Manager