



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
 Phone / Fax : (011) 22453259, 22434362, 9599482307  
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Calibration Certificate No.	22100005046	Issued On	22/03/2022
ULR No.	CC214622100005046F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5046	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

EQUIPMENT DESCRIPTION			
Name	Glass Thermometer for Refrigerator		
Make/Model	Zeal	Visual Inspection	Ok
Serial No.	---	I.D. No.	GEN/005
Range	-10°C to 110°C	Least Count	1°C
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS			
Temperature	25.0 °C ± 3.0 °C	Humidity	50 ± 10 % RH

STANDARD REFERENCE DETAILS					
Traceable to National / International Standard					
Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Methanol Bath
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS-CAL/SOP-TH01, ITS-90 (Comparison Method)

RESULTS: Thermal Calibration		
U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (In °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (In °C)
2	1.97	± 0.61
4	3.95	± 0.61
5	4.92	± 0.61
7	6.89	± 0.61
8	7.86	± 0.61

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.  
 \*\*\* END OF REPORT \*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
 Indian Calibration Services

Checked By

Authorised Signatory



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
 Phone / Fax : (011) 22453259, 22434362, 9599482307  
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Calibration Certificate No.	22100005047	Issued On	22/03/2022
ULR No.	CC214622100005047F	Calibration Date	19/03/2022
Job Identification No.	ICS/C/T/03/5047	Next Due Date	18/03/2023
Ref. No.	SRF, Dated - 16/03/2022		

<b>CALIBRATED FOR:</b>	M/s Path Kind Diagnostics Pvt. Ltd. Ground Floor, Dev Sagar Lodge, Kalaam Chouraha, DM Road, Bulandsahar-203001, (U.P.)
------------------------	---

EQUIPMENT DESCRIPTION			
Name	Glass Thermometer for Refrigerator		
Make/Model	Zeal	Visual Inspection	Ok
Serial No.	---	I.D. No.	GEN/006
Range	-10°C to 110°C	Least Count	1°C
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS			
Temperature	25.0 °C ± 3.0 °C	Humidity	50 ± 10 % RH

STANDARD REFERENCE DETAILS					
Traceable to National / International Standard					
Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Methanol Bath
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90 (Comparison Method)

RESULTS: Thermal Calibration		
U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
2	2.04	± 0.61
4	4.07	± 0.61
5	5.08	± 0.61
7	7.11	± 0.61
8	8.13	± 0.61

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\* END OF REPORT \*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

Authorised Signature

**CONTROLLED COPY**  
 AUTH: *[Signature]*  
**QA FUNCTION**



PATHKIND DIAGNOSTICS PVT LTD  
 DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER

ISSUE DATE:  
01-06-2017  
 ISSUE NO:  
01

AMENDMENT DATE:  
 AMENDMENT NO:

KOTHYAT, GROUND FLOOR,  
 KALA AAM CHORAKA, DW  
 ROAD, BULANDSHEHAR,  
 UTTAR PRADESH, 203001

YEAR: 2022 (Celfrost)

SR NO:

DATE	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
JAN	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER	A																													
	SIGNATURE	(A)	(A)	(A)	(A)	0	0	0	0	0	0	0	0	0	0	0	0	0	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	
FEB	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER																														
	SIGNATURE	(A)	(A)	(A)	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	
MAR	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER																														
	SIGNATURE	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	(A)	0	0	0	0	0	0	0	0	0	0	0	(A)	(A)	(A)	
APR	REFRIGERATOR	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	6°C	7°C	7°C	7°C	7°C	7°C	7°C	7°C	7°C	7°C	6°C	6°C	6°C	6°C	6°C	
	FREEZER																														
	SIGNATURE	(A)	(A)	(A)	(A)	(A)	(A)	(A)	0	0	0	0	(A)	(A)	(A)	(A)	0	0	0	0	0	0	0	0	0	0	0	0	0	0	
MAY	REFRIGERATOR																														
	FREEZER																														
	SIGNATURE																														
JUN	REFRIGERATOR																														
	FREEZER																														
	SIGNATURE																														


ACCEPTABLE LIMITS  
 REFRIGERATOR: 2-4° C      FREEZER: -10 to -30° C

SECTION HEAD SIGN:

DOC NO.: P005/GEN/007.0 *F 12*      DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER      Page 01 of 02  
 ISSUED BY: DR ASHOK RATTAN

Scanned with CamScanner

**CONTROLLED COPY**

AUTH: 

**QA FUNCTION**

	59 m 5479 PATHIND DIAGNOSTICS PVT LTD	ISSUE DATE: 01-08-2017	AMENDMENT DATE:  	KOTHNAT GROUND FLOOR, KALA AAM CHORAMA, DM ROAD, BULANDSHAHAR, UTTAR PRADESH 203001
DAILY TEMPERATURE MONITORING LOG FOR REFRIGERATOR AND FREEZER		ISSUE NO: 01	AMENDMENT NO:  	

YEAR 2022

SR NO:

DATE	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
JAN	REFRIGERATOR	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°
	FREEZER	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C
	SIGNATURE	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R
FEB	REFRIGERATOR	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°
	FREEZER	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C
	SIGNATURE	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R
MAR	REFRIGERATOR	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°
	FREEZER	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C
	SIGNATURE	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R
APR	REFRIGERATOR	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°	19°
	FREEZER	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C	5°C
	SIGNATURE	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R	R
MAY	REFRIGERATOR	19°	19°	19°	19°																										
	FREEZER	5°C	5°C	5°C	5°C																										
	SIGNATURE	R	R	R	R																										
JUN	REFRIGERATOR																														
	FREEZER																														
	SIGNATURE																														

ACCEPTABLE LIMITS  
 REFRIGERATOR: 2-8° C      FREEZER: -10 to -30° C

SECTION HEAD SIGN: 

Scanned with CamScanner