

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2022



PC-1024

Lab Name **DR REETU NAGE** Lab No **9324**
Constituent Group **Chemistry II** Date of Result Entered : **16/03/2022**
PT item **Lyophilized Serum** Date of Report Published : **06/04/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	109.92	9.69	10.65	121.4 mg/dL	1.08	0.76
2	UREA II	Urease UV / GLDH II	Transasia / Erba	601	80.81	11.24	9.09	93.9 mg/dL	1.44	0.74
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	459	3.84	11.18	0.43	3.84 mg/dL	0.00	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	668	1.57	19.28	0.30	1.64 mg/dL	0.23	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	593	5.34	8.62	0.46	5.5 g/dL	0.35	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	633	3.15	9.30	0.29	3.08 g/dL	-0.24	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	637	8.01	11.65	0.93	9.7 mg/dL	1.81	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	700	105.41	11.28	11.89	106 mg/dL	0.05	0.90
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

LAB ADDRESS :


DR REETU NAGE
C/O DR ANIL NAGE, SHOP NO. 32, MOHAN RAKHIWALI GALI, MOHANA ROAD, BALLABGARH
FARIDABAD
HARYANA121004

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **DR REETU NAGE** Lab No **9324**
 Month **February** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 15/02/2021

Date of Report Published : 06/03/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	747	251.19	10.54	26.72	274.3 mg/dl	0.33	1.55
2	UREA II	Urease UV / GLDH II	Transasia / Erba	584	54.62	11.45	6.25	50 mg/dl	-0.74	0.52
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	452	2.80	10.81	0.30	2.7 mg/dl	-0.33	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	532	1.22	19.52	0.24	1.18 mg/dl	-0.17	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	580	5.56	10.15	0.56	6.2 g/dl	1.13	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	570	3.31	9.68	0.32	3.1 g/dl	-0.66	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	576	3.83	16.03	0.61	3.5 mg/dl	-0.54	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	655	110.72	11.41	12.63	92.4 mg/dl	-1.45	0.99
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **DR REETU NAGE** Lab No **9324**
 Month **March** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 23/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	160.91	10.47	16.84	161.7 mg/dl	0.05	1.20
2	UREA II	Urease UV / GLDH II	Transasia / Erba	605	42.26	12.78	5.40	45.4 mg/dl	0.58	0.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	491	1.67	11.16	0.19	1.72 mg/dl	0.72	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	658	3.60	13.31	0.48	4.3 mg/dl	1.46	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.89	9.22	0.54	5.5 g/dl	-0.72	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	599	3.52	9.28	0.33	3.5 g/dl	-0.06	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	668	7.46	13.58	1.01	6.9 mg/dl	-0.54	0.08
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	717	118.97	12.88	15.32	110 mg/dl	-0.59	1.14
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ± 1.0 to ± 2.0		Good.								
Between ± 2.0 to ± 3.0		Accept with caution. Warning Signal.								
Beyond ± 3.0		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

VIEW LAB MONTHLY SUMMARY

Lab Name	DR REETU NAGE	Details About Robust Analysis
Lab No	9324	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 17/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	713	318.19	10.10	32.15	334.6 mg/dl	0.51	2.41
2	UREA II	Urease UV / GLDH II	Transasia / Erba	522	27.05	13.69	3.70	28.4 mg/dl	0.36	0.32
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	468	6.55	11.06	0.72	6.52 mg/dl	-0.04	0.07
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	676	3.28	15.32	0.50	3.76 mg/dl	0.96	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	537	5.61	8.16	0.46	5.7 g/dl	0.20	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	550	3.32	9.16	0.30	3.1 g/dl	-0.72	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	601	4.45	17.19	0.76	4.2 mg/dl	-0.33	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	677	113.01	11.66	13.18	107 mg/dl	-0.46	1.01
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
--------------------------------	--------------------------------	------------------------------------	--

Yearly Summary	Change Password	Home	Sign-out
--------------------------------	---------------------------------	----------------------	--------------------------

VIEW LAB MONTHLY SUMMARY

Lab Name	DR REETU NAGE	Details About Robust Analysis
Lab No	9324	Detail About Monthly Summary
Month	May	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 26/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	670	132.95	10.22	13.59	136.1 mg/dl	0.23	1.05
2	UREA II	Urease UV / GLDH II	Transasia / Erba	562	63.66	12.61	8.03	67.3 mg/dl	0.45	0.68
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	457	1.89	11.86	0.22	1.6 mg/dl	-1.29	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	634	2.47	17.92	0.44	2.4 mg/dl	-0.16	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	588	5.74	9.09	0.52	5.4 g/dl	-0.65	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	567	3.39	9.40	0.32	3.5 g/dl	0.34	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	636	7.08	13.17	0.93	7.2 mg/dl	0.13	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	671	109.08	12.80	13.96	101 mg/dl	-0.58	1.08
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
--------------------------------	--------------------------------	------------------------------------	--

Yearly Summary	Change Password	Home	Sign-out
--------------------------------	---------------------------------	----------------------	--------------------------

VIEW LAB MONTHLY SUMMARY

Lab Name	DR REETU NAGE	Details About Robust Analysis
Lab No	9324	Detail About Monthly Summary
Month	June	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	725	197.93	10.20	20.20	200.4 mg/dl	0.12	1.50
2	UREA II	Urease UV / GLDH II	Transasia / Erba	579	76.33	12.48	9.52	72.04 mg/dl	-0.45	0.79
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	460	4.43	12.53	0.56	5.1 mg/dl	1.21	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	649	3.70	13.01	0.48	4.4 mg/dl	1.45	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.61	10.40	0.58	5.8 g/dl	0.33	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	597	3.32	9.58	0.32	3.34 g/dl	0.04	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	621	4.70	17.49	0.82	4 mg/dl	-0.85	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	692	105.54	13.56	14.31	123.1 mg/dl	1.23	1.09
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

LAB MONTHLY SUMMARY



Lab Name **DR REETU NAGE** Lab No **9324**
 Month **July** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 17/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	751	148.33	10.85	16.09	137 mg/dl	-0.70	1.17
2	UREA II	Urease UV / GLDH II	Transasia / Erba	570	32.47	13.36	4.34	32 mg/dl	-0.11	0.36
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	479	2.77	11.56	0.32	2.49 mg/dl	-0.88	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	655	2.84	12.80	0.36	2.05 mg/dl	0.31	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	609	5.55	8.57	0.48	5.45 g/dl	-0.21	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	607	3.32	8.78	0.29	3.38 g/dl	0.21	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	686	4.95	17.29	0.86	6.33 mg/dl	1.61	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	783	112.10	13.65	15.31	121.4 mg/dl	0.61	1.09
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****