



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

| | | | |
|---------------------------------|---|---------------------------|--------------------|
| Customer name And Address | M/S: Vision Diagnostic Centre A Unit of Dr. M.C. Agrawal Hospital & Research Centre Pvt. Ltd. Firozabad - 283203 | ULR No. | CC223922000001627F |
| | | Service request No & date | P/03/29/04/2022 |
| | | Date of calibration | 29/04/2022 |
| | | Valid up to | 28/04/2023 |
| | | Date of issue | 30/04/2022 |

| Instrument Details | | | |
|--------------------|----------------------|--------------------------|---------|
| Instrument name | Thermometer | S. No. | D17P14B |
| Make/ Model No. | Made In India/ TM21A | Location | ----- |
| Range / Size | -80 to 35 °C | Accuracy | ----- |
| Least Count | 1 °C | Visual Inspection | OK |
| I.D. No. | ----- | Calibration Performed At | ----- |

| Detail of reference standards & Major equipments used | | | |
|---|----------------------------|--|--|
| Equipment Name | RTD With Digital Indicator | | |
| Make | YCT / SBI | | |
| Model / SR No. | 11080368 | | |
| Certificate No. | CC325722000000135F | | |
| Calibration Validity | 10/02/2023 | | |
| Calibration by | KL Technology | | |

| | | | | |
|----------------------------|-------------------|---------|-----------------------|---------|
| Environmental Condition | Temperature | 25±4 °C | Calibration Reference | IS:6274 |
| | Relative Humidity | 50±20 % | Calibration Procedure | CP-02T |

| Calibration Results | | | |
|---------------------|-------------------------|-----------------------------|--|
| Serial No. | Std. Reading in (°C) | Measured Reading in (°C) | Uncertainty At 95% C.L. (coverage factor K=2) |
| 01. | -25.13 | -24 | ±0.3 °C |
| 02. | -20.15 | -20 | |
| 03. | -10.85 | -10 | |
| 04. | -5.71 | -5 | |
| 05. | 0.02 | 0 | |

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(D.B SAXSENA)
Form No. QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1