



PC-1034

**CMC VIROEQAS**  
**By National Reference Laboratory (NRL)**  
**Under National AIDS Control Organisation (NACO)**  
**PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu**  
**Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455**

### **PANEL: HIV Serology**

Laboratory ID: KL-SRL-03-IC20

Distribution Round: Round -2, 2021-2022

Panel dispatch date: 7-Dec-21

Result received month: 07-Jan-22

| SL NO | SAMPLE ID | INTENDED RESULT    | YOUR RESULT        | MATCHING YES / NO | % MATCHING |
|-------|-----------|--------------------|--------------------|-------------------|------------|
| 1     | S022101   | Positive for HIV-1 | Positive for HIV-1 | YES               | 100        |
| 2     | S022102   | Positive for HIV-1 | Positive for HIV-1 | YES               | 100        |
| 3     | S022103   | Negative           | Negative           | YES               | 100        |
| 4     | S022104   | Negative           | Negative           | YES               | 100        |

**REMARKS: PERFORMANCE IS SATISFACTORY**

Comments:

Participants are scored based on qualitative result.

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore  
The data in this CMCVIROEQAS reports are confidential.



PC-1034

**CMC VIROEQAS**  
**By National Reference Laboratory (NRL)**  
**Under National AIDS Control Organisation (NACO)**  
**PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu**  
**Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455**

**Assigned Value:**

Assigned value is determined using multiple assays/multiple testing of the same analyte. The proficiency testing material is a pooled lyophilized plasma sample prepared by following institutional proficiency testing standard operating protocol (PT-SOP) by trained staff. None of the work related to testing, preparation and packaging of sample is subcontracted. The homogeneity of the PT material is determined using multiple testing of multiple aliquots. The stability of the testing material is determined by assessing the reactivity of the specimen till the closing date.

Enquires: For queries, please contact CMCVIROEQAS coordinator at email: [nrl.cmc@gmail.com](mailto:nrl.cmc@gmail.com)

**Name of CMCVIROEQAS coordinator**

**Signature**

**Dr. Rajesh Kannangai**

**Department of Clinical Virology**

**Report Dispatch Date: 21-Feb-22**

Report authorized by: CMCVIROEQAS Coordinator

**END OF REPORT**

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore  
The data in this CMCVIROEQAS reports are confidential.