



BIO-CHEMISTRY-01

ALL METHOD REPORT

Lab Code: 2160

Cycle - 11/2022

Round No - 3

Date: 05/04/2022

Parameters	Units	No.of. Participants	Group Mean	Standard deviation (SD)	Uncertainty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index (SDI)
Albumin	g/dL	155	2.9	0.2	0.02	2.5-3.3	2.8	-0.5
Alkaline Phosphatase	U/L	145	310.2	82.2	8.53	145.8-474.7	.	-
Bilirubin Total	mg/dL	156	5.0	0.6	0.06	3.9-6.1	4.7	-0.5
Calcium(Total)	mg/dL	132	12.1	1.0	0.11	10.1-14.0	.	-
Cholesterol (Total)	mg/dL	152	273.7	16.1	1.63	241.5-306	254	-1.2
Creatinine	mg/dL	158	4.2	0.4	0.04	3.4-5.0	4.6	1.0
Glucose	mg/dL	158	264.9	15.5	1.54	234.0-295.9	278	0.8
HDL	mg/dL	143	96.2	33.9	3.54	28.4-164	95.0	0.0
Potassium	mEq/L	134	5.8	0.3	0.03	5.1-6.4	5.9	0.3
Protein (Total)	g/dl	154	4.7	0.4	0.04	3.8-5.5	4.1	-1.5
Sodium	mEq/L	134	154.5	5.1	0.55	144.2-164.8	159	0.9
SGPT/ALT	U/L	155	121.4	13.4	1.35	94.5-148.2	109	-0.9
SGOT/AST	U/L	155	133.3	12.7	1.28	107.9-158.7	130	-0.3
Triglyceride	mg/dL	152	252.2	16.0	1.62	220.1-284.2	234	-1.1
Urea	mg/dL	157	116.1	9.7	0.97	96.8-135.4	105	-1.1
Uric Acid	mg/dl	157	8.9	0.9	0.09	7.2-10.7	8.8	-0.1

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Need Improvement	Unacceptable Performance Needs Urgent action

Legends	(*) Excluded From Group	{ } Not Reported	(#) Late Result	(\$) Reported in
---------	-------------------------	------------------	-----------------	------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By:

Prepared by: SV

Programme Director

Dr. Bandana Mehrotra

Page 1 of 7



Address: B-171, Nirala Nagar, Lucknow - 226 020 (Ph) : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence