

Calibration Certificate

Calibration Report

**Name of the customer:** SRL DIAGNOSTIC LTD. DANKUNI

**Instrument type:** Fully Auto Bio-Chemistry Analyser

**Model number:** Sphera

**Serial number:** 201878003

**Date of Calibration :** 13.12.2021

**Next Calibration Due Date :** 13.06.2022

**Power checks:**

Power supply

AC 229.8V-50 Hz

Plug top

5 Amps

Earth voltage

1.1 Volt

**Calibration :**

Clean Dust from Instrument's inside & outside. Check & Done  
Mechanical calibration.

**Instrument performance Check :**

- |                                       |         |         |
|---------------------------------------|---------|---------|
| 1. Filter Gain Check for all Filter - | Checked | Passed  |
| 2. Reaction Cuvette Blank Check :     | Checked | Passed. |
| 3. Calibration of all Parameter :     | Done    | Passed. |

**Operational check:**

Precision test

Checked

Passed

Quality control test


Checked

Passed

The equipment Sphera, serial. No: **201878003** is Calibrated.

The operational and performance is done and observations are:

- (1) All system components are operating correctly.
- (2) All tests were carried out in accordance with qualification protocol and standard operating procedure.
- (3) The performance of the Sphera (s/n **201878003**) is as per the acceptance criteria.
- (4) This Calibration Valid for 06 ( Six ) months.

  
Service Engineer

Date: 13.12-21

**Accurex Biomedical Pvt.Ltd** 212, Udyog Mandir No. 1, 7-C, Bhagoji Keer Marg, Mahim, Mumbai - 400 016  
(INDIA)

☎: (9122) 6744,6744, Fax : (9122) 67446755 E-Mail : [accurex@vsnl.com](mailto:accurex@vsnl.com) Website:<http://www.accurexbiomedical.com>