

HORIBA

Explore the future

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2021/133879

24th September 2021

CALIBRATION CERTIFICATE

This is to certify that the hematology analyzer **ABX Micros ES 60** bearing serial no: 005ESOH16385 installed at SRL LIMITED was calibrated on 24th September 2021.

Calibrator : ABX MINOCAL
Lot No. : CX462
Expiry Date : 05th October 2021

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 23rd September 2022



Shrish Dixit

(Marketing Manager)

For Horiba India Pvt Ltd