

## CALIBRATION REPORT

Details of Device Under Calibration :

Date:05-03-2022

<b>Customer Name &amp; Address:</b> SR DIAGNOSTIC CENTRE,SEWA RURAL JHAGADIA	<b>Certificate ID</b> : AZ/21-22/ 1418
<b>Name of Equipment:</b> Centrifuge Machine	<b>Calibration Date</b> : 22-02-2022
<b>Manufacturer</b> : LABLINE	<b>Calibration Due Date</b> : 21-02-2023
<b>Model</b> : NS	<b>Location of equipment</b> : LABORATORY
<b>Serial No</b> : NS	<b>Ambient Temperature</b> : 27 °C

### Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali Date	Due Date
1	Digital Tachometer	System	HMT 560	13-1011	01-02-2022	31-01-203

Condition Of Equipment at the time of Calibration : GOOD


### Test Report:

Sr No	Parameter	Range	Set Value	Observed	Error	Acceptable Range (+/-)	Result
1	RPM	0 to 3500	1000	994	6	5%	Pass
2			2000	1990	10	5%	Pass
3			3000	2980	20	5%	Pass
4			3500	3470	30	5%	Pass

### Note:

- (1) Results reported are valid at the time of and under the stated conditions of measurement.
- (2) Calibration has been done as per customer's requirement.
- (3) Measurements are traceable to the National Standard
- (4) NS means Not Specified and DUC means Device Under Calibration
- (5) The reported measurement uncertainty is estimated at a level of confidence of approx 95 % with a coverage factor  $k = 2$ .
- (6) Cali means calibration .

Calibrated By:   
Gaurang Jansari

Approved By:   
Gaurav Desai



THIS DOCUMENT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL OF "A TO Z BIOMEDICAL SERVICES"

## CALIBRATION REPORT

Details of Device Under Calibration :

Date:05-03-2022

<b>Customer Name &amp; Address:</b> SR DIAGNOSTIC CENTRE,SEWA RURAL JHAGADIA	<b>Certificate ID</b> : AZ/21-22/ 1420
<b>Name of Equipment:</b> Centrifuge Machine	<b>Calibration Date</b> : 22-02-2022
<b>Manufacturer</b> : R-8C PLUS	<b>Calibration Due Date</b> : 21-02-2023
<b>Model</b> : REMI	<b>Location of equipment</b> : LABORATORY
<b>Serial No</b> ZHGN-2438	<b>Ambient Temperature</b> : 27 °C

**Test Equipment Used:**

sr no	Test equipment	Company	Model	Sr no.	Cali Date	Due Date
1	Digital Tachometer	System	HMT 560	13-1011	01-02-2022	31-01-203

Condition Of Equipment at the time of Calibration : GOOD

**Test Report:**

Sr No	Parameter	Range	Set Value	Observed	Error	Acceptable Range (+/-)	Result
1	RPM	0 to 3500	1000	995	5	5%	Pass
2			2000	1995	5	5%	Pass
3			3000	2985	15	5%	Pass
4			3500	3475	25	5%	Pass

**Note:**

- Results reported are valid at the time of and under the stated conditions of measurement.
- Calibration has been done as per customer's requirement.
- Measurements are traceable to the National Standard
- NS means Not Specified and DUC means Device Under Calibration
- The reported measurement uncertainty is estimated at a level of confidence of approx 95 % with a coverage factor  $k = 2$ .
- Cali means calibration .

Calibrated By:

  
Gaurang Jansari

Approved By:

  
Gaurav Desai



THIS DOCUMENT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL OF "A TO Z BIOMEDICAL SERVICES"