

## CALIBRATION REPORT

Details of Device Under Calibration :

Date:05-03-2022

Customer Name & Address: SR DIAGNOSTIC CENTRE,SEWA RURAL JHAGADIA		Certificate ID : AZ/21-22/ T1
Name of Equipment: THERMOMETER		Calibration Date : 22-02-2022
Manufacturer : SYMBOL		Calibration Due Date : 21-02-2023
Model : ST-1A		Location of equipment : LABORATORY
SN : NS		Environment Condition Temperature: 29° C
		Humidity: 35%

### Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali. Date	Due date
1	Temperature Indicator with sensor	Omron	DS708	19090715	01-02-2022	31-01-2023

### Test Report:


Condition of Equipment at the time of Calibration: Good

Sr No	Test	Unit	Set Value	Observed Value	Error	Acceptable Range (+/-)	Result
1	Temperature	° C	10	10	0	5%	Pass
2			20	20	0	5%	Pass
3			30	30	0	5%	Pass
4			40	39.9	0.1	5%	Pass

### Remarks:

- (1) Results reported are valid at the time of and under the stated conditions of measurement.
- (2) The Calibration due date is decided by user, However, the due date shall be recommended based on the customer's written advice.
- (3) Calibration has been done as per customer's requirement.
- (4) Measurements are traceable to the National Standard.
- (5) This document shall not be reproduced, except in full, without the written approval of "A to Z Biomedical Services"
- (6) DUC means Device Under Calibration, NS means Non Specified.
- (7) The Calibration was performed by comparison of DUC's Values ( Observed Values) and Test Equipment's Values ( Set Value).

Calibrated By:  Gaurang Jansari	Approved By:  Gaurav Desai
---	---



## CALIBRATION REPORT

Details of Device Under Calibration :

Date:05-03-2022

<b>Customer Name &amp; Address:</b> SR DIAGNOSTIC CENTRE,SEWA RURAL JHAGADIA		Certificate ID : AZ/21-22/ T2
Name of Equipment: THERMOMETER		Calibration Date : <b>22-02-2022</b>
Manufacturer : SYMBOL		Calibration Due Date : <b>21-02-2023</b>
Model : ST-1A		Location of equipment : LABORATORY
SN : NS		Environment Condition Temperature: 29° C
		Humidity: 35%

### Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali. Date	Due date
1	Temperature Indicator with sensor	Omron	DS708	19090715	01-02-2022	31-01-2023

### Test Report:

Condition of Equipment at the time of Calibration: Good

Sr No	Test	Unit	Set Value	Observed Value	Error	Acceptable Range (+/-)	Result
1	Temperature	° C	10	10	0	5%	Pass
2			20	20	0	5%	Pass
3			30	29.8	0.2	5%	Pass
4			40	39.9	0.1	5%	Pass

### Remarks:

- (1) Results reported are valid at the time of and under the stated conditions of measurement.
- (2) The Calibration due date is decided by user, However, the due date shall be recommended based on the customer's written advice.
- (3) Calibration has been done as per customer's requirement.
- (4) Measurements are traceable to the National Standard.
- (5) This document shall not be reproduced, except in full, without the written approval of "A to Z Biomedical Services"
- (6) DUC means Device Under Calibration, NS means Non Specified.
- (7) The Calibration was performed by comparison of DUC's Values ( Observed Values) and Test Equipment's Values ( Set Value).

Calibrated By:  Gaurang Jansari	Approved By:  Gaurav Desai
---	---

