



APPENDIX E

THERMOMETER CALIBRATION SHEET

Calibration of Thermometer

Name and ID of equipment: IC7C Refrigerator -01 (Kasturba Hospital)
ID of Thermometer to be calibrated: IC7C 7-1
Temperature range of thermometer: -40°C to +50°C
Name and ID of Calibrator: Cal T-2
Temperature range of Calibrator: -10°C to +110°C
Thermometer calibration done on: 02/04/2022
Thermometer calibration due on: 01/10/2022
Thermometer calibration done at: HIV Lab, MAMC

Time of reading	Temperature of Working Thermometer	Temperature of Calibrator	Difference in readings
After 1 hour	4°C	4°C	0
After 2 hour	4°C	4°C	0
After 3 hour	4°C	4°C	0

Correction Factor: 0

Comments: Nil

Calibrated By: Ramesh Kumar (LT)

Verified By: Priya Shukla (TO)

J. S. Mem
Laboratory Technician
02.04.22

Shukla
02/04/2022
Technical Officer

Document Type: SOP for operation and monitoring of Refrigerator	Document No: MAMC/HIV/SOP/07	
Issue No.:02	Issue Date: 10.09.2016	Page No.: 1 of 1

Department of Microbiology
Maulana Azad Medical College
New Delhi-110002